PM-8000

Portable Multi-parameter

Patient Monitor

Operation Manual

Intellectual Property Statement

SHENZHEN MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD. (hereinafter called Mindray) owns the intellectual property rights to this product and this manual. This manual may refer to information protected by copyrights or patents and does not convey any license under the patent rights of Mindray, nor the rights of others. Mindray does not assume any liability arising out of any infringements of patents or other rights of third parties.

Mindray intends to maintain the contents of this manual as confidential information. Disclosure of the information in this manual in any manner whatsoever without the written permission of Mindray is strictly forbidden. Release, amendment, reproduction, distribution, rent, adaption and translation of this manual in any manner whatsoever without the written permission of Mindray is strictly forbidden

and MINDRAY are the registered trademarks or trademarks owned by Mindray in China and other countries. All other trademarks that appear in this manual are used only for editorial purposes without the intention of improperly using them. They are the property of their respective owners.

Contents of this manual are subject to changes without prior notice.

For this Operation Manual, the issued Date is August 2005 (Version: 2.5).

© 2002-2005 Shenzhen Mindray Bio-Medical Electronics Co., Ltd. All rights reserved.

CE 0123

Manufacturer’s Responsibility

All information contained in this manual is believed to be correct. Mindray shall not be liable for errors contained herein nor for incidental or consequential damages in connection with the furnishing, performance, or use of this manual.

Mindray is responsible for safety, reliability and performance of this product only in the condition that:

- all installation operations, expansions, changes, modifications and repairs of this product are conducted by Mindray authorized personnel; and,
- the electrical installation of the relevant room complies with the applicable national and local requirements; and,
- this product is operated under strict observance of this manual.

Warranty

This warranty is exclusive and is in lieu of all other warrantyies, expressed or implied, including warranties of merchantability or fitness for any particular purpose.

Exemptions

Mindray's obligation or liability under this warranty does not include any transportation or other charges or liability for direct, indirect or consequential damages or delay resulting from the improper use or application of the product or the use of parts or accessories not approved by Mindray or repairs by people other than Mindray authorized personnel.

This warranty shall not extend to

- any Mindray product which has been subjected to misuse, negligence or accident; or
- any Mindray product from which Mindray's original serial number tag or product identification markings have been altered or removed; or
- any product of any other manufacturer.

Return Policy

In the event that it becomes necessary to return a unit to Mindray, follow the instructions below.

1. Obtain a return authorization.

Contact the Mindray Service Department and obtain a Mindray Customer Service Authorization Number. The Mindray Customer Service Authorization Number must appear on the outside of the shipping container. Return shipments will not be accepted if the Mindray Customer Service Authorization Number is not clearly visible. Please provide the model number, serial number, and a brief description of the reason for return.

2. Freight policy

The customer is responsible for freight charges when this product is shipped to Mindray for service (including any relevant customs fees or other freight related charges).

3. Return address

Please send the part(s) or equipment to the address offered by Customer Service Department.

Contact Information

Manufacturer:	Shenzhen Mindray Bio-Medical Electronics Co., Ltd.
Address:	Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, Shenzhen 518057 P.R. China
Tel:	+86 755 26522479 +86 755 26582888
Fax:	+86 755 26582500 +86 755 26582501
Website:	www.mindray.com.cn

EC Representative:	Shanghai International Holding Corp. GmbH (Europe)
Address:	Eiffestraße 80, 20537 Hamburg Germany
Tel:	0049-40-2513175
Fax:	0049-40-255726

Equipment Symbols

	This symbol means 'BE CAREFUL '. Refer to the manual.
	This symbol indicates that the instrument is IEC 60601-1 Type CF equipment. The unit displaying this symbol contains an F-Type isolated (floating) patient applied part providing a high degree of protection against shock, and is suitable for use during defibrillation.
	Equipotential grounding system.
	Protective earth ground.
	Power On/Off
	This mark means that this device is fully in conformance with the Council Directive Concerning Medical Devices 93/42/EEC. The number adjacent to the CE marking (0123) is the number of the EU-notified body that certified meeting the requirements of Annex II of the Directive.
	The following definition of the WEEE label applies to EU member states only. This symbol indicates that this product should not be treated as household waste. By ensuring that this product is disposed of correctly, you will help prevent bringing potential negative consequences to the environment and human health. For more detailed information with regard to returning and recycling this product, please consult the distributor from whom you purchased it. * For system products, this label may be attached to the main unit only.
	NOTE: Points to be noted.
	CAUTION: Points to be noted to avoid damage to the equipment.
	WARNING: Points to be noted to avoid injury to the patient and the operator.

FOR YOUR NOTES

Contents

Chapter 1 Introduction 1-1
1.1 General Information 1-3
1.2 Screen Display 1-5
1.3 Button Functions 1-8
1.4 Interfaces 1-10
1.5 Batteries 1-13
Chapter 2 Getting Started 2-1
2.1 Open the Package and Check 2-1
2.2 Connect the Power Cables 2-1
2.3 Power on the Monitor 2-2
2.4 Connect Patient Sensors 2-2
2.5 Check the Recorder 2-3
Chapter 3 System Menu 3-1
3.1 Patient Information Setup 3-2
3.2 Default Setup 3-3
3.3 System Setup 3-5
3.4 Selection Setup 3-11
3.5 Monitor Version 3-12
3.6 Drug Calculation 3-13
3.7 Maintenance 3-13
3.8 DEMO Function 3-18
Chapter 4 Face Select 4-1
4.1 Select Operating Screen 4-1
4.2 Standard Screen 4-1
4.3 Trend Screen 4-2
4.4 oxyCRG Screen 4-3
4.5 Viewbed Screen 4-4
Chapter 5 Alarm 5-1
5.1 Alarm Modes 5-1
5.2 Alarm verification during power on 5-4
5.3 Alarm Cause 5-4
5.4 SILENCE and PAUSE 5-5
5.5 Parameter Alarm 5-6
5.6 When an Alarm Occurs 5-7
Chapter 6 Freeze 6-1
6.1 General 6-1
6.2 Enter/Exit Freeze Status 6-1
6.3 Frozen Menu 6-2
6.4 Reviewing Frozen Waveform 6-2

6.5 Recording Frozen Waveform 6-3
Chapter 7 Recording 7-1
7.1 General Information on Recording 7-1
7.2 Recording Type 7-1
7.3 Recording Startup 7-4
7.4 Recorder Operations and Status Messages 7-5
Chapter 8 Trend and Event 8-1
8.1 Trend Graph 8-1
8.2 Trend Table 8-3
8.3 NIBP Recall 8-4
8.4 Alarm Event Recall 8-5
Chapter 9 Drug Calculation and Titration Table 9-1
9.1 Drug Calculation 9-1
9.2 Titration Table 9-3
Chapter 10 Patient Safety 10-1
Chapter 11 Care / Cleaning 11-1
11.1 System Check 11-1
11.2 General Cleaning 11-2
11.3 Cleaning Agents 11-2
11.4 Sterilization 11-3
11.5 Disinfection 11-3
Chapter 12 ECG/RESP Monitoring 12-1
12.1 What Is ECG Monitoring 12-1
12.2 Precautions during ECG Monitoring 12-1
12.3 Monitoring Procedure 12-2
12.4 ECG Screen Hot Keys 12-6
12.5 ECG Menu 12-7
12.6 ECG Alarm Information and Prompt 12-11
12.7 ST Segment Monitoring (optional) 12-12
12.8 Arr. Monitoring (optional) 12-16
12.9 Measuring RESP 12-20
12.10 Maintenance and Cleaning 12-24
Chapter 13 SpO2 Monitoring 13-1
13.1 PART 1 (MASIMO SpO2 board configuration) 13-1
13.1.1 Precautions 13-3
13.1.2 Monitoring Procedure 13-5
13.1.3 Sensors and Accessories: 13-8
13.1.4 Alarm Description and Prompt 13-10
13.1.5 Masimo Information 13-13
13.2 PART 2 (MINDRAY SpO2 board configuration) 13-14
13.2.1 What is SpO2 Monitoring 13-14
13.2.2 Precautions during SpO2/Pulse Monitoring 13-15
13.2.3 Monitoring Procedure 13-16
13.2.4 Limitations for Measurement 13-18

13.2.5 SpO2 Menu 13-18
13.2.6 Alarm Description and Prompt..... 13-21
13.2.7 Maintenance and Cleaning..... 13-22
Chapter 14 NIBP Monitoring 14-1
14.1 Introduction..... 14-1
14.2 NIBP Monitoring 14-1
14.3 NIBP SETUP menu 14-6
14.4 NIBP Alarm Message 14-9
14.5 Maintenance and Cleaning..... 14-11
Chapter 15 TEMP Monitoring..... 15-1
15.1 TEMP Monitoring..... 15-1
15.2 TEMP SETUP Menu..... 15-2
15.3 TEMP Alarm message..... 15-2
15.4 Care and Cleaning 15-3
Chapter 16 IBP Monitoring..... 16-1
16.1 Introduction..... 16-1
16.2 Precautions during IBP Monitoring..... 16-1
16.3 Monitoring Procedure 16-2
16.4 IBP Menu..... 16-3
16.5 Alarm Information and Prompts..... 16-9
16.6 Maintenance and Cleaning..... 16-11
16.7 ICP Transducer ICT/B （Optional Accessory） 16-12
Chapter 17 Accessories..... 17-1
17.1 ECG Accessories 17-1
17.2 SpO2 Accessories 17-2
17.3 NIBP Accessories..... 17-3
17.4 TEMP Accessories 17-3
17.5 IBP Accessories..... 17-4
Appendix A EC Declaration of Conformance..... A-1
Appendix B Product Specification B-1
Appendix C EMC C-1
Appendix D System Alarm Prompt..... D-1

FOR YOUR NOTES

Chapter 1 Introduction

- For an overall introduction to the monitor, please refer to **General Information**.
- For various messages displayed on the screen, please refer to **Screen Display**.
- For basic operating instructions, please refer to **Button Function.**
- For allocation of interface sockets, please refer to **Interfaces**.
- For important facts to be noted during the battery recharging procedure, please refer to **Built-in Battery**.
- For safety precautions of the monitor, please refer to **Patient Safety**.

⚠ **Warning** ⚠

PM-8000 Portable Multi-Parameter Patient Monitor is intended for clinical monitoring application with operation only granted to appropriate medical staff.

⚠ **Warning** ⚠

Monitor can only monitoring one patient at a time.

⚠ **Warning** ⚠

There could be hazard of electrical shock by opening the monitor casing. All servicing and future upgrading to this equipment must be carried out by personnel trained and authorized by Mindray.

⚠ **Warning** ⚠

Possible explosion hazard if used in the presence of flammable anesthetics or other flammable substance in combination with air, oxygen-enriched environments, or nitrous oxide.

.

⚠ **Warning** ⚠

You must verify if the device and accessories can function safely and normally before use.

⚠ **Warning** ⚠

You must customize the alarm setups according to individual patient situation and make sure that alarm sound can be activated when alarm occurs.

⚠ **Warning** ⚠

Do not use cellular phone in the vicinity of this device. High level electromagnetic radiation emitted from such devices may greatly affect the monitor performance.

⚠ **Warning** ⚠

Do not touch the patient, table, or the device during defibrillation.

⚠ **Warning** ⚠

Devices connected to the monitor shall form an equipotential system (protectively earthed).

⚠ **Warning** ⚠

When used with Electro-surgery equipment, you (doctor or nurse) must give top priority to the patient safety.

⚠ **Warning** ⚠

Do not place the monitor or external power supply in any position that might cause it to fall on the patient. Do not lift the monitor by the power supply cord or patient cable, use only the handle on the monitor.

⚠ **Warning** ⚠

Consult IEC-601-1-1 for system interconnection guidance. The specific requirements for system interconnection are dependent upon the device connected to the monitor and the relative locations of each device from the patient, and the relative location of the connected device to the medically used room containing the monitor. In all circumstance the monitor must be connected to a grounded AC power supply. The monitor is referred to as an IEC 601/F device in the summary of situations table contained in IEC 601-1-1.

⚠ **Warning** ⚠

Dispose of the packaging material, observing the applicable waste control regulations and keeping it out of children's reach.

⚠ **Warning** ⚠

This equipment is accord with the standard CISPR11 (EN55011) class A.

⚠ **Warning** ⚠

Grounding:

Connect the monitor only to a three-wire, grounded, hospital-grade receptacle. The three-conductor plug must be inserted into a properly wired three-wire receptacle; if a three-wire receptacle is not available, a qualified electrician must install one in accordance with the governing electrical code.

Do not under any circumstances remove the grounding conductor from the power plug.

Do not use extension cords or adapters of any type. The power cord and plug must be intact and undamaged.

If there is any doubt about the integrity of the protective earth conductor arrangement, operate the monitor on internal battery power until the AC power supply protective conductor is fully functional.

⚠ **Note** ⚠

The software was developed per IEC601-1-4. The possibility of hazards arising from errors in the software program is minimized.

⚠ **Caution** ⚠

At the end of its service life, the product described in this manual, as well as its accessories, must be disposed of in compliance with the guidelines regulation the disposal of such products. If you have questions concerning disposal of the product, please contact MINDRAY or its representatives.

If you have any doubt to the grounding layout and its performance, you must use the built-in battery to power the monitor.

1.1 General Information

Environment:

Temperature

Working	0 ~ 40 °C
Transport and Storage	-20 ~ 60 °C

Humidity

Working	15% ~ 95 %
Transport and Storage	10% ~ 95 %

Altitude

Working	-500 ~ 4,600m (-1,600 to 15,000ft)
Transport and Storage	-500 ~ 13,100m (-1,600 to 43,000ft)

Power Supply

100/240 (V) AC, 50/60 (Hz)
Pmax = 100 VA
FUSE T 1.6A

Contraindications: None

General instruction:

PM-8000 is a Portable Patient Monitor that has abundant monitoring functions and is used for the clinical monitoring of adult, pediatric and neonate. In addition, the user may select the different parameter configuration according to different requirements.

PM-8000 can be connected to the central monitoring system via the Mindray network so as to form a network monitoring system.

PM-8000 (Figure 1-1) can monitor vital signals as ECG, Respiratory Rate, SpO2, NIBP, TEMP,

IBP. It integrates parameter measuring modules, display and recorder in one device, featuring in compactness, lightweight and portability. Replaceable built-in battery facilitates transportation of patient. Large high-resolution display provides clear view of 5 waveforms and full monitoring parameters.

The power switch (POWER) of the monitor is on the upper left part of the front panel (② in the figure below). The POWER LED (③ in the figure below) is used to indicate the AC Mains condition. This LED is illuminated when AC Mains is connected. BATT LED (④ in the figure below) is on the left side of POWER LED, used to indicate the battery condition. This LED is off when no battery is loaded. After the battery is loaded, this BATT. LED will be illuminated if AC Mains is connected, or lighted off when AC Mains is disconnected and at the same time the monitor is not powered on, or flashes when though AC Mains is disconnected but the monitor is powered on. The alarm indicator (ALARM) is on the upper part of the front panel (① in the figure below), which will flash as soon as any alarm event happens. Sensor and probe sockets are on the left side of the monitor while the recorder is on the right side. Other sockets and power connector is on the rear panel of the monitor.

PM-8000 is a user-friendly device with operations conducted by a few buttons on the front panel (⑤in Figure 1-1) and a rotary knob (⑥in Figure 1-1). Refer to 1.3 Button Functions for details.

Figure 1-1 PM-8000 Portable Patient Monitor

The visible LEDs are CLASS 1 LED PRODUCT according with EN 60825-1 A11 Oct 1996.

PM-8000 Portable Patient Monitor performs monitoring of:

ECG	Heart Rate (HR)
	2-channel ECG waveforms
	Arrhythmia and S-T segment analysis（optional）
RESP	Respiratory Rate (RR)
	Respiration Waveform

SpO2	Oxygen Saturation (SpO2), Pulse Rate (PR)
	SpO2 Plethysmogram
NIBP	Systolic Pressure (NS), Diastolic Pressure (ND), Mean Pressure (NM)
TEMP	Channel-1 Temperature (T1), Channel-2 Temperature (T2), Temperature Difference between two channels (TD)
IBP	IBP SYS, DIA, MAP IBP waveform

PM-8000 provides extensive functions as visual & audible alarm, storage and report printout for trend data, NIBP measurements, and alarm events, oxyCRG, viewbed, and drug dose calculation function is provided either.

1.2 Screen Display

The display of PM-8000 parameter monitor is a color LCD, which can display the collected patient parameters, waveforms, alarm information as well as bed number, time and monitor status, etc.

The screen is divided into three areas(Figure 1-2): Information area①④; waveform area②; parameter area③.



Figure 1-2 PM-8000 Main Display

Information Area

The Message Area is at the top part of the screen, displaying the current status of both the monitor and the patient.

- Patient information include:

BED NO	Bed numbers of all patients under monitoring
Patient type	Three options: Adult, Pediatric, Neonate
“01-01-2000”	Current date
“13：51：32”	Current date and time
Male	Patient sex, Male or Female
ZHANG SHAN	Patient name This item will display blank if the operator does not input patient name

Other information in the Message Area will appear and disappear together with the reported status. According to the content, the information is divided into:

■ Prompt information, reporting the current status of the monitor or sensor/probe, which always appears to the right of the system time. When this information appears, it will cover patient sex and name.

■ flag for alarm PAUSE. Press “SILENCE” button once (less than 1 second) to mute all alarm sounds are muted for the time being and the flag appears at the same time. Press the button again to terminate the PAUSE status. The duration for PAUSE status can be 1 minute, 2 minutes or 3 minutes.

■ flag for alarm SILENCE. Press “SILENCE” button once (more than 1 second) to manually mute the alarm sound and this flag appears at the same time. The SILENCE status terminates when you discharge the status or new alarm occurs.

■ flag for Alarm Volume Off. It appears indicating that you have closed the alarm sound permanently. This status terminates when you discharges the status.

⚠ **Note** ⚠

If **symbol appears, the system will no longer give audible alarm sound. You must be very careful in using this function. Two ways can be used to discharge this status. One is set the alarm volume to an option other than OFF in the USER MAITAIN menu. The other method is to press SILENCE button to make the flag turn to** **. And then press SILENCE again and the system will restore the normal alarm status.**

■ Parameter alarm information is displayed always in the upper right corner of the screen.

■ When the waveforms on the screen are frozen, the FREEZE prompt will appear in the bottom part of the screen.

Waveform / Menu Area

The waveform area can maximally display 5 waveforms. The displaying order of the waveforms on the screen can be adjusted. For the maximum configuration, the waveforms provided by the system for selection are: 2 ECG waveforms, SpO2 waveform, IBP waveforms,

RESP waveform.

All the waveforms in the system are listed out in the “WAVEFORM SETUP” menu. The user may select the waveform to be displayed and adjust their displaying positions. The specific method is illustrated in the part: Tracing Waveforms Selection.

The name of the waveform is displayed on the upper left part of the waveform. The user may choose ECG lead based on the requirements. The gain of the channel and the filter way are also displayed on each ECG waveform. A 1mV scale bar is also displayed to the right side of ECG waveform. The IBP waveform scale can also be selected according to the actual requirement. Its range is described in the part: Measure IBP. In the IBP waveform area, the waveform scale is displayed. The three dotted lines for each IBP waveform form up to down represent respectively the upper limit scale, reference scale and lower limit scale. The values of these three scales can be set. The specific method is given in the part: Measure IBP.

When menu is wanted during screen operation, the menu always occupies the fixed position in the middle part of the waveform area, therefore part of waveform can not be viewed temporarily. After exiting the menu, the system will restores the original screen.

The user may set up the rate to refresh the waveform. The method to adjust the refreshing rate of each waveform is discussed in the setup description of each parameter.

Parameter Area

The parameter area lies to the right side of the waveform area, whose position basically corresponds to the waveform. The parameters displayed in the parameter area include:

ECG

- heart rate or pulse rate (unit: beats/minute)
- The ST analyzing result of channel 1 and 2: ST1, ST2 (unit: mV)
- PVCs（unit: times/minute）

NIBP

- From left to right, there are Systolic pressure, Mean pressure and Diastolic pressure（unit: mmHg or kPa）

SpO_2

- SpO_2（unit: %）
- Pulse Rate（unit: beats/minute）（When “BOTH” item is selected）

IBP

- Blood Pressure: Systolic, Mean, and Diastolic values are displayed from left to right.（unit: mmHg or kPa）.

RESP

- Respiration Rate（unit: times/minute）

TEMP

- Temperature of channel 1 and 2: T1, T2 and the difference between them TD. (unit: ℃ or ℉）

Alarm lamp and alarm status:

In normal status: the alarm lamp is not on.

When alarm exists, the alarm lamp flashes or lights on. The color of the lamp corresponds to

the alarm level. Refer to related chapter: Alarm.

For the details of alarm information and prompt information, refer to the related content of each parameter in related chapter.

⚠Warning⚠

Always verify the audible and visual (LED) alarms when PM-8000 powers on.

1.3 Button Functions

All the operations to PM-8000 are through the buttons and a knob at the bottom of the screen. The names of the buttons are above them. They are (from left to right, Figure 1-3):



Figure 1-3 PM-8000 Buttons and Knob

- POWER（Figure 1-3 ①）

Press to turn on/off the monitor.

- FREEZE（Figure 1-3 ②）

Press this button and the system will access the FREEZE status. In this status the user may review the waveform of 40 seconds. Also, the frozen waveform can be printed out. In the FREEZE status, press this button again to discharge the FREEZE status. For detailed information, refer to related chapter: Freeze.

- SILENCE（Figure 1-3 ③）

Push this button to suspend alarm for maximum 3 minutes (with 1 minute, 2 minutes and 3 minutes selectable). In Alarm PAUSE status, a symbol appears in the Message Area. Push this button for more than 1 second to mute all kinds of sounds (including alarm sound, heart beat, pulse tone, key sound). At the same time, a symbol appears in the Message Area. Push this button again to restore all kinds of sounds and the symbol

appears from the screen.

⚠ **Note** ⚠

If new alarm occurs in Alarm Pause/Silence status, the system will discharge Pause/Silence status automatically. For specific rules, see Chapter Alarm.

⚠ **Note** ⚠

The system will begin to give alarm information again once there exist alarm-triggering event. Nevertheless, remember pushing SILENCE button can permanently shut off audible alarm sound of ECG LEAD OFF and SPO2 SENSOR OFF alarms.

- REC/STOP（Figure 1-3 ④）

Press to start a real time recording. The recording time is set in REC TIME of RECORD SETUP submenu. Press during recording to stop the recording. For detailed information, refer to related chapter.

- NIBP（Figure 1-3 ⑤）

Press to inflate the cuff to start a blood pressure measurement. When measuring, press to cancel the measurement and deflate the cuff.

- Rotary knob（Figure 1-3 ⑥）

The user may use the rotary knob to select the menu item and modify the setup. It can be rotated clockwise or counter-clockwise and pressed like other buttons. The user may use the knob to realize the operations on the screen and in the system menu and parameter menu.

Method to use the knob to operate on the screen:

The rectangular mark on the screen that moves with the rotation of the knob is called “cursor”. Operation can be performed at any position at which the cursor can stay.

When the cursor is in the waveform area, the user may immediately modify the current setup. When the cursor is in the parameter area, the user may open the setup menu of the corresponding parameter module so as to set up the menu items of the module.

Operating method:

- Move the cursor to the item where the operation is wanted
- Press the knob
- One of the following four situations may appear:
  1. The cursor with background color may become into the frame without background color, which implies that the content in the frame can change with the rotation of the knob.
  2. Menu or measuring window may appear on the screen, or the original menu is replaced by the new menu.
  3. A check mark “ √ ” appears at the position, indicating that the item is confirmed.
  4. The system immediately executes a certain function.

1.4 Interfaces

For the convenience of operation, the different kinds of interfaces are in different parts of the monitor.

At the right side is the recorder's paper inlet cover as shown in Figure 1-4.

Figure 1-4 Right Side

At the left side are the connectors to patient cables and the sensors, as shown in Figure 1-5.

① Socket for Spo2 Sensor
② Socket for IBP transducer
③ Socket for TEMP probe
④ Socket for ECG cable
⑤ Socket for NIBP cuff



Figure 1-5 Left Side

This symbol means "BE CAREFUL". Refer to the manual.

Indicates that the instrument is IEC 60601-1 Type CF equipment. The unit displaying this symbol contains an F-Type isolated (floating) patient applied part providing a high degree of protection against sh

Figure 1-6 Rear Panel

On the rear panel are the following sockets, shown in Figure 1-6

- Power Supply: 100 ~ 240 (VAC), 50/60 (Hz). (Socket ⑤)
- VGA MONITOR: (Socket ③)

  Monitor interface for external standard VGA color monitor.

  Working mode:800 × 600, 16 color, APA mode.

Signal:	analog R G B	0.7 Vpp / 750 ohm
	Hor. / Vert.	TTL pos. / Neg.
Interface	D-sub 15 pin	
	Pin 1.	Red Video
	Pin 2.	Green Video
	Pin 3.	Blue Video
	Pin 4.	Ground
	Pin 5.	NC
	Pin 6.	Red Ground
	Pin 7.	Green Ground
	Pin 8.	Blue Ground
	Pin 9.	NC
	Pin 10.	Ground
	Pin 11.	NC
	Pin 12.	NC
	Pin 13	Horizontal Sync.
	Pin 14.	Vertical Sync.

Pin 15. NC

Appliance:

1) Install the VGA monitor in the same room with the patient but keep away from the patient for more than 1.5m. The monitor is intended to be used as an assistant monitoring device.
2) Plug and insert the connection cable while the VGA monitor is in power off status.
3) Power on at the same time, or power on the PM-8000 patient monitor after VGA.
4) Adjust brightness and contrast properly.

■

(Socket ④)

Equipotential grounding terminal for connection with the hospital's grounding system.

■ AUX OUTPUT (socket ②)

This port is used for both Analog Output and NURSE CALL.

The user could select the function of this port in "NURSE CALL SETUP" menu of "USER MAINTAIN" menu. Refer to the section about "USER MAINTAIN" menu to know the detailed information.

ANALOG OUTPUT: connected to oscillograph and pen recorder. BNC Jack.

NURSE CALL: connected to the CALL system of the hospital by using dedicated NURSE CALL cable.

⚠ **Note** ⚠

The output terminal of NURSE CALL cable has two leads in free status (ie., no distinction between positive or negative). Before use, the service engineer from MINDRAY or equipment engineer of the hospital must first install the accompanying connectors according to the real situation of the CALL system of the hospital.

■ Network Interfaces (Socket ①): Standard RJ45 Socket.

■ Fixing hole of supporter(Socket ⑥).

⚠ **Note** ⚠

Monitor must be connected with specific network equipment such as Harb during using net function.

⚠ **Warning** ⚠

Accessory equipment connected to the analog and digital interfaces must be certified according to the respective IEC standards (e.g. IEC 60950 for data processing equipment and IEC 60601-1 for medical equipment). Furthermore all configurations shall comply with the valid version of the system standard IEC 60601-1-1. Everybody who connects additional equipment to the signal input part or signal output part configures a medical system, and is therefore responsible that the system complies with the requirements of the valid version of the system standard IEC 60601-1-1. If in doubt, consult the technical service department or your local representative.

1.5 Batteries

This monitor is designed to operate run battery power when during transport or whenever the power supply is interrupted. The battery is charged automatically when the monitor is connected to AC power, no matter the monitor is powered on or not.

The battery symbol displayed on the main screen tells the status of the battery.

- The battery is installed in the battery slot. The solid part indicates its capacity.
- No battery is installed in the battery slot.

Besides, the battery indicator also indicates the status of the battery.

- ON: The battery is being charged or the battery is fully charged.
- OFF: No battery is installed. If the battery is installed but the monitor is not connected to AC power and not turned on, the indicator will also be off.
- Flashes: The monitor is powered by the internal battery.

The capacity of the internal battery is limited. When the battery capacity is too low, a high level alarm is triggered and the “Battery two low” message is given in the technical alarms area. At this moment, the AC power shall be applied to the monitor.

⚠ **Note** ⚠

Remove the battery before transport, or if the monitor is not likely to be used for an extended period of time.

⚠ **Warning** ⚠

Keep the battery out of the reach of children.

Use only the battery specified by the manufacturer.

1.5.1 Battery Maintenance

Conditioning a Battery

A battery should be conditioned before it is used for the first time. A battery conditioning cycle is one uninterrupted charge of the battery, followed by an uninterrupted discharge of the

battery. Batteries should be conditioned regularly to maintain their useful life. Condition a battery once when it is used or stored for two months, or when its run time becomes noticeably shorter.

To condition a battery, follow this procedure:

1. Disconnect the monitor from the patient and stop all monitoring or measuring.
2. Insert the battery in need of conditioning in the battery slot of the monitor, and leave the other slot empty if your minitor has two slots.
3. Apply AC power to the monitor and allow the battery to charge uninterrupted for 10 hours.
4. Remove AC power and allow the monitor to run from the battery until it shuts off.
5. Apply AC power again to the monitor and allow the battery to charge uninterrupted for 10 hours.
6. This battery is now conditioned and the monitor can be returned to service.

Checking a Battery

The performance of a rechargeable battery may deteriorate over time. To check the performance of a battery, follow this procedure:

1. Disconnect the monitor from the patient and stop all monitoring or measuring.
2. Apply AC power to the monitor and allow the battery to charge uninterrupted for 10 hours.
3. Remove AC power and allow the monitor to run from the battery until it shuts off.
4. The operating time of battery reflects its performance directly.

If your monitor has two battery slots, you can check two batteries at the same time. Please replace the battery or contact with the maintenance personnel if its operating time is significantly lower than the specified time.

⚠ Note ⚠

Life expectancy of a battery depends on how frequent and how long it is used. For a properly maintained and stored lead-acid or lithium ion battery, its life expectancy is about 2 or 3 years respectively. For more aggressive use models, life expectancy can be less. We recommend replacing lead acid batteries every 2 years and lithium ion batteries every 3 years.

The battery might be damaged or malfunctioned if its operating time is too short after being fully charged. The operating time depends on the configuration and operation. For example, measuring NIBP more frequently will also shorten the operating time.

1.5.2 Battery Recycling

When a battery has visual signs of damage, or no longer holds a charge, it should be replaced. Remove the old battery from the monitor and recycle it properly. To dispose of the batteries, follow local laws for proper disposal.

⚠ **Warning** ⚠

Do not disassemble batteries, or dispose of them in fire, or cause them to short circuit. They may ignite, explode, leak or heat up, causing personal injury.

FOR YOUR NOTES

Chapter 2 Getting Started

- Open the package and check
- Connect the power cables
- Power on the monitor
- Connect patient sensors
- Check the recorder

⚠ **Note** ⚠

To ensure that the monitor works properly, please read Chapter Patient Safety, and follow the steps before using the monitor.

2.1 Open the Package and Check

Open the package and take out the monitor and accessories carefully. Keep the package for possible future transportation or storage. Check the components according to the packing list.

- Check for any mechanical damage.
- Check all the cables, modules and accessories.

If there is any problem, contact the distributor immediately.

2.2 Connect the Power Cables

Connection procedure of the AC power line:

- Make sure the AC power supply complies with following specification: 100 ~ 240 VAC, 50/60 Hz.
- Apply the power line provided with the monitor. Plug the power line to INPUT interface of the monitor(Socket ⑤ in Figure 1-6). Connect the other end of the power line to a grounded 3-phase power output.

⚠ **Note** ⚠

Connect the power line to the jack special for hospital usage.

Mindray does not provide MULTIPLE PORTABLE SOCKET-OUTLETS. IF use it, please do not place it on the floor. Mindray advises that every one monitor uses one MULTIPLE PORTABLE SOCKET-OUTLETS.

- Connect to the ground line if necessary. Refer to **Chapter Patient Safety** for details.

⚠ **Note** ⚠

Make sure that the POWER lamp now lights. If it does not light, check your local power supply. If the problem still exists, contact the local Customer Service Center.

The battery needs to be charged after transportation or storage. If the power supply is not properly connected before turning on the monitor, it may not work properly because of insufficient power. Connect the power supply to charge the battery.

2.3 Power on the Monitor

Press **POWER**(① in Figure 1-1) to power on the monitor. Then a beep will be heard and at the same time the indicator will flash twice in yellow and red. After 10 seconds or so, the system will enter monitoring screen after self-test, and you can perform normal monitoring now.

During self-test, the software version will display.

⚠ **Note** ⚠

If the monitor finds any fatal error during self-test, it will alarm.

Check all the functions that may be used to monitor and make sure that the monitor is in good status.

The battery must be recharged to the full electricity after each use to ensure adequate electricity reserve.

The interval between twice pressing of POWER should be more than 1 minute.

⚠ **Warning** ⚠

If any sign of damage is detected, or the monitor displays some error messages, do not use it on any patient. Contact biomedical engineer in the hospital or Mindray Customer Service Center immediately.

2.4 Connect Patient Sensors

Connect all the necessary patient sensors between the monitor and the patient.

⚠ **Note** ⚠

For information on correct connection, refer to related chapter 12-16.

2.5 Check the Recorder

If your monitor is equipped with a recorder, open the recorder door to check if paper is properly installed in the output slot. If no paper present, refer to **Chapter Recording** for details.

Chapter 3 System Menu

- Patient Setup
- Default Setup
- System Setup
- Selection Setup
- Monitor Version
- Drug Calculation
- Maintenance
- Demo Function

PM-8000 Portable Multi-Parameter Patient Monitor features flexible configurations. You can customize monitoring content, waveform sweep speed, sound volume, and output content.
Turn knob to select the MENU hot key on the lower right part of the screen to call up the "SYSTEM MENU" menu. You can perform following operations in this menu.



Figure 3-1 SYSTEM MENU

In this chapter, the submenus will be described one by one, except the "TREND GRAPH", "TREND TABLE", "NIBP RECALL" and "ALARM RECALL", which will be described in **Chapter 8 Trend and Event**.

3.1 Patient Setup

⚠ **Note** ⚠

To clear current patient data, refer to New Patient for details.

Pick the [PATIENT SETUP] item in the “SYSTEM MENU” to call up the following menu.



Figure 3-2 PATIENT SETUP

You can setup following patient information:

DEPT.	Department in which the patient receives treatment.
PAT NO	Patient No.
BED NO	Patient bed number (Range: 1-100)
DOCTOR	Name of the doctor.
NAME	Patient name (Valid characters: A-Z, 0-9 and space bar; Max. length: 12 characters)
SEX	Patient gender (Available options: "F" for Female, "M" for Male)
PAT TYPE	Patient type (Available options: ADU, PED, and NEO)
ADMIT	Hospitalization starting date (format: year\month\ day)
BIRTH	Patient date of birth (format: year\month\day)
HT. (cm/in)	Patient height (turning the knob with the increase/decrease of 0.5 cm/inch each time)The other HT. unit in the other menus accord with the unit which

you choosed here.

WT. (kg/lb) Patient weight (turning the knob with the increase/decrease of 0.5 kg/lb each time)The other WT. unit in the other menus accord with the unit which you choosed here.

BLOOD Patient blood type (Pick A, B, O, AB, or N. "N" represents unknown blood type)

NEW PATIENT Admission of new patient

Also in this menu, you may select the [NEW PATIENT] item to access the “CONFIRM TO UPDATE PATIENT” dialog box as shown below, in which you can decide whether to monitor a new patient.



Figure 3-3 Confirm To Update Patient Menu

Pick [YES] to delete all information of the patient being currently monitored and exit the menu. Pick [NO] to give up updating the patient and the system will keep the information of the current patient and exit the menu.

⚠ **Note** ⚠

If you select [YES], the system will delete all information of the patient being currently monitored.

3.2 Default Setup

⚠ **Note** ⚠

After selecting any item in this sub-menu, the selected item will replace the current setup of the system and accordingly become the system default configuration.



Figure 3-4 DEFAULT Menu

In this sub-menu, you can select both the factory default and the user-defined default. Also in this sub-menu, you can save the current system configuration as the user-defined default configuration. But at this time, the system will automatically save all the setups in the parameter menu, ECG gain and filter way as the user-defined default configuration according to the patient type. Also, the dialog box as shown below will pop up.



Figure 3-5 CONFIRM DEFAULT CONFIG

⚠ **Note** ⚠

After selecting any item in the DEFAULT menu and exiting the box, the "CONFIRM DEFAULT CONFIG" Dialog box will pop up, in which you can select [YES] to confirm your selection or [NO] to give up your selection.

⚠**Warning**⚠

All configurations in the system will be replaced by "default configurations".

3.3 System Setup

Select the [system setup] item in the [system menu]:

Figure 3-6 System setup

In the [System setup] menu , users can setup the following items.

3.3.1 Face select

Select “FACE SELECT” item in “SYSTEM SETUP” menu to access “FACE SELECT” dialog box as shown below, in which four selections are available: STANDARD SCREEN, TREND SCREEN, oxyCRG SCREEN and VIEWBED SCREEN. Only one selection can be chosen for each time.



Figure 3-7 FACE SELECT

3.3.2 Alarm setup

The system provides three levels of alarm volume. You can select any of them as per the clinical requirement. The procedures are:

Select the [ALARM SETUP] item in the “SYSTEM SETUP” sub-menu of the “SYSTEM SETUP” menu. The menu as shown below will pop up, in which you can set up the alarm volume and other alarm information. For detailed information, refer to Chapter **Alarm.**



Figure 3-8 Alarm Setup

Pick “ALARM VOL” item, turn the knob to set the volume. The options are from “10” to “1”. “10” indicates the maximum volume while “1” the minimum. If your machine does not have "Multilevel Volume" function, only three options are available for volume, “3”, “2” and “1”.

3.3.3 Time Setup

Select the [TIME SETUP] item in the “SYSTEM SETUP” menu. The menu as shown below will pop up. System time is in the format of year, month, day, hour, minute and second. Use cursor to highlight the item that you want to modify and turn the knob to select time. Then select [EXIT].

⚠ **Note** ⚠

You shall set up the system time upon turning on the monitor (if you need to set up the system time); otherwise, when you review the content with time information, the system may not display the correct time.



Figure 3-9 System Time Setup

When this monitor is linked to the Central monitor system, its system time will keep consistent with that of the Central monitor system. Method to adjust time:Once link is successfully established, the Central monitor system will send its current time to the monitor. The monitor will automatically adjust its system time accordingly. Besides, the Central monitor system will keep on sending its current time to the monitor once per hour to maintain consistent time between them. However, the monitor will not adjust its time if it is different from the Central monitor system only in second. Please note that if you are setting up the system time when link is just established successfully, the monitor will immediately close the setup menu of system time. The setup button of system time in the system setup menu is disabled when the monitor is linked to the Central monitor system. That means you cannot open the setup menu of system time. (If the Central monitor system has no this function, you can skip over this paragraph.)

3.3.4 Recorder setup

Select the [RECORD] in the “SYSTEM SETUP” menu to call up the following menu:



Figure 3-10 Recorder Setup

In this menu, the user can set up to output two waveforms. The waveforms that can be selected include:

ECG1-ECG6	there are six ECG waveforms in multi-leads display (If no ECG waveform is currently displayed on the screen, this item cannot be picked).
SPO2	SpO2 Plethysmogram.
IBP	The IBP waveform on the screen (If no IBP waveform is currently displayed on the screen, this item cannot be picked).
RESP	RESP waveform (If no RESP waveform is currently displayed on the screen, this item cannot be picked,).
OFF	No display for this waveform.

- RT REC TIME this item has two options, CONTINUAL and 8s. “CONTINUAL” means once pushing the “REC/STOP” button on the recorder panel or the monitor panel, the recorder will continuously print out the waveform or parameter until this button is pushed again.
- TIMING REC TIME used to set up the time interval between two recordings. 10 selections are available: “OFF, 10min, 20min, 30min, 40min, 50min, 1hour, 2hours, 3hours and 4hours”. The system will start the recording process according to the selected time interval. The recording time is always 8 seconds.

⚠ **Note** ⚠

RT REC TIME takes priority over TIMING REC TIME.

- REC RATE: this item has two options, 25.0 and 50.0 mm/s.
- REC GRID: used to decide output format: OFF is without grid, and ON is with grid.
- CLEAR REC TASK: used to clear the alarm event that has been generated and is waiting for recording out.

⚠ **Note** ⚠

If two same waveforms are selected, the system will automatically change one of the waveform to a different one.

3.3.5 Analog

The monitor can output an analog waveform, whose time delay is less than 30ms. The output terminal is on the rear panel.

Select “ANALOG” item in “SYSTEM SETUP” menu to call up the ANALOG menu. The first item is for setting up On/Off of the switch of the analog output. The second item is for selecting the waveform name to be output.

Select “EXIT” item to return to the previous menu.



Figure 3-11 ANALOG

⚠ **Note** ⚠

In the USER MAINTAIN menu, If the AUX OUTPUT item being selected with NURSE CALL, the AUX OUTPUT port will be used to realize NURSE CALL function while “ANALOG OUT” function is switched off at the same time.

3.3.6 Module Setup

Select the [MODULE SETUP] item in the “SYSTEM SETUP” menu to call up the following menu:



Figure 3-12 Module Setup

You can choose the parameters to be monitored in this menu. This can avoid the interference from the parameters that need not attention.

3.3.7 Tracing Waveforms Selection

Select the [TRACE SETUP] in the “SYSTEM SETUP” menu to call up the following menu.



Figure 3-13 Tracing Waveforms Selection

You can define the traces displayed on the screen in this menu. The waveforms available for selection are those whose modules have been selected in “MODULE SETUP” menu.

This user can only decide the display sequence of the waveforms on the screen. Select the “WAVE SEQUENCE” item in the menu to access the sub-menu of the same name as shown in the figure below.



Figure 3-14 Wave sequence

3.3.8 Event Setup

The monitor has four types of events. You can specify their representations by yourself.

Select the [MARK EVENT] item in the “SYSTEM SETUP” to call up the following menu:

Figure 3-15 MARK EVENT Menu

How to mark the event: Use the rotary knob to select one from event A, B, C and D. The @ symbol will appear in the frame of the event being selected. Once making a wrong selection, you can push the knob on the event again to give up the selection. Select [EXIT] to exit the menu and consequently the selection will come into effect.

Event function has following significance:
To classify the records into different categories, such as those having influence on patients and those having influence on parameter monitoring including dose taking, injection, therapy status. Event will be displayed on the trend graph/table in order to assist the analysis on the patient parameters when the event happens.

3.4 Selection Setup

Select the [SELECTION] item in the “SYSTEM MENU” to call up the following menu.



Figure 3-16 Selection Setup

KEY VOL:

Pick “KEY VOL” item in “SELECTION” menu, turn the knob to set the volume. The options are from “10” to “0”. “10” indicates the maximum volume while “0” indicates close the volume. If your machine does not have "Multilevel Volume" function, only four options are available for volume, “3”, “2”,”1” and “OFF”.

HELP:

The system provides On-line Help to menu operations. You can choose any help information as per your need. The method is:

Select the [SELECTION] item in the “SYSTEM MENU” to access the “SELECTION” sub-menu, in which you can highlight the [HELP] item and turn the knob to select “ON” or “OFF”. When it is “ON”, you can browse the on-line help information. When it is “OFF”, the system will turn off the on-line help function.

SCAN TYPE:

The system can display all waveforms about monitored patient on the screen either in “Refresh” or “Scroll” way. The method is:

Select “SELECTION” item in “SYSTEM MENU” to access “SELECTION” sub-menu, in which there is the item “SCAN TYPE”. The user may decide the way to display the waveform by choosing either “REFRESH” or “SCROLL”.

ALM LIMIT:

The system can display the alarm limits. You can choose this function as per your need. The method is:

Select the [SELECTION] in the “SYSTEM MENU” to call up the “SELECTION” menu. You can set the “ALM LIMIT” switch to “ON” or “OFF”.

3.5 Monitor Version

Select the [VERSION] item in the “SYSTEM MENU” to know the software version of the monitor.

Figure 3-17 Monitor Version

Select the [DEVICE CONFIG LIST] to know the configuration of the monitor.



Figure 3-18 Device Configuration List

3.6 Drug Calculation

You can use the drug calculation and titration table function to calculate the concentration of 15 kinds of drugs. For detailed information, please refer to Chapter: ***Drug Calculation and Titration Table***.

3.7 Maintenance

Select the [MAINTAIN] item in the “SYSTEM MENU” to call up the “ENTER MAINTAIN PASSWORD” dialog box as shown below, in which you can enter password and then customize maintenance settings.

ENTER MAINTAIN PASSWORD
USER KEY:
FACTORY KEY:
CONFIRM
CONFIRM
STATUS >>
A B C D E F G H I J K L M N O P Q R S T U
V W X Y Z 0 1 2 3 4 5 6 7 8 9 DEL OK
EXIT

Figure 3-19 Enter Maintain Password

You cannot execute factory maintenance function, which is only available for the service engineers of our company.

Input the password into the “ENTER MAINTAIN PASSWORD” box and press [CONFIRM], the “USER MAINTAIN” menu will pop up, in which you can set up following items.



Figure 3-20 User Maintain

For the [LANGUAGE] language, you can select the screen language you need.

For the [AUX OUTPUT] item, there are two options available:

- ANALOG OUT: if being selected, the AUX OUTPUT port will be used to realize “ANANOG OUT” function while NURSE CALL function is switched off at the same time. And visually the “NURSE CALL SETUP” item in “USER MAINTAIN” menu will become gray indicating that the function is disabled.
- NURSE CALL: if being selected, the AUX OUTPUT port will be used to realize NURSE CALL function while “ANALOG OUT” function is switched off at the same time.

For the [LEAD NAMING] item, you can select “AHA” or “EURO”. To know the difference between these two styles, refer to Chapter: ECG/RESP Monitoring.

For the [ALM SOUND] item, you can set the alarm volume to “ON” or “OFF”.
For the [NET TYPE] item, two selections are available: HYPER III and CMS.
For the [LOCAL NET NO] item, it refers to the net No.

⚠ **Warning** ⚠

When the alarm volume is set to “OFF”, you will not hear the alarm sound if new alarm occurs. Therefore, you must be very careful in using this selection.

If setting the alarm volume to “OFF” when the system is in Silence or Pause status, the system will automatically discharge Silence or Pause status.

If you select “Silence” or “Pause” when the alarm volume is set to “OFF”, the system will restore the alarm volume before the alarm volume is set to “OFF” and enter Silence or Pause status.

⚠ **Note** ⚠

After the alarm volume is set to OFF, a **symbol will appear in the Technical Alarm Area.**

⚠ **Note** ⚠

Setting Alarm Volume to “OFF” is valid only when the monitor is turned on for this time. After turning on the monitor next time, this setup will restore its value of the previous time when the system is turned on.

COLOR SELF-DEFINE: is used by the user to define the color of the waveform displayed on the screen. Five colors can be chosen from: green, cyan, red, yellow and white.



Figure 3-21 Color Self-define

NURSE CALL SETUP: If the NURSE CALL item in AUX OUTPUT being selected, the NURSE CALL SETUP submenu will be available.



Figure 3-22 Nurse Call Setup

- SIGNAL DURATION: “PULSE” and “CONTINUUM” two types of signals are available. Selecting “PULSE” indicates that the NURSE CALL is the pulse signal of 1s duration; selecting “CONTINUUM” indicates that the NURSE CALL signal is synchronous with the alarm signal designated in the triggering condition.
- SIGNAL TYPE: “NORMAL OPEN” or “NORMAL CLOSE”.

  NORMAL OPEN: select this item when the CALL system of the hospital is set to “NORMAL OPEN”;

  NORMAL CLOSE: select this item when the CALL system of the hospital is set to “NORMAL CLOSE”.
- ALM LEV and ALM TYPE: after NURSE CALL function is activated, the monitor provides the following combination options of alarm level and alarm type for the user to choose in order to trigger NURSE CALL signal. “ALM LEV” provides three combination options, i.e., NURSE CALL signal will be triggered when it is “HIGH” alarm, “MED” alarm or “LOW”

alarm. “ALM TYPE” provides two combination options, i.e., NURSE CALL signal will be triggered when it is “TECH” alarm or “PHYS” alarm.

⚠ Warning ⚠

When no option in “ALM TYPE” is selected, the NURSE CALL signal will not be triggered in whatever condition.

⚠ Warning ⚠

When in ALARM SILENCE/PAUSE status, the monitor will automatically switch off NURSE CALL signal; after discharging ALARM SILENCE/PAUSE status, the monitor will automatically return to the status before ALARM SILENCE/PAUSE is activated.

If the user select “CLOSE” in the ALARM SOUND item of the “USER MAINTAIN” menu, it will does not affect the function of NURSE CALL.

⚠ Warning ⚠

The nurse call feature should not be used as the primary source of alarm notification. The audible and visual alarms of the monitor, used in conjunction with clinical signs and symptoms, are the primary source for notifying medical personnel that an alarm condition exists.

3.8 DEMO function

Select the [DEMO] item in the “SYSTEM MENU” to call up the “ENTER DEMO PASSWORD”. After entering the password, the system enters DEMO status.

The purpose of waveform demonstration is only to demonstrate the machine performance, and for training purpose. In clinical application, this function is not forbidden because the DEMO will mislead the medical staff to treat the DEMO waveform and parameter as the actual data of the patient, which may result in the delay of treatment or mistreatment. Therefore before entering this menu, you shall enter password.



Figure 3-23 Input Demo Key

Chapter 4 Face Select

This monitor has four different operating screens, which are “Standard Screen”, “Trend Screen”, “oxyCRG Screen”, and “Viewbed Screen”. When required, you can select different operating screens for necessary information. Let’s probe into these four operating screens one by one.

4.1 Select Operating Screen

In the “SYSTEM MENU”, select the “FACE SELECT” option in the “SYSTEM SETUP” menu to call up the dialog box as shown in the figure below. There are four options in this dialog, which are “STANDARD SCREEN”, “TREND SCREEN”, “oxyCRG SCREEN” and “VIEWBED SCREEN”. Only one item can be selected at one time.



Figure 4-1 FACE SELECT

4.2 Standard Screen

In the “FACE SELECT” menu, Select the “STANDARD SCREEN” option to enter the Standard Screen. The Standard Screen displays to us the parameters in the Parameter area and the waveforms being monitored. This screen is the basic operating screen of the monitor.

Figure 4-2 STANDARD SCREEN

4.3 Trend Screen

- Enter TREND SCREEN

In the “FACE SELECT” menu, select the “TREND SCREEN” option to enter the Trend Screen.

Figure 4-3 TREND SCREEN

- Position of trend graph

Trend graph is located to the right of the corresponding waveform in the Waveform area. Its color is the same as that of the corresponding parameter.

- Trend length

Dynamic trend length is 2 hours. On the trend graph, the scale of the right end of the X-axis is 0 hour while the left end is 2-hour.

- Select trend parameter

If multiple parameters are located at the same position on the trend graph, by selecting the corresponding hot key of a parameter on the trend graph, you can have the trend graph of this parameter displayed on the screen. For example, in ECG trend graph, you can select hot keys such as HR, ST or PVCs, then the system will display their corresponding trend graphs respectively.

- Close trend screen

In the “FACE SELECT” menu, select options of other operating screens to close the Trend Screen.

4.4 oxyCRG Screen

- Enter oxyCRG screen

In the “FACE SELECT” menu, select the “oxyCRG SCREEN” to enter the oxyCRG Screen.

Figure 4-4 oxyCRG SCREEN

- Trend graph of oxyCRG screen

Located at the lower part of the screen, oxyCRG screen consists of three trends: HR

Trend, SpO2 Trend and RR Trend or Compressed Resp. Waveform.

- Select OxyCRG trend length

Three are three hot keys at the bottom part of the oxyCRG Screen, which are 4MIN/2MIN/1MIN, RR/RESP WAVE, and REC.

By using hot keys for trend time, you may select to display trend graphs of three different lengths, i.e., 1 min, 2 min and 4 min.

- Select RR trend or Compressed Resp. Waveform

By using the hot keys for RR/RESP WAVE, you may select either RR trend graph or compressed Resp. Wave. They occupy the same position. Therefore, if select “RR”, the position displays the dynamic trend of RR. If select “RESP WAVE”, the position displays the compressed Resp. Wave.

- Record

Select the “REC” hot key in the “OxyCRG Screen”, you may use the recorder to output the three waveforms in the oxyCRG at the same time.

- Close OxyCRG

In the FACE SELECT menu, select options of other operating screens to close the OxyCRG Screen.

4.5 Viewbed Screen

If another monitor is connected on the same LAN of this monitor, you can use this monitor to view any measured waveform and information about all measured parameters from another monitor.

- Enter Viewbed Screen

Select the “VIEWBED SCREEN” option in the “FACE SELECT” menu. Viewbed Screen window occupies the space of the bottom four waveforms.

Figure 4-5 VIEWBED SCREEN

- Hot key of Viewbed

There are two hot keys in the Viewbed Screen: Select Bed Number and Select Waveform.

The hot key of Select Bed Number displays the bed numbers and patient names of other monitors currently connected on the LAN. You can select a monitor to be monitored according to the patient name and bed number. If at this time no other monitors are connected on the same LAN of this monitor, the hot key of Bed Number will therefore display “N/A”. After you use this hot key to select a monitor to be viewed, the system will toggle to the display of the selected monitor for your view. The selected waveform is one of those listed in the hot key of Select Waveform.

The hot key of Select Waveform is used to select a waveform generated by the monitor being viewed. If the hot key of Select Waveform displays “N/A”, it indicates that the bedside monitor being viewed has no waveforms. You can use this hot key to select and therefore view different waveforms of the monitor being viewed.

- Alarm indicator of Viewbed

On the upper right side of the Viewbed Screen, there is an Alarm Indicator used to tell the alarm status of the monitor being viewed. The activity of this alarm indicator is identical with that of the alarm lamp on the panel of the monitor being viewed. That is to say, if the monitor being viewed occurs medium/low level alarm, this alarm indicator illuminates yellow; if it occurs high level alarm, this alarm indicator illuminates red. If the monitor being viewed has no alarm or the alarm is screened, the icon for this alarm indicator will not be displayed.

- Parameter area of Viewbed Screen

Under the hot key of Select Bed Number is the Parameter area, in which parameters

of all monitors being viewed are displayed.

- Waveform area of Viewbed Screen

Under the hot key of Select Waveform is the Waveform area. The Sweep manner (refreshing or scrolling) of the waveform is identical with that of this monitor. The feature description of the displayed waveform is given above the waveform. Sweep speed is also identical with that set up for the same waveform on this monitor.

- Technical Information area

Technical Information area is to the right of patient name in Viewbed Screen. This area displays related technical information to Viewbed, such as due to network failure or network too busy, Viewbed is disabled.

- Close Viewbed Screen

In the FACE SELECT menu, select options of other operating screens to close the Viewbed Screen.

- Rules for automatically selecting monitor to be viewed and waveform

When you turn on the monitor or enter Viewbed Screen, the system will automatically select a networked bedside monitor and a waveform of this monitor for you to view. If the monitor being currently viewed is disconnected, the viewed monitor will automatically close, clear displays of all alarms, parameters and waveforms. However in this situation, the Viewbed Screen still displays. If you want to view another monitor, you must select again through using hot keys.

If a measure module of the viewed monitor is plugged out or closed, its corresponding waveform will disappear and the waveform in the Waveform area will not be refreshed. Instead this Waveform area will display empty. At this time, if you want to view other waveforms of this monitor, you need to select again.

Chapter 5 Alarm

This chapter gives general information about the alarm and corresponding remedies.

Alarm setup and prompt messages are provided in respective parameter setup sections.

⚠ **Warning** ⚠

When PM-8000 is powered on, the system may verify the audio and visual alarm function.

Upon turning on the monitor, a "Dang" will be heard and at the same time the indicator will flash twice in yellow and red. This is used to verify the audio and visual alarm function of the system. Therefore, the user should be carefully observe the status. If the audio and visual alarm function is not normal, it indicates that the monitor cannot be used to monitor a patient. Please contact Mindray Company or service center.

5.1 Alarm Modes

5.1.1 Alarm Level

Each alarm, either technical or physiological, has its own level. For alarm of higher level, when it occurs, the system will give prompt in a more alert way. Some alarm's level can be set by the user via software. Others can not be changed once defined by the system. Alarms in PM-8000 are divided into three levels, that is, high, medium and low.

High-level alarm indicates the patient's life is in danger or the monitor under using has serious problem in technical respect. It is the most serious alarm.
Medium-level alarm means serious warning.
Low-level alarm is a general warning.

Alarms are classified into three categories, which are physiological alarm, technical alarm and general alarm. Physiological alarm refer to those alarms triggered by patient's physiological situation which could be considered dangerous to his or her life, such as heart rate (HR) exceeding alarm limit (parameter alarms). Technical alarm refer to system failure which can make certain monitoring process technically impossible or make monitoring result unbelievable. Technical alarm is also called System Error Message. General alarm belongs to those situations that can not be categorized into these two cases but still need to pay some attention.

PM-8000 has preset the alarm level for the parameters. You can also modify the alarm level using the method described in this chapter.

Alarm level of the System Error Message (technical alarm) is pre-set in the system.
All technical alarm level and general alarm level, some of the physiological alarm level are pre-set in the system and can not be changed by user.

5.1.2 Alarm Modes

When alarm occurs, PM-8000 may raise the user's attention in at least three ways, which are audio prompt, visual prompt and description. Audio and visual prompt is given by TFT display device, the speaker on the display device and the alarm indicator. Description is displayed on the screen. Physiological alarm is displayed in the Physiological Alarm area. Most of technical alarms are displayed in the Technical Alarm area. Technical alarms related to NIBP measurement are displayed in the NIBP Technical Alarm area at the bottom of NIBP parameter area.

⚠ **Note** ⚠
The Physiological Alarm area is on the upper right part of the screen. The Technical Alarm area is to the left side of the Physiological Alarm area.

⚠ **Note** ⚠
The concrete presentation of each alarm prompt is related to the alarm level.

Alarm prompt of the parameter exceeding the alarm limit.
When physiological alarm of the monitored parameter exceeds the alarm limit, besides using the above-mentioned three ways to give the alarm prompt, the monitor also gives alarm by making the monitored parameter flash in the frequency of 1Hz. If at this time the upper and lower limits of the parameter are displayed, they will flash in the same frequency (1Hz).

Screen Display

When an alarm occurs, the parameter triggering the alarm flashes. "*" signal appears on the screen indicating the occurrence of alarm. Red "***" indicates high-level alarm, yellow "**" indicates medium-level alarm, and yellow "*" indicates low-level alarm. Technical alarm will not prompts "*" signal.

Lamp light

The high/medium/low-level alarms are indicated by the system in following different visual ways:

Alarm level	Visual prompt
High	Alarm indicator flashes in red with high frequency.
Medium	Alarm indicator flashes in yellow with low frequency.
Low	Alarm indicator lights on in yellow.

Alarm Sound

The high/medium/low-level alarms are indicated by the system in following different audio ways:

Alarm level	Audio prompt
High	Mode is “DO-DO-DO------DO-DO, DO-DO-DO------DO-DO”, which is triggered once every 8 seconds.
Medium	Mode is “DO-DO-DO”, which is triggered once every 24 seconds.
Low	Mode is “DO-”, which is triggered once every 24 seconds.

⚠ **Note** ⚠

When alarms of different levels occur at the same time, the monitor prompts the one of the highest level.

Alarm Setup

The setup of the alarms can be realized in the alarm menu.

Press the “ALARM SETUP” button on the SYSTEM SETUP menu to call up “ALARM SETUP” menu (default menu) as shown below. In the “ALM SEL” item, the user may set up the information about common alarm setup (represented by “COMMON ALM SETUP”) and the alarm setup of each parameter.



Figure 5-1 ALARM SETUP

■ **COMMON ALM SETUP**

Select “COMMON ALM SETUP” selection in “ALM SEL” item. This operation may call up the dialog box as the default one.

- ALARM VOL: The options are from “10” to “1”. “10” indicates the maximum volume while “1” the minimum. If your machine does not have "Multilevel Volume" function, only three options are available for volume, “3”, “2” and “1”.
- ALM REC TIME: which has three selections: 8S, 16S, 32S.
- ALM PAUSE TIME: refers to the alarm suspension time span, which has three selections: 1MIN, 2MIN, 3MIN.
- PARA ALM TYPE: which has two selections: LATCH, UNLATCH. LATCH refers to the situation once alarm occurs, the system will alarm always until the intervention of the operator (press SILENCE on the panel). UNLATCH refers to the situation that once the alarm condition is discharged, the alarm will disappear automatically.

■ **Alarm setup of each parameter**

In “ALARM SETUP” menu select “ALM SEL” item to set up the alarm information of following parameters. They are HR, ST, PVC, SPO2, NIBP, IBP (1, 2), RESP, TEMP. For example:

- Method to set up alarm information of HR:

Step 1: Select “HR ALM SETUP” in “ALM SEL” item to call up the dialog box “ALARM SETUP” for HR only.

Step 2: Five items are available for the user to set up, which are HR ALM (on/off of the alarm switch), ALM LEV(alarm level), ALM REC(alarm recording switch), ALM HI (higher limit of HR alarm), ALM LO (lower limit of HR alarm). When use the knob to select each item and press the knob, a pull-down list appears for the user to choose his desired selection.

The method for setting the alarm information of other parameters is the same as HR.

5.2 Alarm verification during power on

During PM-8000 power on, audible and visual alarm capability will be tested by the system.

Every time when PM-8000 powers on, alarm beeps “DO-”, and the LED indicator on the display device flashes. If no beeps heard or no alarm indicator flashing viewed, do not use this device to monitor any patient, and notify Customer Service Center.

5.3 Alarm Cause

Alarm occurs when:

1.Physiological alarm is evoked;

2. Alarm for error of the system (technical alarm) is evoked;

3. General alert occurs.

- **A. Conditions that activate the parameter alarms:**

When the measurement value exceeds the alarm limit and the alarm is set “ON”. Alarm will not activate if the alarm is set “OFF”.

- **B. Conditions that activate the system alarms (technical alarm):**

Upon the system error, the monitor prompts alarm immediately and proceeds corresponding remedy, stops all monitoring and eliminates the final results in order to avoid faulted treatment. If more than one error occurs, they will be displayed by turns.

- **C. General alert**

In some circumstances, alerts will behave as physiological alarm but in normal sense, we don’t regard them as real patient health related items.

5.4 SILENCE and PAUSE

- SILENCE/CLOSE

Push the SILENCE button on the panel for more than 1 second, the system will shut off all sounds. Push the SILENCE button again, the system can exit the SILENCE status and restore the PAUSE status and accordingly suspend the alarm as per the previously defined time duration. Push the SILENCE button for the third time, the system will exit the PAUSE status and restore the normal alarm status by giving the alarm sound again. When the system is in the SILENCE status, any new alarm will terminate the SILENCE status and make the system restore the normal alarm status.

⚠ **Note** ⚠

When the **symbol appears indicating the alarm sound is shut off and accordingly the system will not give alarm sound. Therefore, you must be very careful in using this function. There are two methods to terminate this status. One is to set the alarm volume to “ON” in the MAINTAIN menu. The other method is to push the SILENCE button shortly to make the** **symbol become ; push the SILENCE button again and the system will restore the normal alarm status again.**

- PAUSE

Push the SILENCE button on the panel shortly, the system will shut off all alarm sound and visual prompt as well as description of physiological alarm, and enter the PAUSE status. The countdown of PAUSE status is displayed in the Physiological Alarm area, in which area the symbol is also displayed.

The time duration of the PAUSE status can be set to 1min, 2min or 3min. You can select in

the [ALM PAUSE TIME] item in the “SYSTEM MENU\SYSTEM SETUP\ALARM SETUP”.
After pushing the SILENCE button again, the system will restore the normal status. Besides, the occurrence of any new technical alarm will also terminate the PAUSE status and let the system restore the normal status. The symbol disappears, too.

⚠ **Note** ⚠

After the system goes back to the normal status, the existence of alarm depends on whether the alarm condition is complied with. After pushing the SILENCE button, the system will permanently shut off the alarm sound for LEAD OFF/SENSOR OFF alarm.

5.5 Parameter Alarm

The setup for parameter alarms is in their menus. In the menu for a specific parameter, you can check and set the alarm limit, alarm status. The setup is isolated from each other.

When a parameter alarm is off, a symbol “ ” displays near the parameter. If the alarms are turned off individually, they must be turned on individually.

For the parameters whose alarm is set to ON, the alarm will be triggered when at least one of them exceeds alarm limit. The following actions take place:

1. Alarm message displays on the screen as described in alarm mode;
2. The monitor beeps in its corresponding alarm class and volume;
3. Alarm lamp flashes;
4. Store all parameter values during the alarm and 4,8 or 16 second waveform prior to and after alarm.
5. If alarm recording is on, the recorder starts alarm recording. For further information on alarm recording, please refer to Chapter Recording.

5.6 When an Alarm Occurs

⚠ **Note** ⚠

When an alarm occurs, you should always check the patient's condition first.

The alarm message appears at the top of the screen on the right side. It is needed to identify the alarm and act appropriately, according to the cause of the alarm.

1. Check the patient's condition.
2. Identify the cause of the alarm.
3. Silence the alarm, if necessary.
4. When cause of alarm has been over, check that the alarm is working properly.

You will find the alarm messages for the individual parameter in their appropriate parameter chapters of this manual.

FOR YOUR NOTES

Chapter 6 Freeze

- General
- Freeze & Unfreeze
- Review & Record Frozen Waveforms

6.1 General

When monitoring a patient, you may freeze the waveforms of interest so as to view them carefully. Generally you can review maximally 40 seconds of a frozen waveform. If required, you may also use recorder to print out a frozen waveform. The Freeze function of this monitor has following features:

- Freeze status can be activated on any operating screen;
- At the same time of entering the Freeze status, the system exits all other operating menus. Besides, the system freezes all waveforms in the Waveform area of the Basic Screen, or Full-lead ECG waveforms and the extra waveform (if available) on the Full-lead ECG screen. Nevertheless the Parameter area refreshes normally.
- In the Freeze status, it does not affect the display and refresh of the Trend Graph area on the trend screen, the display and refresh of oxyCRG on the Dynamic Refresh screen, or the display and refresh of the Viewbed window on the Viewbed screen.
- The frozen waveforms can be reviewed or recorded.

6.2 Enter/Exit Freeze Status

Enter Freeze Status

In the Non-Freeze status, press the “FREEZE” button on the front panel of the monitor to let the system exit the Menu being currently displayed (if available), then enter the Freeze status and display the popup “FROZEN” menu. In the Freeze status, except Viewbed waveforms, all other waveforms are frozen. In other words, the system will no longer refresh all other waveforms.

Exit Freeze Status

In the Freeze status, executing any of the following operations will command the system to exit the Freeze status:

- Select the “EXIT” option on the “FROZEN” menu;
- Press the “FREEZE” button on the front panel again;

- Press the non-immediate-to-execute button (such as a button once pressed, a menu will pop up for you to further select an option)on the front panel and system buttons of MAIN and MENU;
- Execute any operation that may trigger the adjustment of the screen or display of a new menu.

After exiting the Freeze status, the system will discharge the Freeze status, clear screen waveforms and resume to display real-time waveforms. In the Screen Refresh mode, the system begins scanning waveforms from the extreme left one. In the Screen Scroll mode, the system begins displaying and scrolling waveforms from the extreme right one.

6.3 FROZEN Menu

Press the “FREEZE” button on the button module, the FROZEN menu will appear on the bottom part of the screen. At the same time, the system enters the Freeze status.



Figure 6-1 FROZEN menu

- WAVE 1： used to select the first frozen waveform to record. The pull-down list of this item gives you the names of all frozen waveforms displayed on the screen.
- WAVE 2： used to select the second frozen waveform to record. The pull-down list of this item gives you the names of all waveforms displayed on the screen.
- RECALL：used to review frozen waveforms.
- REC: after selected, the system begins recording the frozen waveforms selected in “WAVE 1” and “WAVE 2”.
- EXIT： after pressed, the system closes the FROZEN menu and exits the Freeze status。

⚠ **Note** ⚠

Pressing the “FREEZE” button repeatedly in short time period may result in discontinuous waveforms on the screen.

6.4 Reviewing Frozen Waveform

By moving the waveform, you may review a waveform of 40 seconds before the moment when it is frozen. For a waveform less than 40 seconds, the remaining part is displayed as a straight line. Use the rotary snob on the front panel to move the cursor to the “RECALL” option on the FROZEN menu. Press the knob, the option displays “L-RIGHT”. By turning the knob left or right, frozen waveforms on the screen will move left or right correspondingly. There is an arrow indicating upward under the right side of the last waveform. There is also a time scale beside the arrow. “0S” is used to mark the moment when waveforms are frozen. With waveforms moving right, this time mark will in turn change into -1S, -2S, -3S… These time marks are applied to all waveforms on the screen.

6.5 Recording Frozen Waveform

In the Freeze status, you may output displayed frozen waveforms via the recorder. Maximum 2 waveforms can be output at one time. On the FROZEN menu, the pull-down lists of both “WAVE 1” and “WAVE 2” give you all names of frozen waveforms on the screen, from which you may select two. Select the “REC” option on the FROZEN menu to output parameters generated upon the freezing moment and the two selected frozen waveforms. If one of the two selected waveforms is closed or not available, only parameters and the other waveform are recorded. If these two selected waveforms are all closed or not available, only parameters are recorded. As for the function of recording frozen waveforms, you can only record the waveforms displayed upon the freezing moment. The recording time length is the same as the length of the waveform displayed on the screen. For example, if the speed of a waveform is relatively fast, then it needs shorter time to record it. When recording frozen waveforms, the system is still in the Freeze status. After completion of recording, if required, you may select once more the waveform to be output and select “REC” option again to record the whole selected waveforms. If the recorder does not exist, selecting the “REC” option can only call out the prompt “Recorder does not exist” in the STATUS bar. For more detailed information about recording, please refer to the chapter of “Recording”.

FOR YOUR NOTES

Chapter 7 Recording

- General information on recording
- Instructions for configuring and recording
- Recording messages

7.1 General Information on Recording

A thermal dot matrices recorder with 48mm wide printout paper is used for PM-8000 Portable Patient Monitor.

Performance of the Recorder

- Waveform record is printed out at a rate of 25 or 50 mm/s.
- It can record up to 2 waveforms.
- Output with grid selectable.
- English / Chinese printout.
- The real time recording time and waveform are user-configurable.
- Auto recording interval is set by the user, the waveform is in accordance with the real time recording.
- The alarm recording waveform is automatically selected by the monitor.

7.2 Recording Type

PM-8000 provides several stripe recording types:

- Continuous real-time recording
- 8 second real-time recording
- Auto 8 second recording
- Alarm recording
- Waveform freeze recording
- Trend graph/table recording
- ARR events review recording
- Alarm event recording
- NIBP review recording
- Monitor information recording
- Drug calculation titration recording
- OxyCRG recording

Real-time Recording

Real-time recording starts as you press the REC/STOP button on the recorder.

The waveforms for continuous real-time recording and continuous 8 second recording are automatically set by the monitor (usually the first two waveforms displayed on the screen). You can also configure it through the menu. Refer to related section for details.
In RECORD menu, the user can choose two waveforms to be printed out. The User can setup one waveform off. Thus, the real time record will print out one waveform. If two waveforms are off, the real time record will print out measure parameters only.

⚠ **Note** ⚠
If certain recording is in process, and another parameter demands alarm recording, it will only be executed after the earlier recording is finished.

Auto recording

The monitor starts the recorder for 8 seconds according to interval time set in the “TIMING REC TIME” of the “RECORDER ” menu. Refer to **Chapter 3.5 Recorder Setup** for details.

Alarm Recording

Parameter Alarm

The monitor records waveforms 4, 8, or 16 seconds prior to and after the alarm (totally 8, 16 or 32 seconds) (which can be selected in System Menu). All parameter values during the alarm will also be recorded.。

When parameter alarm occurs, two recorded waveforms can be printed out.
In order to avoid repeated printout of alarm waveforms:

- If more than two parameter alarms are switched on and triggered simultaneously, the recorder will print out those of the highest level. If of the same alarm level, the latest alarm will be printed out.
- If an alarm occurs during the alarm of another parameter, it will be printed out after the current recording is finished.
- If many alarms occur at the same time, some of waveforms will be stored for printout in turn.

ST Segment Alarm

The monitor records 2-channel ECG waveforms 4, 8, or 16 seconds prior to and after the alarm (totally 8, 16, or 32 seconds) (which can be selected in the ECG SETUP menu). All parameter values during the alarm will also be recorded.

Arrhythmia Alarm

The monitor records 2-channel ECG waveforms 4 seconds prior to and after the alarm (totally 8 seconds). All measurement results during the alarm will also be recorded.

Freeze Waveform Recording

The monitor prints out the selected waveforms under the FREEZE mode. In this way you can snap the abnormal waveforms on the screen and record it.

Trend Graph / Table Recording

The monitor can print out the trend graph and table in the current TREND GRAPH or TREND TABLE window.

Arrhythmia Review Recording

The monitor can print out the alarm Arrhythmia event in the current ARR RECALL window.

Alarm Review Recording

The monitor can print out the alarm events include waves and parameters in the current ALARM RECALL window.

NIBP Review Recording

The monitor can print out all the NIBP review events in NIBP RECALL window.

Monitor Information

The monitor can print out messages in the current STATUS window.

Titration Table

The monitor can print out the messages in the current TITRATION window.

Notes on Recording

- Recording texts:
  Real time Report
  Periodic Report
  Para Alarm Report: XXX (name of the alarm parameter)
  Arrhythmia Report: XXX (Arrhythmia type)
  Freeze Wave Report
  Trend Graph
  Trend Table
  Para Alarm Review
  NIBP Test Review
  Status Report
  Titration Table
- Alarm parameters, alarm time and freeze time
- Patient bed number, name, sex, height, weight, date of birth, admission date
- Parameter name and value

- Recording time
- Waveform name
- Waveform scale (for ECG waveform)
- ECG lead, scale, filter mode, (if having ECG waveforms, it will be printed out within the first second or when changing the lead, gain and filter mode during real-time recording.)
- IBP scale (the first second of IBP waveform)
- CO2 scale (the first second of CO2 waveform)
- Date and time
- Company name

7.3 Recording Startup

You can start the recording in the following ways:

Continuous real-time recording	Press REC/STOP to start/stop the recording.
8 second real-time recording	Press REC/STOP to start recording. It will automatically stop in 8 seconds.
Auto recording	Record the two waveforms selected in RECORD menu according to the setup time interval in RECORD menu.
Alarm recording	When alarm recording is set ON, it automatically starts when alarm occurs.
Frozen waveform recording	---After accessing FREEZE menu, use knob to select two waveforms to be output. Then press REC button in the menu to print out the waveforms.



	If two waveforms are off, the measure parameters in frozen are printed out only.
Trend graph recording	Pick ”REC” button in the “TREND GRAPH” menu when viewing the trend graph to print out the currently displayed trend graph.
Trend table recording	Pick ”REC” button in the “TREND TABLE” menu when viewing the trend table to printout the currently displayed trend table.
Arrhythmia review recording	Access ARR RECALL window from ARR ANALYSIS of ECG SETUP menu and Pick ”WAVE” button to access “ARR WAVE RECALL” menu. Then press “REC” button to output the Arr. Waveform and related information currently displayed on the screen.
Alarm review recording	Access the “ALARM RECALL” window from “ALARM RECALL CONDITION” menu from “SYSTEM MENU” and pick “REC” button to print out the alarm review waveform

	and related information currently displayed in the “ALARM RECALL” window.
NIBP review recording	Access the “NIBP RECALL” window from “SYSTEM MENU” and pick “REC” button to print out the NIBP information currently displayed in the window.
Monitor information recording	Access the “ENTER MAINTAIN PASSWORD” menu from the “MAINTAIN” menu. Then pick the “STATUS” button to access the ‘’STATUS” window. Pick “REC” button to print out the status monitor information currently displayed in the window.
Titration table recording	Access the “DRUG CALC” menu from the “SYSTEM MENU” menu. Pick the “TITRATION” button in the menu to access the “TITRATION” window. Pick the “REC” button to print out the titration currently displayed in the window.
OxyCRG recording	In oxyCRG screen, pick the “REC” button to print out oxyCRG currently displayed in the window.

⚠ **Note** ⚠

You can press REC/STOP button on the recorder to stop the current recording process.

Access the “RECORD” menu from the “SYSTEM SETUP” menu. Then pick the “CLEAR REC TASK” button to stop all recording tasks.

7.4 Recorder Operations and Status Messages

Record Paper Requirement

Only standard 50 (+0/-1) mm thermosensitive record paper can be used, otherwise the recorder may not function, the recording quality may be poor, and the thermosensitive printhead may be damaged.

Function Properly

- When the recorder is working, the record paper goes out steadily. Do not pull the paper, or the recorder will be damaged.
- Do not operate the recorder without record paper.

Paper Out

When "RECORDER OUT OF PAPER" alarm is displayed, the recorder cannot start. Please insert record paper properly.

Inserting Paper

- Open the recorder catch.
- Pull down the switch on the left axis of the recorder.
- Insert a new roll of paper into the paper cassette, printing side facing the thermosensitive printhead.

- When the paper can be seen from the other side, pull it out. Ensure proper position and tidy margin.
- Pull back the switch on the left axis of the recorder.
- Give out the paper from the recorder outlet.
- Close the recorder catch.

⚠ **Note** ⚠

Be careful when inserting paper. Avoid damaging the thermosensitive printhead. Unless when inserting paper or shooting troubles, do not leave the recorder catch open.

Removing Paper Jam

When the recorder functions or sounds improperly, open the recorder catch to check for a paper jam. Removing the paper jam in the following way:

- Cut the record paper from the feeding edge.
- Pull up the switch on the left axis of the recorder.
- Pull the paper from below.
- Re-insert the paper.

Recorder Status Message (Technical Alarms)

Message	Cause	Alarm Level	Remedy
RECORDER HEAD HOT	The thermal terminal is too hot.	low	Stop operation
REC HEAD IN WRONG POS.	The thermal head is not in recording place.	low	Push down the switch on the left axis of the recorder.
RECORDER OUT OF PAPER	Record paper runs out.	low	Insert a new roll of record paper.
RECORDER COMM ERR	Operating status error	low	Reset the recorder.
RECORDER PAPER JAM	Recording continuously for more than 30m	low	Re-insert paper.
RECORDER INITIALIZING	The recorder is in initialization process.	low	Wait for the completion of initialization
TOO MANY REC TASKS	Too many alarm events take place simultaneously.	Low	Send recording order after a while.
RECORDER PAPER W.P.	The paper is in wrong position.	low	Insert the record paper again.
RECORDER BUSY	In the status of printing out	low	Wait for the completion of printing out
REC NOT AVAILABLE	Recorder stops working.	Low	Gives recording order after the recorder restores to the normal

			status or the failure is removed.
RECORDER VLT HIGH	The voltage of the recorder is too high.	Low	Stop recording until the recorder restores normal status.
RECORDER VLT LOW	The voltage of the recorder is too low.	Low	Stop recording until the recorder restores normal status.
RECORDER S. COMM ERR	Unrecoverable serial port communication error.	Low	Shut down the monitor and re-start it again.
RECORDER SELFTEST ERR	Possibly caused by the RAM, ROM, CPU or WATCHDOG.	Low	Reset the recorder.
RECORDER INIT ERR	Error occurs during initialization	low	Shutdown and re-start
RECORDER INIT ERR1	Error occurs during initialization	low	Shutdown and re-start
RECORDER INIT ERR2	Error occurs during initialization	low	Shutdown and re-start
RECORDER INIT ERR3	Error occurs during initialization	low	Shutdown and re-start
RECORDER INIT ERR4	Error occurs during initialization	low	Shutdown and re-start
RECORDER INIT ERR7	Error occurs during initialization	low	Shutdown and re-start
RECORDER INIT ERR8	Error occurs during initialization	low	Shutdown and re-start

If after shutdown and re-start, error still exists, contact out service engineers.

FOR YOUR NOTES

Chapter 8 Trend and Event

PM-8000 provides 72-hour trend data of all parameters, storage of 400 NIBP measurement results and 60 alarm events. This chapter gives detailed instruction for review of all data.

8.1 Trend Graph

- The latest 1-hour trend is displayed every 1 or 5 seconds;
- The latest 72-hour trend is displayed every 1, 5 or 10 minutes;

Pick "TREND GRAPH" in the SYSTEM MENU to call up the following menu:



Figure 8-1 TREND GRAPH Menu

The uppermost part is the name of the parameter, in which y-axis stands for value and x-axis time. "▼" Indicates the value of the parameter, which it points to, is below the x-axis, with corresponding time displayed beyond the trend graph. Other trends except NIBP trend are displayed as continuous curves. In NIBP trend graph, "▼" indicates systolic value, "▲" indicates diastolic value, and "*" indicates mean value.

To select trend graph of a specific parameter:

Pick PARA SELECT item (the first selection of the upper line) and select a requested parameter name by turning the knob.

To select 1-hour or 72-hour trend graph:

Pick RESOLUTION item (the latter selection of the upper line), choose 1 or 5 sec for 1-hour trend graph and 1, 5 or 10 min for 72-hour trend graph.

To view other trend curves:

When "" appears on the right part of the screen, pick "L-RIGHT" (the button at the extreme left of the lower line), turn the knob clockwise to view later trend curves. When "" appears on the left part of the screen, pick the same item, turn the knob counterclockwise to view earlier trend curve.

To change the display scale

Pick the “ZOOM” button in the lower line to adjust the y-axis scale and thus change the trend curve in proportion. The value beyond maximum value will be represented by the maximum value.

To obtain trend data of a specific time

The time to which the cursor points will change as the knob is turned. Parameter at this time is displayed below the x-axis. When "" appears on the right part of the screen, the trend graph pages down for later trend curve as the cursor moves here. When "" appears on the left part of the screen, the trend graph pages up for earlier trend curve as the cursor moves here.

To print out the trend curve

Press REC button to print out the trend curve of current selected parameter.

Mark event

If an event is marked A, B, C, or D, then the corresponding event type will display on the axis time of the trend graph. The event sign (A, B, C or D) is displayed in a frame.

Operation example

To view the NIBP trend graph of the last 1 hour:

Pick the MENU hot key lower right of the screen.

Pick TREND GRAPH item.

Pick the first item and switch to NIBP by turning the knob.

Adjust the second item to be 1 or 5 sec.

Pick the ZOOM button and turn the knob to view changes of the trend graph time and trend curve.

Stop at requested trend time section for careful review. Pick the ZOOM button to adjust the display scale if necessary.

For measurement result of a specific time, pick CURSOR to move the cursor to the point, corresponding time and value will display on above and below respectively.

For printout of trend graph, pick REC to start report printing of NIBP trend of this hour.

Pick EXIT to return to trend graph display.

8.2 Trend Table

■ The latest 72-trend table data can be displayed at every 1, 5, 10, 30, or 60 minutes.

Pick TREND TABLE in the SYSTEM MENU to call up the following menu:



Figure 8-2 TREND TABLE Menu

Time in response to each group of trend data is displayed at the leftmost list with date in bracket. Marked event corresponds to marking time.

NIBP trend data presents different specificity. A certain NIBP measuring time is displayed below the TEST AT item, as well as the measurement value. For more than one measurement in one time, it can display only one group, and mark a "*" on the MORE to indicate two and above measurement results.

To choose trend table of different resolution

Pick the leftmost item and change the time interval of trend data.

To view other trend data:

When "↑" appears on the upper part of the screen, pick UP-DOWN button and turn the knob clockwise to view later trend data. When "↓" appears on the lower part of the screen, pick the same item and turn the knob counterclockwise to view earlier trend data.

To obtain trend data of different parameter

Pick L-RIGHT to select one from the groups of parameters. A ">" by the rightmost item indicates following page available. And "<" by the leftmost item indicated previous page available.

To print out the trend data

Pick REC to print out the trend data of current displayed parameter.

Mark event

If an event is marked A, B, C, or D, the corresponding event type will display on the axis time of the trend table.

Operation example

To view a NIBP trend table:

- Pick MENU hot key lower right of the screen to access “SYSTEM MENU”.
- Pick TREND TABLE.
- Pick L-RIGHT and switch to NIBP by turning the knob.
- Pick the first item from the left and select requested time interval.
- Pick UP-DOWN and turn the knob to view NIBP trend data of different time.
- For printout of trend table, pick REC to start report printing of all trend data including NIBP of this time span.
- Pick EXIT to return to SYSTEM MENU.

8.3 NIBP Recall

PM-8000 can review the latest 400 NIBP measurement data.
Pick NIBP RECALL in the SYSTEM MENU to invoke the result and time of the latest 10 measurements, as shown in the figure below.



Figure 8-3 NIBP RECALL

Data is listed chronologically from the latest to the earliest. 10 measurements can be displayed in one screen. Pick UP-DOWN to view other trend curve up to 400 results. Pick REC to print out all measurement data of NIBP RECALL.

8.4 Alarm Event Recall

PM-8000 can display the latest 60 alarm events.

- Select “ALARM RECALL” in the SYSTEM MENU to access ALARM RECALL CONDITION menu as shown below.



Figure 8-4 ALARM RECALL CONDITION Menu

In this menu, the user may select the conditions for alarm review, including:

1. Start and End time of review:
   The user may select the start time of review in the item of START.
   Then the user may select the end time of review. Two selections are available: current time and the user-defined time.
   For user-defined end time, the user can use the knob to select.
2. ALARM RECALL EVENT
   In the pull-down list of ALARM RECALL EVENT, the user can select the parameter whose alarm events he wants to review. The selections include ALL(alarm events of all parameters), ECG, REST, SPO2, NIBP, IBP, TEMP, HR_H>180(the value of HR is higher than the upper alarm limit), HR_L<60(the value of HR is below the lower alarm limit), SPO2<90%, IBP_H>200mmHg, IBP_L<40mmHg, RR_H>40, RR_L<10, TEMP_H>40℃, TEMP_L<34℃.
   After setting up all the review conditions, press the “ALARM RECALL” button to access “ALARM RECALL” window.

- ALARM RECALL

The ALARM RECALL window is as shown below, in which following data are displayed:

① Time span (Format: month-day-year hour: minute- month-day-year hour: minute).
② Event type.
③ Serial number (Format: NO. xx of XX).
④ The value at the time of alarm. NIBP result is with time.

⑤ Two 8/16/32-second waveforms.



Figure 8-5 ALARM RECALL Menu

To view all waveforms during the alarming process

Pick L-RIGHT and turn the knob to view all 8/16/32-second waveforms stored.

To view other alarm events

Events of up to 60 are listed chronologically from the latest to the earliest. Pick UP-DOWN button and turn the knob to view later or earlier events.

Recording

Pick REC to print our all data and waveform of this event.

Chapter 9 Drug Calculation and Titration Table

PM-8000 Portable Patient Monitor provides Drug calculation and titration table display functions for fifteen drugs and outputs the content of titration table on the recorder.

9.1 Drug Calculation

The drug calculations that can be performed by the system are AMINOPHYLLINE, DOBUTAMINE, DOPAMINE, EPINEPHRINE, HEPARIN, ISUPREL, LIDOCAINE, NIPRIDE, NITROGLYCERIN and PITOCIN. Besides DRUG A, DRUG B, DRUG C, DRUG D and DRUG E are also provided to flexibly replace any of the drugs.

Select “DRUG CALC” in SYSTEM MENU, the following “DRUG CALC” display appears:



Figure 9-1 DRUG CALC

The following formulas are applied for dose calculation:

Concentrat = Amount / Volume

INF Rate =DOSE / Concentrat

Duration = Amount / Dose

Dose = Rate × Concentrat

Operating method:

In the Drug Calculation window, the operator should first select the name of the drug to be

calculated, and then confirm the patient weight. Afterwards, the operator should also enter other known values.

Turn the knob to select the value of the item to be calculated. Turn the knob to change the value. When it is the required value, press the knob to view the calculation result. Each item has its calculation range. If the result exceeds the range, display "---.--"

⚠ **Note** ⚠

For the drug calculation, the prerequisite is that the operator must first of all enter the patient weight and drug name. The system first gives a group of random initial values, which cannot be used by the operator as the calculation reference. Instead, he should enter a new group of values at the doctor's instruction.

⚠ **Note** ⚠

Each drug has its fixed unit or unit series. Operator must select the proper unit at the doctor's instruction. If the result exceeds the system-defined range, it will display "---".

⚠ **Note** ⚠

After entering a value, a conspicuous prompt will appear in the menu warning the operator to confirm the correctness of the entered value. The correct value is the guarantee for the reliability and safety of the calculated results.

⚠ **Note** ⚠

In neonate mode, Drip Rate and Drop Size items are disabled.

⚠ **Note** ⚠

For each entered value, the system will always give a dialog box asking for the user's confirmation. You must be careful when answering each box. The calculated result is reliable only after the entered value is confirmed to be correct.

Select the drug name：

Turn the knob to pick the DRUG NAME item in DRUG CALC menu. The user may select the drug name in the pull-down list, including AMINOPHYLLINE、DOBUTAMINE、DOPAMINE、EPINEPHRINE、HEPARIN、ISUPREL、LIDOCAINE、NIPRIDE、NITROGLYCERIN、PITOCIN、Drug A、Drug B、Drug C、Drug D and Drug E. Calculation for only one type can be generated each time.

NOTE：A、B、C、D、E are only codes for drugs instead of their real names. The units for these five drugs are fixed. The operator may select the appropriate units according to the convention of using these drugs. The rules for expressing the units are:

"mg" series units are fixedly used for drug A, B and C: g, mg, mcg.

"unit" series units are fixedly used for drug D: unit, k unit, m unit.

"mEq" is fixedly used for drug E.

Patient weight:

After accessing the DRUG CALC window, the operator should enter the patient weight into the first or the second item. The entered weight will be used as the independent data only for the calculation of drug concentration.

⚠ **Note** ⚠

This drug calculation function acts only as a calculator. That means the patient weight in Drug Calculation menu and the patient weight in Patient Information menu are independent from each other. Therefore if the Weight in Drug Calculation changes, the Weight in Patient Information does not change. In this way, we can say, the Drug Calculation menu is independent from other menus in the system. Any change of it will not affect other information about the patient being currently monitored.

9.2 Titration Table

Access titration table:

Select TITRATION item in DRUG CALC menu to enter titration table display.

Titration table display for drug is as following:



Figure 9-2 TITRATION

- Method to operate the titration table:

1. In the TITRATION table, turn the knob to pick BASIC item. Press and turn the knob to select either FLOW RATE or DOSE or DROP RATE.
2. Then turn the knob to pick STEP item. Press and turn the knob to select step. 1 ~ 10 are available for selection with the increment being 1.
3. Turn the knob to pick DOSE TYPE item. Press and turn the knob to select the unit in the

pull-down list.

4. Use UP-DOWN item in the table to view the data in previous or following pages.
5. Turn the knob to pick REC item. After pressing the knob, the recorder prints out the data displayed in the current titration table.
6. Turn the knob to pick EXIT to return to DRUG CALC menu.

Total amount, dose, volume, flow-rate, drop rate and patient weight and drug name are displayed on the top of the titration table. Meaning of each English identifier is:

AMOUNT：drug amount

VOLUME：liquid volume

DOSE/min：drug dose

FLOW RATE：flow rate

DROP RATE：drop rate

WEIGHT：patient weight

Chapter 10 Patient Safety

The PM-8000 Portable Patient Monitor is designed to comply with the International National Safety requirements for medical electrical equipment. This device has floating inputs and is protected against the effects of defibrillation and electrosurgery. If the correct electrodes are used and applied in accordance with the manufacturer instructions, the screen display will recover within 10 seconds after defibrillation.

This symbol indicates that the instrument is IEC 60601-1 Type CF equipment. The unit displaying this symbol contains an F-Type isolated (floating) patient applied part providing a high degree of protection against shock, and is suitable for use during defibrillation.

⚠ **Warning** ⚠

Do not touch the patient, bed or instrument during defibrillation.

Environment

Follow the instructions below to ensure a completely safe electrical installation. The environment where the PM-8000 Portable Patient Monitor will be used should be reasonably free from vibration, dust, corrosive or explosive gases, extremes of temperature, humidity, and so on. For a cabinet mounted installation, allow sufficient room at the front for operation and sufficient room at the rear for servicing with the cabinet access door open.

The PM-8000 Portable Patient Monitor operates within specifications at ambient temperatures between 0℃ and 40℃. Ambient temperatures that exceed these limits could affect the accuracy of the instrument and cause damage to the modules and circuits. Allow at least 2 inches (5cms) clearance around the instrument for proper air circulation.

Power Source Requirements

Refer to **chapter Production Specification.**

Grounding the PM-8000 Portable Patient Monitor

To protect the patient and hospital personnel, the cabinet of the PM-8000 Portable Patient Monitor must be grounded. Accordingly, the PM-8000 Portable Patient Monitor is equipped with a detachable 3-wire cable which grounds the instrument to the power line ground (protective earth) when plugged into an appropriate 3-wire receptacle. If a 3-wire receptacle is not available, consult the hospital electrician. If completeness of the protective grounding wire is in doubt, the equipment must be operated with internal power supply.

⚠ Warning ⚠

Do not use a 3-wire to 2-wire adapter with this instrument.

Connect the grounding wire to the equipotential grounding terminal on the main system. If it is not evident from the instrument specifications whether a particular instrument combination is hazardous or not, for example due to summation of leakage currents, the user should consult the manufacturers concerned or else an expert in the field, to ensure that the necessary safety of all instruments concerned will not be impaired by the proposed combination.

Equipotential Grounding

When other equipments are used together with the monitor, a grounding cable should be used to connect the equipotential grounding connectors of the monitor and of other equipments. This helps to reduce the potential differences between different pieces of equipment, and ensure the safety of the operator and patient.

⚠ Warning ⚠

If the protective grounding (protective earth) system is doubtful, the monitor must be supplied by inner power only.

Condensation

Make sure that during operation, the instrument is free of condensation. Condensation can form when equipment is moved from one building to another, thus being exposed to moisture and differences in temperature.

⚠ Warning ⚠

Possible explosion hazard if used in the presence of flammable anesthetics.

Chapter 11 Care / Cleaning

11.1 System Check

Before using the monitor, do the following:

- check if there is any mechanical damage;
- check all the outer cables, inserted modules and accessories;
- check all the functions of the monitor to make sure that the monitor is in good condition.

If you find any damage on the monitor, stop using the monitor on patient, and contact the biomedical engineer of the hospital or Mindray Customer Service immediately.

The overall check of the monitor, including the safety check, should be performed only by qualified personnel once every 6 to 12 month, and each time after fix up.

You should check the synchronism of the defibrillator in the frequency described in the hospital regulations. At least every 3 months, it should be checked by a qualified customer service technician.

All the checks that need to open the monitor should be performed by qualified customer service technician. The safety and maintenance check can be conducted by persons from Mindray. You can obtain the material about the customer service contract from the local Mindray office.

Warning

If the hospital or agency that is responding to using the monitor does not follow a satisfactory maintenance schedule, the monitor may become invalid, and the human health may be endangered.

Note

To ensure maximum battery life, it is recommended that, at least once a month, the monitor be run on battery until it turns itself off and then recharged.

Warning

Refer the battery replacement only to Mindray service technician.

11.2 General Cleaning

⚠ **Warning** ⚠

Before cleaning the monitor or the sensor, make sure that the equipment is switched off and disconnected from the power line.

The PM-8000 Patient Monitor must be kept dust-free.

Regular cleaning of the monitor shell and the screen is strongly recommended. Use only non-caustic detergents such as soap and water to clean the monitor shell.

⚠ **Note** ⚠

Please pay special attention to the following items:

1. **Avoid using ammonia-based or acetone-based cleaners such as acetone.**
2. **Most cleaning agents must be diluted before use. Follow the manufacturer's directions carefully to avoid damaging the monitor.**
3. **Don't use the grinding material, such as steel wool etc.**
4. **Don't let the cleaning agent enter into the chassis of the system.**
5. **Don't leave the cleaning agents at any part of the equipment.**

11.3 Cleaning Agents

Examples of disinfectants that can be used on the instrument casing are listed below:

- Diluted Ammonia Water
- Diluted Sodium Hyoichlo (Bleaching agent).

⚠ **Note** ⚠

The diluted sodium hyoichlo from 500ppm(1:100 diluted bleaching agent) to 5000ppm (1:10 bleaching agents) is very effective. The concentration of the diluted sodium hyocihlo depends on how many organisms (blood, mucus) on the surface of the chassis to be cleaned.

- Diluted Formaldehyde 35% -- 37%
- Hydrogen Peroxide 3%
- Alcohol
- Isopropanol

⚠ **Note** ⚠

PM-8000 monitor and sensor surface can be cleaned with hospital-grade ethanol and dried in air or with crisp and clean cloth.

⚠ **Note** ⚠

Mindray has no responsibility for the effectiveness of controlling infectious disease using these chemical agents. Please contact infectious disease experts in your hospital for details.

11.4 Sterilization

To avoid extended damage to the equipment, sterilization is only recommended when stipulated as necessary in the Hospital Maintenance Schedule. Sterilization facilities should be cleaned first.

Recommended sterilization material: Ethylate, and Acetaldehyde.

Appropriate sterilization materials for ECG lead, blood pressure cuff are introduced in **Chapters ECG/RESP Monitoring**, **Chapter NIBP Monitoring** respectively.

⚠ **Caution** ⚠

- **Follow the manufacturer's instruction to dilute the solution, or adopt the lowest possible density.**
- **Do not let liquid enter the monitor.**
- **No part of this monitor can be subjected to immersion in liquid.**
- **Do not pour liquid onto the monitor during sterilization.**
- **Use a moistened cloth to wipe up any agent remained on the monitor.**

11.5 Disinfection

To avoid extended damage to the equipment, disinfection is only recommended when stipulated as necessary in the Hospital Maintenance Schedule. Disinfection facilities should be cleaned first.

Appropriate disinfection materials for ECG lead, SpO2 sensor, blood pressure cuff, TEMP probe, IBP sensor are introduced in **Chapters 12-16** respectively.

⚠ **Caution** ⚠

Do not use EtO gas or formaldehyde to disinfect the monitor.

FOR YOUR NOTES

Chapter 12 ECG/RESP Monitoring

12.1 What Is ECG Monitoring

Monitoring the ECG produces a continuous waveform of the patient's cardiac electric activity to enable an accurate assessment of his current physiological state. Only proper connection of the ECG cables can ensure satisfactory measurement. On the Normal Display, PM-8000 provides display of 2-channel ECG waveforms.

- The patient cable consists of 2 parts (Refer to **Chapter Accessories** for more detailed information of the ECG accessories);
  The cable that connects to the monitor;
  The lead set that connects to the patient.
- Using a 5-lead set, the ECG can derive up to two waveforms from two different leads. For requested lead, you may choose from the left side of ECG waveform.
- The monitor displays the Heart Rate (HR), ST segment and Arrhythmia analysis.
- All of the parameters above can be set as alarm parameters.

⚠ **Note** ⚠

In the default settings of PM-8000, the ECG waveforms are the first two waveforms from top in the Waveform Area.

12.2 Precautions during ECG Monitoring

⚠ **Warning** ⚠

Do not touch the patient, table nearby, or the equipment during defibrillation.

⚠ **Warning** ⚠

Use only the original PM-8000 ECG cable for monitoring.

⚠ **Warning** ⚠

When connecting the cables and electrodes, make sure no conductive part is in contact with the ground. Verify that all ECG electrodes, including neutral electrodes, are securely attached to the patient.

⚠ **Warning** ⚠

When apply the ECG cable with no resistances to Mindray patient monitor or other patient monitors which themselves with no current limit resistance, it can't be applied to defibrillation.

⚠ **Note** ⚠

Interference from a non-grounded instrument near the patient and ESU interference can cause inaccuracy of the waveform.

12.3 Monitoring Procedure

12.3.1 Preparation

1. Prepare the patient's skin prior to placing the electrodes.
   - The skin is a poor conductor of electricity, therefore preparation of the patient's skin is important to facilitate good electrode contact to skin.
   - Shave hair from sites, if necessary.
   - Wash sites thoroughly with soap and water. (Never use ether or pure alcohol, because this increases skin impedance).
   - Rub the skin briskly to increase capillary blood flow in the tissues and remove skin scurf and grease.
2. Attach clip or snap to electrodes prior to placement.
3. Put the electrodes on the patient. Before attaching, apply some conductive jelly on the electrodes if the electrodes are not electrolyte self-supplied.
4. Connect the electrode lead to the patient's cable.
5. Make sure the monitor is ready with power supply.

⚠ **Warning** ⚠

Check everyday whether there is skin irritation resulted from the ECG electrodes. If so, replace electrodes every 24 hours or change their sites.

⚠ **Note** ⚠

For protecting environment, the electrodes must be recycled or disposed of properly.

⚠ **Warning** ⚠

Verify lead fault detection prior to the start of monitoring phase. Unplug the ECG cable from the socket, the screen will display the error message "ECG LEAD OFF" and the audible alarm is activated.

12. 3. 2 Installing ECG lead

Placing the Electrodes for ECG Monitoring

Electrode placement for 5-lead set (Figure 12-1)

- Red (R) electrode - Be placed near the right shoulder, directly below the clavicle.
- Yellow (L) electrode - Be placed near the left shoulder, directly below the clavicle.
- Black (N) electrode - Be placed on the right hypogastrium.
- Green (F) electrode - Be placed on the left hypogastrium.
- White (C) electrode - Be placed on the chest as illustrated in the F Figure 12-2

Note: the following table gives the corresponding lead names used in Europe and America respectively. (Lead names are represented by R, L, N, F and C respectively in Europe, whose corresponding lead names in America are RA, LA, RL, LL and V.)

America		Euro	
Lead names	color	Lead names	color
RA	White	R	Red
LA	Black	L	Yellow
LL	Red	F	Green
RL	Green	N	Black
V	brown	C	White



Figure 12-1 Electrode placement for 5-lead set

⚠ **Note** ⚠

To ensure patient safety, all leads must be attached to the patient.

For 12-lead set, attach the C-electrode to one of the indicated positions as below (Figure 12-2):

- V1 On the 4th intercostal space at the right sterna margin.
- V2 On the 4th intercostal space at the left sterna margin.
- V3 Midway between V2 and V4 electrodes.
- V4 On the 5th intercostal space at the left clavicular line.
- V5 On the left anterior axillary line, horizontal with V4 electrode.
- V6 On the left middle axillary line, horizontal with V4 electrode.
- V3R-V7R On the right side of the chest in positions corresponding to those on the left.
- VE Over the xiphoid position.
- V7 On the 5th intercostal space at the left posterior axillary line of back.
- V7R On the 5th intercostal space at the right posterior axillary line of back.



Figure 12-2 C-electrode placement for 5-lead set

Recommended ECG Lead Placement for Surgical Patients

⚠ **Warning** ⚠

When using Electrosurgery equipment, leads should be placed in a position in equal distance from Electrosurgery electrotome and the grounding plate to avoid cautery. Electrosurgery equipment wire and ECG cable must not be tangled up.

The placing of the ECG leads will depend on the type of surgery that is being performed. For example, with open chest surgery the electrodes may be placed laterally on the chest or on the back. In the operating room, artifacts can sometimes affect the ECG waveform due to the use of ES (Electrosurgery) equipment. To help reduce this you can place the electrodes on the right and left shoulders, the right and left sides near the stomach, and the chest lead on

the left side at mid-chest. Avoid placing the electrodes on the upper arms, otherwise the ECG waveform will be too small.

⚠ **Warning** ⚠

When using Electrosurgery equipment, never place an electrode near the grounding plate of the Electrosurgery device, otherwise there will be a great deal of interference with the ECG signal.

⚠ **Note** ⚠

If a ECG waveform is not accurate, while the electrodes are tightly attached, try to change the lead.

⚠ **Note** ⚠

Interference from a non-grounded instrument near the patient and ESU interference can cause inaccuracy of the waveform.

Normal QRS complex should be:

- Tall and narrow with no notches.
- With tall R-wave completely above or below the baseline.
- With pacer spike no higher than R-wave height.
- With T-wave less than one-third of the R-wave height.
- With P-wave much smaller than the T-wave.

For getting 1 mv calibrated ECG wave, choose ECG CAL button in ECG SETUP menu. A message "when CAL, can't monitor! " prompts on the screen.



Figure 12-3 Standard ECG Waveform

⚠ **Warning** ⚠

Do not touch the patient, table nearby, or the equipment during defibrillation.

12.4 ECG Screen Hot Keys



Figure 12-4 the hot key for ECG

① Leads of channel 1:

1） The selectable leads are I, II, III, aVR, aVL, aVF, V.

2） When the ECG is 5-lead, the selectable leads are: I, II, III, aVR, aVL, aVF, V; when ECG is 3-lead, the selectable leads are: I, II, III.

3） Leads on the ECG wave must not have the same name. Otherwise, the system will automatically change the ECG waveform name that has the same name as the waveform being currently adjusted to another name.

② Waveform gain of channel 1: used to adjust the size of ECG waveforms

Select gain value for each channel from ×0.25,×0.5,×1,×2, and auto. Under "auto" mode, the monitor chooses an appropriate level automatically. A 1mv scale displays on each ECG channel's right side. The height of 1mV bar is directly proportional to the waveform amplitude.

⚠ **Note** ⚠

When the input signals are too large, the peak of the waveform may be not able to be displayed. In this case the user may manually change the setup method of ECG waveform according to the actual waveform so as to avoid the occurrence of the unfavorable phenomena.

③ Filter method: used for displaying clearer and more detailed waveform

There are three filter modes for selection. DIAGNOSTI, MONITOR and SURGERY modes may reduce perturbance and interference from Electrosurgery equipment. The filter method is the item applicable for both channels, which is always displayed at the waveform place of the channel 1 ECG waveform.

⚠ **Note** ⚠

Only in Diagnosis mode, the system can provide non-processed real signals. In Monitor or Sugery mode, ECG waveforms may have distortion of different extent. In either of the latter two modes, the system can only show the basic ECG and the results of ST analysis may also be greatly affected. In Surgery mode, results of ARR analysis may be somewhat affected. Therefore, it is suggested that in the environment having relative small interference, you'd better monitor a patient in Diagnosis mode.

④ Leads of channel 2: refer to ① for detailed information.

⑤ Waveform gain of channel 2: refer to ② for detailed information.

⚠ **Note** ⚠

Pacemaking signal detected is marked by a "|" above the ECG waveform.

12.5 ECG Menu

ECG SETUP Menu

Pick the ECG hot key on the screen, and the following menu will popup.



Figure 12-5 ECG SETUP menu

ECG alarm setting

- HR ALM: pick "ON" to enable prompt message and data record during the ECG alarm; pick "OFF" to disable the alarm function, and there will be a beside "ECG".
- ALM LEV: selectable from HIGH, MED, LOW. Level HIGH represents the most serious case.
- ALM REC: pick "ON" to enable report printing upon ECG alarm.

- ALM HI: used to set up the upper limit of ECG alarm.
- ALM LO: used to set up the lower limit of ECG alarm.

ECG alarm is activated when the heart beat exceeds set ALM HI value or falls below ALM LO value.

ECG alarm limits:

	Max. ALM HI	Min. ALM LO	Step
HR ADU	300	15	1
HR PED	350	15	1
HR NEO	350	15	1

⚠ **Note** ⚠

Please set the alarm limits according to clinical condition of individual patient. The upper limit shall not exceed 20 beat/min higher than the patient's heart rate.

■ HR FROM

ECG, SpO_2, AUTO and BOTH may detect heart rate. AUTO distinguishes heart rate source according to the quality of signal. By picking ECG, the monitor prompts HR and activates HR beep. By picking SpO_2, the monitor prompts PULSE and activates pulse beep. BOTH mode displays HR and PR simultaneously, when this item is picked, PR parameter is displayed to the right side of SpO_2. As for the sound of HR or PR in BOTH mode, HR is given the priority, i.e., if HR is available, whose sound will be sent out, but if HR is not available, then the sound will be for PR.

■ HR CHANNEL

"CH1" to count the heart rate by CH 1 waveform

"CH2" to count the heart rate by CH 2 waveform

"AUTO" the monitor selects a channel automatically

■ LEAD TYPE: used to select either 5 LEADS or 3 LEADS.

■ SWEEP

Available options for ECG SWEEP are 12.5, 25.0, and 50.0 mm/s.

■ ST ANALYSIS

Pick this item to access ST ANALYSIS menu, the detailed information about the menu is to be discussed in the following section.

■ ARR ANALYSIS

Pick this item to access ARR ANALYSIS menu, the detailed information about the menu is to be discussed in the following section.

■ OTHER SETUP

Pick this item to access ECG SETUP menu as shown below:



Figure 12-6 ECG SETUP menu

In the sub-menu, following functions are available:

- ECG DISPLAY: Select NORMAL DISPLAY to display 2 ECG waveforms for 5-lead (for 3-lead, only 1 ECG waveform is displayed.). Select MULTI-LEADS DISPLAY, the waveform area on the screen displays 6 ECG waveforms. Select HALF-SCAN MULTI-LEADS, there are 4 ECG waveforms are displayed on the screen.
  Note: If 3 LEADS is selected in the ECG SETUP menu, only NORMAL DISPLAY can be selected for ECG DISPLAY item in the sub-menu.
- BEAT VOL
  The options are from “10” to “0”. “10” indicates the maximum volume while “1” the minimum,”0” indicates OFF. If your machine does not have "Multilevel Volume" function, only four options are available for volume, “3”, “2”, “1” and ”OFF”.

⚠ **Note** ⚠

PITCH TONE volume is controlled through adjusting the heart beat volume. However, if SPO2 is selected as “HR FROM” in “ECG SETUP”, the PITCH TONE volume will accordingly controlled through adjusting “PR SOUND” in “SPO2 SETUP” menu. Refer to the chapter about SPO2 to know the detailed information about PITCH TONE. (If your machine does not have "PITCH TONE" function, you can jump over this paragraph)

- PACE
  "ON" detected signal will be marked by a "|" above the ECG waveform.
  "OFF" for non-pacemaking patient.

⚠ **Note** ⚠

If monitoring a patient with the pacemaker, set “PACE” to On. If monitoring a patient without pacemaker, set “PACE” to Off.

If “PACE” is on, the system will not perform some types of ARR analysis. For detailed information, please refer to the section: ARR ALARM. In the table, the ARR type marked by All types applies to the analysis in all situations, marked by Non-paced applies only to the analysis in the situation when the patient does not use pacemaker.

- CASCADE: switch for ECG cascade. CASCADE: wave of each channel is displayed in two lines. This function effects only when NORMAL DISPLAY is selected for ECG DISPLAY.
- ECG CAL: pick this item to start calibrating ECG. The method to end CAL: re-select the CAL key in the menu or re-select the lead name on the screen.
- ADJUST WAVE POS: used to adjust the position of the waveform on the screen. Pick to access ADJUST WAVE POS dialog box. The user may use CH NAME item to select the channel to be adjusted, UP-DOWN to adjust the position of the selected channel on the screen, BACK TO DEFAULT to let the wave go back to the default position on the screen.



Figure 12-7 ADJUST WAVE POS menu

- DEFAULT: pick this item to access the ECG DEFAULT CONFIG dialog box, in which the user may select whether the FACTORY DEFAULT CONFIG or the USER DEFAULT CONFIG is to be used. After selecting any of the items and exiting the dialog box, the system will pop up the dialog box asking for the user's confirmation.

⚠ **Warning** ⚠

For pacemaker patient, the pacing impulse analysis function must be switched on, otherwise, the pacing impulse may be counted as normal QRS complex, which results in failure of "ECG LOST" error detection.

■ For monitor with ST segment & Arrhythmia analysis software, refer to **ST Segment Monitoring** and **Arrhythmia Analysis** for details.

⚠ **Note** ⚠

When Pacer Switch is On, the Arrhythmia events related to PVCs will not be monitored. At the same time, the ST analysis will not be performed either.

12.6 ECG Alarm Information and Prompt

Alarm Message

Alarms occurring in the process of ECG measurement contain two types: physiological alarm and technical alarm. Prompt message may also appear in the mean time. For the audio and visual features during the appearance of these alarms and prompt messages in the process of ECG measurement, please refer to the related description in Chapter Alarm. In the screen, physiological alarm messages and the prompt messages able to trigger alarms (general alerts) all displayed in the alarm area of the monitor while technical alarms and prompt messages unable to trigger alarms are then displayed in the information area of the monitor. This section does not describe the content about Arr. and ST analysis.

Among physiological alarms, those belonging to the type that the parameter has exceeded the limits may activate the recorder to automatically output the parameters and related measured waveforms when the alarms occur on the condition that the alarm record switch in the related menu is On.

Tables below describe respectively the possible various alarms those may occur during the measurement.

Physiological alarms:

Message	Cause	Alarm level
ECG LOST	No ECG signal of the patient is detected.	HIGH
HR TOO HIGH	HR measuring value is above the upper alarm limit	User-selectable
HR TOO LOW	HR measuring value is below the lower alarm limit	User-selectable

Technical alarms:

Message	Cause	Alarm level	Remedy
ECG LEAD OFF	ECG electrodes fall off the skin or ECG cables fall off the monitor.	LOW	Make sure that all electrodes, leads and patient cables are properly connected.
ECG V LEAD OFF or ECG C LEAD OFF			
ECG LL LEAD OFF or ECG F LEAD OFF			
ECG LA LEAD OFF or ECG L LEAD OFF			
ECG RA LEAD OFF or ECG R LEAD OFF			
ECG INIT ERR	ECG module failure	HIGH	Stop using measuring function provided by ECG module, notifies biomedical engineer or Mindray service staff.
ECG INIT ERR1			
ECG INIT ERR2			
ECG INIT ERR3			
ECG INIT ERR4			

ECG INIT ERR5			
ECG INIT ERR6			
ECG INIT ERR7			
ECG INIT ERR8			
ECG COMM STOP	Occasional communication failure	HIGH	If failure persists, notify biomedical engineer or Mindray service staff.
ECG COMM ERR	Occasional communication failure	HIGH	If failure persists, notify biomedical engineer or Mindray service staff.
HR ALM LMT ERR	Functional safety failure	HIGH	Stop using HR alarm function, notify biomedical engineer or Mindray service staff.
ECG NOISE	ECG measuring signal is greatly interfered.	LOW	Make sure the patient is quiet, the electrodes are properly connected and AC power system is well grounded.

Prompt messages (include general alerts):

Message	Cause	Alarm Level
HR EXCEED	HR measuring value exceeds the measurement range.	HIGH

12.7 ST Segment Monitoring (optional)

ST Segment Monitoring (Optional)

- ST segment monitoring function is shutoff by default. You can switch it to ON when necessary.

  NOTE: When setting ST ANALYSIS on, the monitor will select “DIAGNOSTIC” mode. You can set it to “MONITOR” mode or “OPERATE” mode as required. However at this time ST value has been severely distorted.
- It is available to measure the variance of ST segment with ST analysis at the waveform tracks for selected lead. The corresponding ST measurement result displays numerically at ST1 and ST2 in the Parameter Area. The trend can be viewed with table or graphic form.
- Measurement unit of ST segment: mv.
- Measurement symbol of ST segment: "+" = elevating, "-" = depressing.
- Measurement range of ST segment: -2.0 mv, ~ + 2.0 mv.

Pick the ST ANALYSIS item in the ECG SETUP menu to access the ST ANALYSIS sub-menu as shown below.

ST ANALYSIS menu



Figure 12-8 ST ANALYSIS menu

ST analysis alarm setting

- ST ANAL: the switch for ST analysis. Set it to ON to activate the ST analysis or OFF to disable the ST analysis.
- ST ALM: pick "ON" to enable prompt message and data record during the ST analysis alarm; pick "OFF" to disable the alarm function, and there will be a beside ST. ST alarm is activated when the result exceeds set ST HI value or falls below ST LO value.
- ALM LEV: used to set up the ST alarm level. There are three selections: HIGH, MED and LOW.
- ALM REC: pick "ON" to enable report printing upon ST analysis alarm.
- ALM HI: used to set up the upper limit of ST alarm. The max. higher limit is 2.0. The minimum higher limit is 0.2 larger than the set lower limit.
- ALM LOW: used to set up the lower limit of ST alarm. The minimum lower limit is –2.0. The max. lower limit is 0.2 lower than the set higher limit.

ST analysis alarm limits:

	Max. ST HI	Min. ST LO	Step
ST	2.0 mv	-2.0 mv	0.1

- DEF POINT pick this item to access the DEF POINT window, in which the position of ISO and ST point can be set up.
  - ISO Base point. Default is 78 ms.
  - ST Measurement point.



Figure 12-9 DEF POINT window

The operator can adjust the position of both ISO and ST measurement points.

The reference point is the position where the peak of R-wave locates.



Figure 12-10 DEF Point

The ST measurement for each beat complex is the vertical difference between the two measurement points.

⚠ **Note** ⚠

the ST measurement point should be adjusted if the patient's HR or ECG morphology changes significantly.

□ Adjusting ISO, ST

These two points can be adjusted turning the knob.

When adjusting ST measurement point, the system will show the ST Measurement Point Window. The QRS complex template displays in the window (If the template is not established, a horizontal line will display. If the channel is not at ON position, a horizontal line will also display). It is adjustable of the highlight bar in the window. You may select ISO or ST, then switch the knob left or right to move the cursor line. When the cursor is at the required position, you may select the base point or the measurement point.

⚠ **Note** ⚠

Abnormal QRS complex is not considered in ST segment analysis.

ST Alarm Message

Note: The alarm limits for two ST measurements are identical. No setting of alarm limits can be made only for one channel.

Among physiological alarms, those belonging to the type that the parameter has exceeded the limits may activate the recorder to automatically output the parameters and related measured waveforms when the alarms occur on the condition that the alarm record switch in the related menu is On.

Tables below describe the possible physiological alarms, technical alarms and prompt messages during ST measurement.

Physiological alarms:

Message	Cause	Alarm Level
ST1 TOO HIGH	ST measuring value of channel 1 is above the upper alarm limit.	User-selectable
STI TOO LOW	ST measuring value of channel 1 is below the lower alarm limit.	User-selectable
ST2 TOO HIGH	ST measuring value of channel 2 is above the upper alarm limit.	User-selectable
ST2 TOO LOW	ST measuring value of channel 2 is below the lower alarm limit.	User-selectable

Technical alarms:

Message	Cause	Alarm Level	Remedy
ST ALM LMT ERR	Functional safety failure	HIGH	Stop using ST alarming function, notify biomedical engineer or Mindray service staff.

Prompt messages (include general alerts):

Message	Cause	Alarm Level
ST1 EXCEED	ST measuring value of channel 1 exceeds the measurement range.	HIGH
ST2 EXCEED	ST measuring value of channel 2 exceeds the measurement range.	HIGH

12.8 Arr. Monitoring (optional)

Arrhythmia Analysis

The arrhythmia algorithm is used to monitor ECG of neonate and adult patient in clinical, detect the changing of heart rate and ventricular rhythm, and also save arrhythmia events and generate alarming information. Arrhythmia algorithm can monitor paced and non-paced patients. Qualified personnel can use arrhythmia analysis to evaluate patient's condition (such as heart rate, PVCs frequency, rhythm and ectopic beat) and decide the treatment. Besides detecting changing of ECG, arrhythmia algorithm can also monitor patients and give proper alarm for arrhythmia.

- The arrhythmia monitoring is shutoff by default. You can enable it when necessary.
- This function can call up the doctor's attention to the patient's heart rate by measuring and classifying the arrhythmia and abnormal heart beat and triggering the alarm.
- The monitor can conduct up to 13 different arrhythmia analyses.
- The monitor can store the latest 60 alarm events when taking arrhythmia analysis to a peculiar buffer. The operator can edit these arrhythmia events through the menu below.

Pick the item ARR ANALYSIS in ECG SETUP menu to access the ARR ANALYSIS sub-menu.

ARR ANALYSIS Menu



Figure 12-11 ARR ANALYSIS Menu

- ARR ANAL: Pick "ON" during monitoring. Default set is "OFF".
- PVCs ALM: pick "ON" to enable prompt message and data record when alarm occurs; pick "OFF" to disable the alarm function, and there will be a beside "PVCs".
- ALM LEV: selectable from HIGH, MED, LOW. Level HIGH represents the most serious case.
- ALM REC: pick "ON" to enable report printing upon PVCs alarm.

  PVCs alarm is activated when the PVCs exceeds set PVCs ALM HI value.

PVCs alarm upper limits:

	Max	Min	Step
PVCs	10	1	1

PVCs alarm and prompt message:

Among physiological alarms, those belonging to the type that the parameter has exceeded the limits may activate the recorder to automatically output the parameters and related measured waveforms when the alarms occur on the condition that the alarm record switch in the related menu is On.

Tables below describe the possible physiological alarms, technical alarms and prompt messages occurring during pvcs measurement.

Physiological alarms:

Message	Cause	Alarm Level
PVCs TOO HIGH	PVCs measuring value is above upper alarm limit.	User-selectable

Technical alarms:

Message	Cause	Alarm Level	Remedy
PVCs ALM LMT ERR	Functional safety failure	HIGH	Stop using PVCs alarming function, notify biomedical engineer or Mindray service staff.

- ARR RELEARN Pick this item to start a learning procedure.
- ARR ALARM Pick this item to access the ARR ALARM dialog box to set arrhythmia alarm parameters.

Set ALM to ON/OFF to enable/disable the alarm function; Set REC to ON/OFF to enable/disable alarm record function, turn the knob under LEV column to set alarm level to HIGH, MED or LOW.



Figure 12-12 ARR ALARM Menu

You can pick ALL ALM ON to enable alarm function of all arrhythmia types and pick ALL ALM

OFF to disable this function. Likewise, you can pick ALL REC ON to enable recording function for all arrhythmia types and pick ALL REC OFF to disable this function. Changing the ALM LEV can reset alarm level of all arrhythmia types to the same value.

■ ARR RECALL Pick this item to review and edit the ARR analysis result.
The latest arrhythmia events (up to 60) are displayed.



Figure 12-13 ARR RECALL Menu

- □ UP-DOWN Observe other event lists of other page.
- □ CURSOR Select the Arr. event, whose name is displayed in a protruding frame.
- □ DELETE Delete the selected Arr. event.
- □ RENAME Rename the selected Arr. event, whose name is displayed in a sunken frame.
  Switch the knob until the name you want appears.
- □ WAVE To display the Arrhythmia waveform, time and parameter value.
  - ❍ UP-DOWN To observe waveforms of other Arrhythmia events.
  - ❍ L_RIGHT To observe 8-second waveform of Arrhythmia events.
  - ❍ REC To print out displayed Arrhythmia event.
  - ❍ EXIT To return to ARR RECALL menu of Arrhythmia event.



Figure 12-14 ARR WAVE RECALL Menu

⚠ **Note** ⚠

If there are more than 60 Arrhythmia events, the latest will be retained.

ARR ALARM

The alarm is triggered when an Arrhythmia occurs. If the ALM is ON, the alarm sounds and the alarm indicator flashes. If the REC is ON, the alarm record will be printed out (4 seconds prior to and after the alarm, with the ECG waveforms of analysis channel).

Physiological alarms:

Prompt	Applicable Patient Type	Occurring Condition	Alarm Level
ASYSTOLE	All patients	No QRS is detected for 4 consecutive seconds	User-selectable
VFIB /VTAC	Without pacemaker	Fibrillatory wave for consecutive 4 seconds; or The number of continuous Vent beats is larger than the upper limit of cluster Vent beats (>5). The RR interval is less than 600ms.	User-selectable
VT>2	Without pacemaker	3 < the number of cluster PVCs < 5	User-selectable
COUPLET	Without pacemaker	2 consecutive PVCs	User-selectable
BIGEMINY	Without pacemaker	Vent Bigeminy	User-selectable
TRIGEMINY	Without pacemaker	Vent Trigeminy	User-selectable
R ON T	Without pacemaker	A type of single PVC under the condition that HR<100， R-R interval is less than 1/3 the average interval, followed by a compensating pause of 1.25X the average R-R interval(the next R wave advances onto the previous T wave).	User-selectable

PVC	Without pacemaker	Single PVCs not belonging to the type of above mentioned PVCs.	User-selectable
TACHY	All patients	5 consecutive QRS complex , RR interval is less than 500ms.	User-selectable
BRADY	All patients	5 consecutive QRS complex, RR interval is longer than 1.5s.	User-selectable
MISSED BEATS	Without pacemaker	When HR is less than 100 beats/min., no heart beat is tested during the period 1.75 times of the average RR interval; or When HR is larger than 100 beats/min., no beat is tested with 1 second.	User-selectable
PNP	With pacemaker	No QRS complex and pacing pulse are availabe during the period 1.75 times of the average R-R interval (only considering patients with pacemaker.)	User-selectable
PNC	With pacemaker	When pacing pulse is available, no QRS exists during the period 1.75 times of the average RR interval (only considering patients with pacemaker.)	User-selectable

Patient type:

All patients: refers to perform Arr.analysis on patients either with pacemakers or without pacemakers.

Without pacemaker: refers to perform Arr. Analysis only on the patients without pacemakers.

With pacemaker: refers to perform Arr. Analysis only on the patients with pacemakers.

Prompt message:

Message	Cause	Alarm Level
ARR LEARNING	The QRS template building required for Arr. Analysis is in process.	No alarm

⚠ **Note** ⚠

Arrhythmia name displays in the Alarm Message Area.

12.9 Measuring RESP

12.9.1 How to measure RESP?

The monitor measures respiration from the amount of thoracic impedance between two ECG electrodes. The change of impedance between the two electrodes, (due to the thoracic movement), produces a respiratory waveform on the screen.

12.9.2 Setting Up RESP measurement

For RESP monitoring, it is not necessary for additional electrodes, however, the placing of electrodes is important.

Some patients, due to their clinical condition, expand their chest laterally, causing a negative intrathoracic pressure. In these cases it is better to place the two RESP electrodes laterally in the right axillary and left lateral chest areas at the maximum point of breathing movement to optimize the respiratory waveform.

⚠ **Note** ⚠

The RESP monitoring is not recommended to be used on patients who are very active, as this can cause false alarms.
Please select the ECG cable with no resistance for RESP monitoring.

Checklist for RESP Monitoring

1. Prepare the patient's skin prior to placing the electrodes.
2. Attach snap or clip to the electrodes and attach the electrodes to the patient as described below.
3. Switch on the monitor.

12.9.3 Installing electrode for RESP measurement

Placing the Electrodes for Respiratory Monitoring



Figure 12-15 Electrodes placement (5-lead)

⚠ **Note** ⚠

Place the red and green electrodes diagonally to optimize the respiration waveform. Avoid the liver area and the ventricles of the heart in the line between the RESP electrodes so as to avoid cardiac overlay or artifacts from pulsating blood flow. This is particularly important for neonates.

12.9.4 RESP menu

RESP SETUP Menu

Pick RESP hot key on the screen to call up the following menu:



Figure 12-16 RESP SETUP Menu

RESP alarm setting

- ALM: pick "ON" to enable prompt message and data record during the RESP alarm; pick "OFF" to disable the alarm function, and there will be a beside “RESP”.
- ALM REC: pick "ON" to enable report printing upon RESP alarm.
- ALM LEV: selectable from HIGH, MED and LOW. Level HIGH represents the most serious case.
- ALM HI: used to set up the upper alarm limit.
- ALM LO: used to set up the lower alarm limit.
  RESP alarm is activated when the respiration rate exceeds set ALM HI value or falls below ALM LO value.

RESP alarm limits:

	Max. RR HI	Min. RR LO	Step
RESP ADU	120	0	1
RESP NEO/PED	150	0	1

- APNEA ALM: to set the standard of judging an apnea case. It ranges from 10 to 40 seconds, increases / decreases by 5.
- SWEEP: Available options for RESP SWEEP are 6.25, 12.5 and 25.0 mm/s.
- WAVE AMP: The user may set up the displaying amplitude of the RESP waveform. The selections are 0.25, 0.5, 1, 2, 3, 4, 5.
- HOLD TYPE: AUTO/MANUAL adjustable. When it is AUTO mode, HOLD HI and HOLD LO menus cannot be used and the monitor automatically calculates the RESP RATE.
- HOLD HI and HOLD LO: When it is AUTO mode, HOLD HI and HOLD LO menus cannot

be used and the monitor automatically calculates the RESP RATE. When the HOLD TYPE is MANUAL, the user can use the knob to pick either HOLD HI or HOLD LO and turn the knob to adjust the two dashed lines in the RESP WAVEFORM area respectively. The positions of the dashed lines will be used to calculate the upper and lower limits of RESP RATE by the monitor.

- DEFAULT: pick this item to access the RESP DEFAULT CONFIG dialog box, in which the user may select whether the FACTORY DEFAULT CONFIG or the USER DEFAULT CONFIG is to be used. After selecting any of the items and exiting the dialog box, the system will pop up the dialog box asking for the user's confirmation

RESP Alarm Message

Among physiological alarms, those belonging to the type that the parameter has exceeded the limits may activate the recorder to automatically output the parameters and related measured waveforms when the alarms occur on the condition that the alarm record switch in the related menu is On.

Tables below describe the possible physiological alarms, technical alarms and prompt messages occurring during resp measurement.

Physiological alarms:

Message	Cause	Alarm Level
RR TOO HIGH	RESP measuring value is above upper alarm limit.	User-selectable
RR TOO LOW	RESP measuring value is below lower alarm limit.	User-selectable
RESP APNEA	RESP can not be measured within specific time interval.	HIGH

Technical alarms:

Message	Cause	Alarm Level	Remedy
RESP ALM LMT ERR	Functional safety failure	HIGH	Stop using RESP alarming function, notify biomedical engineer or Mindray service staff.

Prompt message (general alerts):

Message	Cause	Alarm Level
RR EXCEED	RR measuring value exceeds the measure range.	HIGH

12.10 Maintenance and Cleaning

Care and Cleaning

⚠ **Warning** ⚠

Before cleaning the monitor or the sensor, make sure that the equipment is switched off and disconnected from the power line.

If there is any sign that the ECG cable may be damaged or deteriorated, replace it with a new one instead of continuing its application on the patient.

- Cleaning:
  Use fine-hair cloth moistened in mild soap liquid or cleaning agent containing 70% ethanol to clean the equipment.
- Sterilization
  To avoid extended damage to the equipment, sterilization is only recommended when stipulated as necessary in the Hospital Maintenance Schedule. Sterilization facilities should be cleaned first.
  Recommended sterilization material:
  - Ethylate: 70% alcohol, 70% isopropanol
  - Acetaldehyde
- Disinfection
  To avoid extended damage to the equipment, disinfection is only recommended when stipulated as necessary in the Hospital Maintenance Schedule. Disinfection facilities should be cleaned first.

Chapter 13 SpO2 Monitoring

13.1 PART 1 (MASIMO SpO2 board configuration)

General description

SpO_2 is a non-invasive measurement of the functional oxygen saturation.
The measurement is taken by placing a sensor on a patient , usually on the fingertip for adults, and the hand or foot for neonates. The sensor is connected to the patient monitor with pulse oximetry measurement module (Masimo Set, which is called MS-7). The monitor displays the calculated data from MS-7 in three ways: 1)as a percent value for arterial oxygen saturation (SpO_2) ; 2)as a pulse rate (PR) and 3)as a plethysmographic waveform on the screen.

Principles of Operation

This MS-7 is based on three principles:

- Oxyhemoglobin and deoxyhemoglobin differ in their absorption of red and infrared light(spectrophotometry).
- The volume of arterial blood in tissue and the light absorbed by the blood changes during the pulse(plethysmography)
- Arterio-venous shunting is highly variable and that fluctuating absorbance by venous blood is a major component of noise during the pulse.

This MS-7 determines SpO_2 by passing red and infrared light into a capillary bed and measuring changes in light absorption during the pulsatile cycle. Red and infrared light-emitting diodes (LEDs) in oximetry sensors serve as the light sources, a photodiode serves as the photodetector.

Traditional pulse oximeter assumes that all pulsations in the light absorbance signal are caused by oscillations in the arterial blood volume. This assumes that the blood flow in the region of the sensor passes entirely through the capillary bed rather than through any arterio-venous shunts. The traditional pulse oximeter calculates the ratio of pulsatile absorbance (AC) to the mean absorbance (DC) at each of two wavelengths, 660 nm and

940 nm:

S(660) = AC(660)/DC(660)

S(940) = AC(940)/DC(940)

This traditional instrument then calculates the ratio of these two arterial pulse-added absorbance signals:

R = S(660)/S(940)

This value of R is used to find the saturation SpO_2 in a look-up table built into the instrument's software. The values in the look-up table are based upon human blood studies against a laboratory co-oximeter on healthy adult volunteers in induced hypoxia studies.

This MS-7 assumes that arterio-venous shunting is highly variable and that fluctuating absorbance by venous blood is the major component of noise during the pulse. The MS-7 decomposes S(660) and S(940) into an arterial signal plus a noise component and calculates the ratio of the arterial signals without the noise:

S(660) = S1 + N1
S(940) = S2 + N2
R = S1/S2

Again, R is the ratio of two arterial pulse-added absorbance signals and its value is used to find the saturation SpO_2 in an empirically derived equation into the software. The values in the empirically derived equation are based upon human blood studies against a laboratory co-oximeter on healthy adult volunteers in induced hypoxia studies.

The above equations are combined and a noise reference (N') is determined:

N' = S(660) - S(940) x R

If there is no noise N' = 0: then S(660) = S(940) x R which is the same relationship for the traditional pulse oximeter.

The equation for the noise reference is based on the value of R, the value being seeked to determine the SpO_2. This instrument's software sweeps through possible values of R that correspond to SpO_2 values between 1% and 100% and generates an N' value for each of these R values. The S(660) and S(940) signals are processed with each possible N' noise reference through an adaptive correlation canceler (ACC) which yields an output power for each possible value of R (i.e., each possible SpO_2 from 1% to 100%). The result is a Discrete Saturation Transform (DST™) plot of relative output power versus possible SpO_2 value as shown in the following figure where R corresponds to SpO_2 = 97%:



The DST plot has two peaks: the peak corresponding to the higher saturation is selected as the SpO_2 value. This entire sequence is repeated once every two seconds on the most recent four seconds of raw data. The SpO_2 value therefore corresponds to a running average of arterial hemoglobin saturation that is updated every two seconds.

13.1.1 Precautions

⚠ **Warning** ⚠

This pulse wave from M-7 should NOT be used as an apnea monitoring.

⚠ **Warning** ⚠

This monitor with M-7 should be considered an early warning device for SpO2. As a trend towards patient deoxygenation is indicated, blood samples should be analyzed by a laboratory co-oximeter to completely understand the patient's condition.

⚠ **Warning** ⚠

If an alarm condition (other than exceptions listed herein) occurs while the alarm silence period is set to off, the only alarm indications will be visual displays and symbols related to the alarm condition.

⚠ **Warning** ⚠

Measure the monitor's leakage current whenever an external device is connected to the serial port. Leakage current must not exceed 100 microamperes.

⚠ **Warning** ⚠

To ensure patient electrical isolation, connect only to other equipment with electronically isolated circuits.

⚠ **Warning** ⚠

Do not connect to an electrical outlet controlled by a wall switch or dimmer.

⚠ **Warning** ⚠

As with all medical equipment, carefully route patient cabling to reduce the possibility of patient entanglement or strangulation.

⚠ **Warning** ⚠

Interfering Substances: Carboxyhemoglobin may erroneously increase readings. The level of increase is approximately equal to the amount of carboxyhemoglobin present. Dyes, or any substance containing dyes, that change usual arterial pigmentation may cause erroneous readings.

⚠ **Warning** ⚠

Do not use this instrument and the sensors during magnetic resonance imaging (MRI) scanning. Induced current could potentially cause burns. The monitor may affect the MRI image, and the MRI unit may affect the accuracy of the oximetry measurements.

⚠ **Warning** ⚠

Pulse oximetry can overestimate the SpO2 value in the presence of Hb-CO, Met-Hb or dye dilution chemicals.

⚠ **Warning** ⚠

Verify sensor cable fault detection before beginning of monitoring phase. Unplug the SpO_2 sensor cable from the socket, the screen will display the error message "SpO2 SENSOR OFF" and the audible alarm is activated.

⚠ **Warning** ⚠

Do not use the sterile supplied SpO_2 sensors if the packaging or the sensor is

damaged and return them to the vendor.

⚠ **Warning** ⚠

Prolonged and continuous monitoring may increase jeopardy of unexpected change of dermal condition such as abnormal sensitivity, rubescence, vesicle, repressive putrescence, and so on. It is especially important to check the sensor placement of neonate and patient of poor perfusion or immature dermogram by light collimation and proper attaching strictly according to changes of the skin. Check per 2~3 hours the sensor placement and move it when the skin deteriorates. More frequent examinations may be required for different patients.

⚠ **Note** ⚠

Do not perform SpO_2 measuring and NIBP measuring in same arm at one time, because obstruction of blood flow during NIBP measuring may adversely affect the reading of SpO_2 value.

⚠ **Note** ⚠

- **Make sure the nail covers the light window;**
- **The wire should be on the backside of the hand.**

⚠ **Note** ⚠

SpO_2 value always displays at the same position. Pulse Rate will display when HR FROM is set at "SpO2", “BOTH” in the ECG SETUP menu.

⚠ **Note** ⚠

SpO_2 waveform is not proportional to the pulse volume.

13.1.2 Monitoring Procedure

⚠ **Note** ⚠

This monitor with MS-7 is intended for continuous noninvasive monitoring of functional oxygen saturation of arterial hemoglobin (SpO2) and pulse rate (measured by an SpO2 sensor) for adult, pediatric, and neonatal patients in a hospital and mobile environment.

Limitations for Measurement

If the accuracy of any measurement does not seem reasonable, first check the patient’s vital

signs by alternate means and the check the instrument for proper functioning.

Inaccurate measurements may be caused by:

- incorrect sensor application or use
- significant levels of dysfunctional hemoglobins (e.g., carboxyhemoglobin or methemoglobin)
- intravascular dyes such as indocyanine green or methylene blue.
- exposure to excessive illumination, such as surgical lamps (especially ones with a xenon light source), bilirubin lamps, fluorescent lights, infrared heating lamps, or direct sunlight (exposure to excessive illumination can be corrected by covering the sensor with a dark or opaque material)
- excessive patient movement
- venous pulsations
- placement of a sensor on an extremity with a blood pressure cuff, arterial catheter, or intravascular line

the monitor can be used during defibrillation, but the readings may be inaccurate for a short time.

Loss of pulse signal can occur in any of the following situation:

- the sensor is too tight
- there is excessive illumination from light sources such as a surgical lamp, a bilirubin lamp, or sunlight
- a blood pressure cuff is inflated on the same extremity as the one with a SpO2 sensor attached.
- the patient has hypotension, severe vasoconstriction, severe anemia, or hypothermia
- there is arterial occlusion proximal to the sensor
- the patient is in cardiac arrest or is in shock

SpO_2 plethysmogram measurement: (MASIMO SET™ ONLY)

1. Switch on the monitor.
2. Attach the sensor to the appropriate site of the patient finger.
3. Plug the connector of the sensor extension cable into the SpO_2 socket on the PM-8000.



Figure 13-1 mounting of the sensor

SpO2 SETUP Menu

Pick the SpO2 hot key on the screen to call up the SpO2 SETUP menu as shown below.



Figure 13-2 SpO2 SETUP menu

⚠ Warning ⚠

Setting the SpO_2 upper alarm limit to 100% is equivalent to switching off the alarm on upper limit. High oxygen levels may predispose a premature infant to retrolental fibroplasia. Therefore, the upper alarm limit for oxygen saturation must be carefully selected in accordance with commonly accepted clinical practices.

SpO_2 alarm setting

- ALM: pick "ON" to enable prompt message and data record during the SpO_2 alarm; pick "OFF" to disable the alarm function, and there will be a beside "SpO_2".
- ALM REC: pick "ON" to enable report printing upon SpO_2 alarm.
- ALM LEV: used to set up alarm level, selectable from HIGH, MED and LOW. HIGH represents the most serious case.
- SpO2 ALM HI and SpO2 ALM LO:SpO2 alarm is activated when the result exceeds set SpO2 ALM HI value or falls below SpO2 ALM LO value. Use the knob to pick the SpO2 ALM HI or SpO2 ALM LO item and turn the knob to select the desired alarm limit.
- PR ALM HI and PR ALM LO: PR alarm is activated when the pulse rate exceeds set PR ALM HI value or falls below PR ALM LO value. Use the knob to pick the PR ALM HI or PR ALM LO item and turn the knob to select the desired alarm limit.

⚠ Warning ⚠

Check alarm limits each time the monitor is used to ensure that they are appropriate for the patient being monitored.

SpO_2 and PR alarm limits:

	Max. Upper Limit	Min. Lower Limit	Step
SpO_2	100	0	1
PR	240	25	1

- SWEEP
  Available options are 12.5, 25.0 mm/s.
- PR SOUND
  Pulse beep volume. The options are from “10” to “1”. “10” indicates the maximum volume while “1” the minimum. If your machine does not have "Multilevel Volume" function, only three options are available for volume, “3”, “2” and “1”.
- AVG TIME
  2~4S,4~6S, 8S,10S,12S,14S, 16S represent times that SpO_2 average value is counted.
- SENSITIVITY MODE
  Available options are normal and high.
- DEFAULT:
  Pick this item to access the SpO_2 DEFAULT CONFIG dialog box, in which the user may select whether the FACTORY DEFAULT CONFIG or the USER DEFAULT CONFIG is to be used. After selecting any of the items and exiting the dialog box, the system will pop up the dialog box asking for the user’s confirmation.

PITCH TONE function (OPTIONAL)

When SPO2 changes, if the “BEAT VOL” in “ECG SETUP” menu is set to a value other than “0” (which means “BEAT VOL” is switched ON), the heart beat volume will change automatically according to SPO2 value. This monitor has 20 kinds of PITCH TONE; the higher the SPO2 value is, the higher the PITCH TONE will be.

Although these 20 kinds of PITCH TONE could not be adjusted in menu, their volume could be controlled. For example, when “SPO2” is selected as “HR SOURCE” in “ECG SETUP” menu, the volume of PITCH TONE will be controlled by “PR SOUND” in “SPO2 SETUP” menu. If other item than “SPO2” is selected as “HR SOURCE” in “ECG SETUP” menu, the volume of PITCH TONE will be consequently controlled by “BEAT SOUND” in “ECG SETUP” MENU.

⚠**Note**⚠

When SPO2 module is switched OFF, PITCH TONE function will become disabled automatically.

13.1.3 Sensors and Accessories:

Before use, carefully read the LNOP sensor Directions for Use.
Use only Masimo oximetry sensors for SpO2 measurements. Other oxygen transducers or sensors may cause improper Radical Pulse Oximeter performance.

Tissue damage can be caused by incorrect application or use of an LNOP sensor, for example by wrapping the sensor too tightly. Inspect he sensor site as directed in he sensor Directions for Use to ensure skin integrity and correct positioning and adhesion of he sensor.

⚠ **Caution** ⚠

Do not use the damaged sensors. Do not use a sensor with exposed optical components. Do not immerse the sensor in water, solvents, or cleaning solutions (the sensors and connectors are not waterproof). Do not sterilize by irradiation, steam, or ethylene oxide.

⚠ **Caution** ⚠

Do not use damaged patient cables. Do not immerse the patient cables in water, solvents, or cleaning solutions (the sensors and connectors are not waterproof). Do not sterilize by irradiation, steam, or ethylene oxide.

- **Selecting a Masimo sensor:**

When selecting a sensor, consider, the patient's weight, the adequacy of perfusion, the available sensor sites, and the duration of monitoring. For more information refer to the following table or contact Masimo. Use only Masimo sensors and sensor cables. Select an appropriate sensor, apply it as directed, and observe all warnings and cautions presented in the directions for use accompanying the sensor.

High ambient light sources such as surgical lights(especially those with a xenon light source), bilirubin lamps, fluorescent lights, infrared heating lamps, and direct sunlight can interfere with the performance of an SpO2 sensor. To prevent interference from ambient light, ensure that the sensor is properly applied, and cover the sensor site with opaque material, if required. Failure to take this precaution in high ambient light conditions may result in inaccurate measurements.

SENSOR	USAGE	PATIENT WEIGHT
LNOP ADT	SINGLE USE	ADULTS > 30 kg
LNOP PDT	SINGLE USE	Adults >10 kg and <50 kg
LNOP NEO	SINGLE USE	Neonate <10 kg
LNOP NEO PT	SINGLE USE	Neonate <1 kg,or with poor skin integrity
LNOP DCI	REUSABLE	Adults and Pediatrics >30 kg
LNOP DCIP	REUSABLE	Pediatrics >10 kg and <50 kg
LNOP DCSC	REUSABLE	Adult and Pediatrics >30 kg,for spot check applications
LNOP Ear sensor	REUSABLE	Adult/Pediatric >30 kg
LNOP YI Multi-site	REUSABLE	Adult/Pediatric/Infant/Neonatal > 1 kg

Cleaning and reuse of Masimo LNOP sensors

Reusable sensors can be cleaned per the following procedure:

- Remove the sensor from the patient.
- Disconnect the sensor from the monitor.

- Wipe the entire sensor clean with a 70%isopropyl alcohol pad.
- Allow the sensor to air dry before returning it to operation.

Reattachment of single use adhesive sensors

- LNOP single use sensors may be reapplied to the same patient if the emitter and detector windows are clear and the adhesive still adheres to the skin.
- The adhesive can be partially rejuvenated by wiping with a 70%isopropyl alcohol wipe and allowing the sensor to thoroughly air dry prior to replacement on the patient.

⚠ **Warning** ⚠

To avoid cross contamination only use Masimo LNOP single use sensors on the same patient.

⚠ **Note** ⚠

If the sensor fails to track the pulse consistently, the sensors may be incorrectly positioned. Reposition the sensor or choose a different monitoring site.

⚠ **Caution** ⚠

Do not reprocess any LNOP single use sensors.

- **MASIMO SET PATIENT CABLES:**

Reusable patient cables of various lengths are available. All cables that display the Masimo SET logo are designed to work with any Masimo LNOP sensor and with any pulse oximeter or multiparameter instrument displaying the Masimo SET logo.
Only use Masimo oximetry patient cables for SpO2 measurements. Other patient cables may cause improper Radical pulse oximeter performance.

⚠ **Caution** ⚠

Carefully route patient cables to reduce the possibility of patient entanglement or strangulation.

⚠ **Caution** ⚠

Do not soak or immerse patient cables in any liquid solution. Do not sterilize patient cables by irradiation steam or ethylene oxide see the cleaning instructions in the directions for use for reusable Masimo patient cables.

- **Maintenance**

⚠ **Warning** ⚠

Before cleaning the monitor or the sensor, make sure that the equipment is switched off and disconnected from the power line.

For cleaning:

- Use a cotton ball or a soft mull moistened with hospital-grade ethanol to wipe the surface of the sensor, and then dry it with a cloth. This cleaning method can also be applied to the luminotron and receiving unit.
- The cable can be cleaned with 3% hydrogen dioxide, 70% isopropanol, or other active reagent. However, connector of the sensor shall not be subjected to such solution.

13.1.4 Alarm Description and Prompt

SpO2 Alarm Message

Among physiological alarms, those belonging to the type that the parameter has exceeded the limits may activate the recorder to automatically output the parameters and related measured waveforms when the alarms occur on the condition that the alarm record switch in the related menu is On.

TABLES BELOW DESCRIBE THE POSSIBLE PHYSIOLOGICAL ALARMS, TECHNICAL ALARMS AND PROMPT MESSAGES OCCURRING DURING SPO2 MEASUREMENT.

Physiological alarm:

Message	Cause	Alarm Level
SPO_2TOO HIGH	SpO2 measuring value is above upper alarm limit.	User-selectable
SpO2 TOO LOW	SpO2 measuring value is below lower alarm limit.	User-selectable
PR TOO HIGH	PR measuring value is above upper alarm limit.	User-selectable
PR TOO LOW	PR measuring value is below lower alarm limit.	User-selectable

Technical alarms:

Message	Possible Cause(s)	Alarm Level	Recommendation
SpO2 NO SENSOR	Sensor not fully inserted into the connector.	LOW	May be an incorrect sensor, or a defective sensor or cable. Insert sensor into the connector. Disconnect and reconnect sensor. Refer to the instructions for the sensor being used.
	Sensor inserted upside down.	LOW	Disconnect and reconnect he sensor with the logos matching.

SpO2 SENSOR OFF	SpO2 sensor may be disconnected from the patient or the monitor.	LOW	Disconnect and reconnect the sensor. Reattach sensor.
SpO2 SENSOR FAULT	This message appears when the sensor is faulty	HIGH	Stop using the measuring function of SpO2 module, notify biomedical engineer or our service staff.
SpO2 UNRECOGNIZED SENSOR	Masimo board does not recognize the sensor.	LOW	Make sure that the monitor and the patient are in correct connection with the cables.
SpO2 INCOMPATIBLE SENSOR	This message is displayed when the masimo sensor is finding incompatible sensor.	LOW	Make sure that the monitor use incompatible sensor.
SpO2 INTERFERENCE	Outside signal or energy preventing reading.	LOW	Remove outside interference.
SpO2 PULSE SEARCH	Unit is searching for the patients pulse.	LOW	If values are not displayed within 30 seconds, disconnect and reconnect sensor. If pulse search continues, remove sensor and replace on a better perfused site.
SpO2 LOW PERFUSION	Signal too small.	LOW	Move sensor to better perfused site.
SpO2 TOO MUCH LIGHT	Too much light on patient(sensor). Inadequate tissue covering sensor detector.	LOW	Remove or reduce lighting. Cover sensor from light. Reposition sensor.
SpO2 LOW SIGNAL IQ	Low signal quality.	LOW	Ensure proper sensor application. Mover sensor to a better perfused site.
SpO2 BOARD FAULT	This message appears when the Masimo Set board malfunctions.	HIGH	Stop using the measuring function of SpO2 module, notify biomedical engineer or our service staff.
SpO2 COMMUNICATION ERROR	This message is displayed when the front end module is having problems communicating (ie: framing errors or bad checksums) with the Masimo board.	HIGH	Stop using the measuring function of SpO2 module, notify biomedical engineer or our service staff.

SpO2 COMMUNICATION STOP	This message is displayed when the host can not receive the data from Masimo board for 5 seconds	HIGH	Stop using the measuring function of SpO2 module, notify biomedical engineer or our service staff.
SpO2 INIT ERR	This message is displayed when the SpO2 module initialization error happened.	HIGH	Stop using the measuring function of SpO2 module, notify biomedical engineer or our service staff.

Prompt message (include general alerts):

Message	Cause	Alarm Level
SpO2 EXCEED	SpO2 measuring value exceeds the range.	HIGH
PR EXCEED	PR measuring value exceeds the range.	HIGH
SEARCH PULSE	SpO2 module is searching for pulse.	No alarm
NO PULSE	SpO2 module cannot detect SpO2 signal for a long time.	HIGH

13.1.5 Masimo Information

The MASIMO SET® Product

Masimo Patents

This device is covered under one or more the following U.S. Patents: 5,482,036; 5,490,505; 5,632,272; 5,685,299; 5,758,644; 5,769,785; 6,002,952; 6,036,642; 6,067,462; 6,206,830, 6,157,850 and international equivalents. U.S.A. and international patents pending.

No Implied License

Possession or purchase of this device does not convey any express or implied license to use the device with replacement parts which would, alone, or in combination with this device, fall within the scope of one or more of the patents relating to this device.

13.2 PART 2 (MINDRAY SpO2 board configuration)

13.2.1 What is SpO2 Monitoring

SpO_2 Plethysmogram measurement is employed to determine the oxygen saturation of hemoglobin in the arterial blood. If, for example, 97% hemoglobin molecules in the red blood cells of the arterial blood combine with oxygen, then the blood has a SpO_2 oxygen saturation of 97%. The SpO_2 numeric on the monitor will read 97% .The SpO_2 numeric shows the percentage of hemoglobin molecules which have combined with oxygen molecules to form oxyhemoglobin. The SpO2/PLETH parameter can also provide a pulse rate signal and a plethysmogram wave.

How the SpO_2 / PLETH Parameter Works

- Arterial oxygen saturation is measured by a method called pulse oximetry. It is a continuous, non-invasive method based on the different absorption spectra of reduced hemoglobin and oxyhemoglobin. It measures how much light, sent from light sources on one side of the sensor, is transmitted through patient tissue (such as a finger or an ear), to a receiver on the other side.
  The sensor measurement wavelengths are nominally 660nm for the Red LED and 940nm for Infrared LED. Maximum optical power output for LED is 4 mW.
- The amount of light transmitted depends on many factors, most of which are constant. However, one of these factors, the blood flow in the arteries, varies with time, because it is pulsating. By measuring the light absorption during a pulsation, it is possible to derive the oxygen saturation of the arterial blood. Detecting the pulsation gives a PLETH waveform and pulse rate signal.
- The SpO_2 value and the PLETH waveform can be displayed on the main screen.
- SPO2 is a non-invasive measurement of the functional oxygen saturation.

⚠ **Warning** ⚠

Pulse oximetry can overestimate the SpO_2 value in the presence of Hb-CO, Met-Hb or dye dilution chemicals.

SpO2 / Pulse Monitoring

⚠ **Warning** ⚠

ES (Electrosurgery) equipment wire and SpO_2 cable must not be tangled up.

⚠ **Warning** ⚠

Do not put the sensor on extremities with arterial catheter or venous syringe.

⚠ **Note** ⚠

Do not perform SpO_2 measuring and NIBP measuring on same arm at one time, because obstruction of blood flow during NIBP measuring may adversely affect the reading of SpO_2 value.

13.2.2 Precautions during SpO2/Pulse Monitoring

⚠ **Note** ⚠

- **Make sure the nail covers the light window;**
- **The wire should be on the backside of the hand.**

⚠ **Note** ⚠

- **SpO_2 value is always displayed at the same position.**
- **Pulse Rate will be displayed only under following situations:**
  - **Select HR FROM as "SPO2" or "BOTH" in the ECG SETUP menu.**
  - **Select HR FROM as "AUTO" in the ECG SETUP menu and there is no ECG signal.**

⚠ **Note** ⚠

SpO_2 waveform is not proportional to the pulse volume.

⚠ **Warning** ⚠

Check if the sensor cable is in normal condition before monitoring. After unplugging the SpO_2 sensor cable from the socket, the system shall display the error message "SPO2 SENSOR OFF" and give the audible alarm.

⚠ **Warning** ⚠

Do not use the SpO_2 sensor once the package or the sensor is found damaged. Instead, you shall return it to the vendor.

⚠ **Warning** ⚠

Prolonged and continuous monitoring may increase jeopardy of unexpected change of dermal condition such as abnormal sensitivity, rubescence, vesicle, repressive putrescence, and so on. It is especially important to check the sensor placement of neonate and patient of poor perfusion or immature dermogram by light collimation and proper attaching strictly according to changes of the skin. Check per 2~3 hours the sensor placement and move it when the skin deteriorates. More frequent examinations may be required for different patients.

13.2.3 Monitoring Procedure

SpO_2 plethysmogram measurement

1. Switch on the monitor.
2. Attach the sensor to the appropriate site of the patient finger.
3. Plug the connector of the sensor extension cable into the SpO_2 socket on the SpO_2 module.



Figure 13-3 Mounting of the sensor

■ **Neonate SpO_2 Measurement**

The process of measuring neonate SpO_2 is similar to that of measuring adult SpO_2. Below is the description of neonate SpO_2 sensor and its installation.

1. Neonate SpO_2 sensor

Neonate SpO_2 sensor consists of Y-form SpO_2 sensor and its sheath. Insert the LED and PD ends of the Y-form SpO_2 sensor respectively into the upper and lower grooves on the sheath (figure 13-4). Figure 13-5 shows us the neonate SpO_2 sensor after insertion.



Figure 13-4 Neonate SpO_2 sensor (1)



Figure 13-5 Neonate SpO_2 sensor (2)

2. Attaching Neonate SpO_2 sensor

Wind the SpO_2 sensor around a hand or foot. Hold the sensor, pull the belt and fit one of its sides with “V” edge into the “V” groove on the corresponding side of the sheath. Appropriately elongate the belt (about 20mm) and fit the “V” edge of the other side of the belt into the “V” groove of the other side of the sheath and then loosen the belt. After the “V” edges of the two sides of the belt fit well into the “V” grooves on the two sides of the sheath, put the belt into the first lock bar to fasten the belt. See figure 13-6. If the belt is too much long, you may put it into the second lock bar. You must position the SpO_2 sensor in this way so as to make the photoelectric component face the correct position. In the mean time, note not to elongate the belt too much, which may lead to inaccurate measurement and also blocking the blood circulation severely.



Figure 13-6 mounting of the neonate sensor

⚠ **Note** ⚠

If the sensor cannot be positioned accurately to the part to be measured, it may result in inaccurate SpO_2 reading, or even that the SpO_2 cannot be measured because no pulse is detected. If this is true, you must position the sensor again.

The excessive patient movement may result in inaccurate reading. In this situation, you must keep the patient quiet or change the part for monitoring to reduce the adverse influence of excessive movement.

⚠ **Warning** ⚠

In the process of extended and continuous monitoring, you should check the

peripheral circulation and the skin every 2 hours. If any unfavorable changes take place, you should change the measured position in time.

In the process of extended and continuous monitoring, you should periodically check the position of the sensor. In case that the position of the sensor moves during monitoring, the measurement accuracy may be affected.

13.2.4 Limitations for Measurement

Measurement Limitations

In operation, the accuracy of oximetry readings can be affected by:

- **High-frequency electrical noise, including noise created by the host system, or noise from external sources, such as electrosurgical apparatus connected to the system.**
- **Do not use oximeters and oximetry sensors during magnetic resonance imaging (MRI) scanning. Induced current could potentially cause burns.**
- **Intravascular dye injections**
- **Excessive patient movement**
- **External light radiation**
- **Improper sensor installation or incorrect contact position of the patient**
- **Sensor temperature (optimal temperature between 28° C and 42° C)**
- **Placement of the sensor on an extremity that has a blood pressure cuff, arterial catheter, or intravascular line.**
- **Significant concentrations of dysfunctional hemoglobin, such as carboxyhemoglobin and methemoglobin.**
- **SpO_2 too low**
- **Bad circular injection of the part being measured**
- **Shock, anemia, low temperature and application of vasomotor may all cause the arterial blood flow to reduce and hence make the measurement impossible.**
- **The absorption of oxyhemoglobin (HbO2) and deoxyhemoglobin to the light of special wavelength may also affect SpO_2 measurement. If there exist other objects (carbon hemoglobin, methemoglobin, methylene blue and indigo carmine) absorbing the light of the same wavelength, they may result in false or low SpO_2 value.**
- **It is recommended to use SpO_2 sensors described in chapter Accessories and Ordering Information.**

13.2.5 SpO2 Menu

SPO2 SETUP Menu

Turn the knob to move the cursor onto the SPO2 hot key in the Parameter area, push the knob to access the SPO2 SETUP menu.



Figure 13-7 SPO2 SETUP menu

⚠ **Warning** ⚠

Setting the SpO_2 upper alarm limit to 100% is equivalent to switching off the alarm on upper limit. High oxygen levels may predispose a premature infant to retrolental fibroplasia. Therefore, the upper alarm limit for oxygen saturation must be carefully selected in accordance with commonly accepted clinical practices.

SpO_2 alarm setting

- ALM: pick "ON", the system will give alarm prompt and store alarm information when SpO_2 alarm occurs; pick "OFF", the system will not give alarm and instead display a beside "SpO_2".
- ALM REC: pick "ON", the system will command the recorder to output alarm information when SpO_2 alarm occurs.
- ALM LEV: used to set up alarm level, selectable from HIGH, MED and LOW. HIGH represents the most serious case.
- SPO2 ALM HI and SPO2 ALM LO: SpO_2 alarm is activated when the result exceeds set SPO2 ALM HI value or falls below SPO2 ALM LO value.
- PR ALM HI and PR ALM LO: PR alarm is activated when the pulse rate exceeds set PR ALM HI value or falls below PR ALM LO value.

SpO2 and PR alarm limits:

	Max. Upper Limit	Min. Lower Limit	Step
SpO2	100	0	1
PR	254	0	1

The default SpO2 and PR alarm limits:

Parameters		Max. Upper Limit	Min. Lower Limit
SpO2	Adult	100	90
	Pediatric	100	90
	Neonatal	95	80
PR	Adult	120	50
	Pediatric	160	75
	Neonatal	200	100

- SWEEP
  Available options are 12.5mm/s, 25.0 mm/s.
- PR SOUND
  Pulse beep volume. The options are from “10” to “1”. “10” indicates the maximum volume while “1” the minimum. If your machine does not have "Multilevel Volume" function, only three options are available for volume, “3”, “2” and “1”.
- AVG TIME
  2~4S,4~6S, 8S,10S,12S,14S, 16S represent times that SpO_2 average value is counted.
- DEFAULT:
  Pick this item to access the SpO_2 DEFAULT CONFIG dialog box, in which the user may select whether the FACTORY DEFAULT CONFIG or the USER DEFAULT CONFIG is to be used. After selecting any of the items and exiting the dialog box, the system will pop up the dialog box asking for the user’s confirmation.

PITCH TONE function(OPTIONAL)

When SPO2 changes, if the “BEAT VOL” in “ECG SETUP” menu is set to a value other than “0” (which means “BEAT VOL” is switched ON), the heart beat volume will change automatically according to SPO2 value. This monitor has 20 kinds of PITCH TONE; the higher the SPO2 value is, the higher the PITCH TONE will be.

Although these 20 kinds of PITCH TONE could not be adjusted in menu, their volume could be controlled. For example, when “SPO2” is selected as “HR SOURCE” in “ECG SETUP” menu, the volume of PITCH TONE will be controlled by “PR SOUND” in “SPO2 SETUP” menu. If other item than “SPO2” is selected as “HR SOURCE” in “ECG SETUP” menu, the volume of PITCH TONE will be consequently controlled by “BEAT SOUND” in “ECG SETUP” MENU.

⚠**Note**⚠

When SPO2 module is switched OFF, PITCH TONE function will become disabled automatically.

13.2.6 Alarm Description and Prompt

SpO2 Alarm Message

When the alarm switches are set to ON in relevant menus, the physiological alarms caused by the parameter exceeding the alarm limit may possibly trigger the recorder to automatically output alarming parameter value and corresponding waveforms.

TABLES BELOW DESCRIBE THE POSSIBLE PHYSIOLOGICAL ALARMS, TECHNICAL ALARMS AND PROMPT MESSAGES OCCURRING DURING SPO_2 MEASUREMENT.

Physiological alarm:

Message	Cause	Alarm Level
SPO2 TOO HIGH	SpO_2 measuring value is above upper alarm limit.	User-selectable
SpO2 TOO LOW	SpO2 measuring value is below lower alarm limit.	User-selectable
PR TOO HIGH	PR measuring value is above upper alarm limit.	User-selectable
PR TOO LOW	PR measuring value is below lower alarm limit.	User-selectable

Technical alarms:

Message	Cause	Alarm Level	Remedy
SPO2 SENSOR OFF	SpO_2 sensor may be disconnected from the patient or the monitor.	LOW	Make sure that the monitor and the patient are in correct connection with the cables.
SPO2 INIT ERR	SpO_2 module failure	HIGH	Stop using the measuring function of SpO_2 module, notify biomedical engineer or Mindray service staff.
SPO2 INIT ERR 1			
SPO2 INIT ERR 2			
SPO2 INIT ERR 3			
SPO2 INIT ERR 4			
SPO2 INIT ERR 5			
SPO2 INIT ERR 6			
SPO2 INIT ERR 7			
SPO2 INIT ERR 8			
SPO2 COMM STOP	SpO_2 module failure or communication error	HIGH	Stop using the measuring function of SpO_2 module, notify biomedical engineer or Mindray service staff.

SPO2 COMM ERR	SpO_2 module failure or communication error	HIGH	Stop using the measuring function of SpO_2 module, notify biomedical engineer or Mindray service staff.
SPO2 ALM LMT ERR	Functional safety failure	HIGH	Stop using the measuring function of SpO_2 module, notify biomedical engineer or Mindray service staff.
PR ALM LMT ERR	Functional safety failure	HIGH	Stop using the measuring function of SpO_2 module, notify biomedical engineer or Mindray service staff.

Prompt message (include general alerts):

Message	Cause	Alarm Level
SPO2 EXCEED	SpO_2 measuring value exceeds the range.	HIGH
PR EXCEED	PR measuring value exceeds the range.	HIGH
SEARCH PULSE	SpO_2 module is searching for pulse.	No alarm
NO PULSE	SpO_2 module cannot detect SpO_2 signal for a long time.	HIGH

13.2.7 Maintenance and Cleaning

Care and Cleaning

⚠ **Warning** ⚠

Turn of the monitor and disconnect the line power before cleaning the monitor or the sensor

⚠ **Warning** ⚠

Do not subject the sensor to autoclaving.

Do not immerse the sensor into any liquid.

Do not use any sensor or cable that may be damaged or deteriorated.

Cleaning:

- Use a cotton ball or a soft mull moistened with hospital-grade ethanol to wipe the surface of the sensor, and then dry it with a cloth. This cleaning method can also be applied to the luminotron and receiving unit.
- The cable can be cleaned with 3% hydrogen dioxide, 70% isopropanol, or other active reagent. However, connector of the sensor shall not be subjected to such solution.

Chapter 14 NIBP Monitoring

14.1 Introduction

- Reference to the European standard EN 1060-1: Specification for Non-invasive sphygmomanometers Part 1, General requirements.
- The Non-invasive Blood Pressure (NIBP) module measures the blood pressure using the oscillometric method.
- It is applicable for adult, pediatric, and neonatal usage.
- There are three modes of measurement available: manual, automatic and continuous. Each mode displays the diastolic, systolic and mean blood pressure.
  - In the MANUAL mode, only one measurement is conducted for each time.
  - In the AUTO mode, the measurement is cycled; you can set the interval time to 1/2/3/4/5/10/15/30/60/90/120/180/240/480 minutes.
  - In the continuous mode, the monitor measures the blood pressure as many times as possible in five minutes.

⚠ **Warning** ⚠

1. **You must not perform NIBP measurements on patients with sickle-cell disease or under any condition which the skin is damaged or expected to be damaged.**
2. **For a thrombasthemia patient, it is important to determine whether measurement of the blood pressure shall be done automatically. The determination should be based on the clinical evaluation.**
3. **Ensure that the correct setting is selected when performing measurements on children. It may be dangerous for the children to use an over pressure level.**

14.2 NIBP Monitoring

14.2.1 NIBP Measuring

⚠ **Warning** ⚠

- **Before starting a measurement, verify that you have selected a setting appropriate for your patient (adult, pediatric or neonate.)**
- **Do not apply the cuff to a limb that has an intravenous infusion or catheter in place. This could cause tissue damage around the catheter when infusion is slowed or blocked during cuff inflation.**

⚠ Warning ⚠

Make sure that the air conduit connecting the blood pressure cuff and the monitor is neither blocked nor tangled.

1. Plug in the air hose and switch on the system.
2. Apply the blood pressure cuff to the patient's arm or leg following the instructions below (Figure 14-1).

■ Ensure that the cuff is completely deflated.

■ Apply the appropriate size cuff to the patient, and make sure that the symbol "Φ" is over the appropriate artery. Ensure that the cuff is not wrapped too tightly around the limb. Excessive tightness may cause discoloration and eventual ischemia of the extremities.



Figure 14-1 Applying Cuff

⚠ Note ⚠

The width of the cuff should be either 40% of the limb circumference (50% for neonates) or 2/3 of the upper arm length. The inflatable part of the cuff should be long enough to encircle 50-80% of the limb. The wrong size of cuff can cause erroneous readings. If the cuff size is in question, then use a larger cuff.

Size of reusable cuff for neonate/children/adult

Patient Type	Limb perimeter	Cuff width	Hose
Infant	10 ~19 cm	8 cm	1.5 m or 3 m
Child	18 ~ 26 cm	10.6 cm	
Adult	25 ~ 35 cm	14 cm	
Large Adult	33 ~ 47 cm	17 cm	
Thigh	46 ~ 66 cm	21 cm	

Size of disposable cuff for neonate/children/adult

Size No.	Limb perimeter	Cuff width	Hose
1	3.1 ~ 5.7 cm	2.5 cm	1.5 m or 3 m
2	4.3 ~ 8.0 cm	3.2 cm	
3	5.8 ~ 10.9 cm	4.3 cm	
4	7.1 ~ 13.1 cm	5.1 cm	

- Make sure that the cuff edge falls within the range of mark <->. If it does not, use a larger or smaller cuff that fits better.

3. Connect the cuff to the air hose. The limb chosen for taking the measurement should be placed at the same level as the patient's heart. If this is not possible you should apply the following corrections to the measured values:
- If the cuff is placed higher than the heart level, add 0.75 mmHg (0.10 kPa) for each centimeter of difference.
- If it is placed lower than the heart level, deduct 0.75 mmHg (0.10 kPa) for each centimeter of difference.

4. Check whether the patient mode is appropriately selected. Access PATIENT SETUP menu from SYSTEM MENU and pick PAT TYPE item and turn the knob to select the required patient type.
5. Select a measurement mode in the NIBP SETUP menu. Refer to the following paragraphs **Operation Hints** for details
6. Press the NIBP button on the front panel to start a measurement.

Operation Hints

1. To start auto measuring:
   Access NIBP SETUP menu and pick the INTERVAL item, in which the user may choose the selections other than MANUAL to set up the time interval for auto measurement. After that, press NIBP button on the front panel to start the auto measuring according to the selected time interval.

⚠ Warning ⚠

Prolonged non-invasive blood pressure measurements in Auto mode may be associated with purport, ischemia and neuropathy in the limb wearing the cuff. When monitoring a patient, examine the extremities of the limb frequently for normal color, warmth and sensitivity. If any abnormality is observed, stop the blood pressure measurements.

2. To stop auto measuring:
   During auto measuring press NIBP button on the front panel at any time to stop auto measurement.
3. To start a manual measuring:
   - Access NIBP SETUP menu and pick the INTERVAL item. Select the MANUAL selection. Then press the NIBP button on the front panel to start a manual measurement.
   - During the idle period of auto measuring process, press the NIBP button on the front panel at any time to start a manual measurement. Then press the NIBP button on the front panel to stop manual measurement and the system continues executes auto-measuring program according to selected time interval.

4. To start a manual measuring during the AUTO mode:
   Press NIBP button on the front panel.
5. To stop a manual measuring
   Repress the NIBP button on the front panel again.
6. To perform continuous measuring:
   Access NIBP SETUP menu and pick the CONTINUAL item to start the continuous measurement. The monitor will measure as many times of NIBP as possible within 5 minutes.

⚠ **Warning** ⚠

Prolonged non-invasive blood pressure measurements in continual mode may be associated with purport, ischemia and neuropathy in the limb wearing the cuff. When monitoring a patient, examine the extremities of the limb frequently for normal color, warmth and sensitivity. If any abnormality is observed, stop the blood pressure measurements.

7. To stop continuous measuring:
   During continuous measuring press NIBP button on the front panel at any time to stop continuous measurement.

⚠ **Note** ⚠

If you are in doubt about the accuracy of any reading(s), check the patient's vital signs by an alternative method before checking the functioning of the monitor.

⚠ **Warning** ⚠

If liquid is inadvertently splashed on the equipment or its accessories, or may enter the conduit or inside the monitor, contact local Customer Service Center.

Measurement Limitations

To different patient conditions, the oscillometric measurement has certain limitations. The measurement is in search of regular arterial pressure pulse. In those circumstances when the patient's condition makes it difficult to detect, the measurement becomes unreliable and measuring time increases. The user should be aware that the following conditions could interfere with the measurement, making the measurement unreliable or longer to derive. In some cases, the patient's condition will make a measurement impossible.

- **Patient Movement**

Measurements will be unreliable or may not be possible if the patient is moving, shivering or having convulsions. These motions may interfere with the detection of the arterial pressure pulses. In addition, the measurement time will be prolonged.

- **Cardiac Arrhythmia's**

Measurements will be unreliable and may not be possible if the patient's cardiac arrhythmia has caused an irregular heartbeat. The measuring time thus will be prolonged.

- **Heart-lung Machine**

Measurements will not be possible if the patient is connected to a heart-lung machine.

- **Pressure Changes**

Measurements will be unreliable and may not be possible if the patient's blood pressure is changing rapidly over the period of time during which the arterial pressure pulses are being analyzed to obtain the measurement.

- **Severe Shock**

If the patient is in severe shock or hypothermia, measurements will be unreliable since reduced blood flow to the peripheries will cause reduced pulsation of the arteries.

- **Heart Rate Extremes**

Measurements can not be made at a heart rate of less than 40 bpm and greater than 240 bpm.

14. 2. 2 NIBP monitoring screen

NIBP measurement result and corresponding message are displayed as follows:



14.3 NIBP SETUP menu

Pick the NIBP hot key on the screen to call up the NIBP menu shown as below:



Figure 14-2 NIBP SETUP Menu

ν NIBP alarm setting

- ALM: pick "ON" to enable prompt message and data record during the NIBP alarm; pick "OFF" to disable the alarm function, and there will be a beside “NIBP” .
- ALM LEV: selectable from HIGH, MED to LOW. HIGH represents the most serious case.
- ALM REC: pick "ON" to enable report printing upon NIBP alarm.
- SYS ALM HI, SYS ALM LOW, MEAN ALM HI, MEAN ALM LO, DIA ALM HI, DIA ALM LO are for the user to set up the alarm limit for each type of pressure. NIBP alarm is activated when the pressure exceeds set upper alarm limits or falls below lower alarm limits.

NIBP alarm limits:

Adult Mode

SYS	40-270 mmHg
DIA	10-215 mmHg
Mean	20-235 mmHg

Pediatric Mode

SYS	40-200 mmHg
DIA	10-150 mmHg
Mean	20-165 mmHg

Neonatal Mode

SYS	40-135 mmHg
DIA	10-100 mmHg
Mean	20-105 mmHg

- RESET

  Restore measurement status.

  Pick this item to restore initial settings of the pressure pump.

  When the pressure does not work properly and the system fails to give message for the problem, pick this item to activate self-test procedure, thus restore the system from abnormal performance.

- CONTINUAL

  Start continuous measuring.

  When this item is picked, the menu will disappear automatically.

- INTERVAL

  Interval time for automatic measuring. Available selections:

  1/2/3/4/5/10/15/30/60/90/120/180/240/480 minutes. Press START/STOP button on the NIBP module to start the first auto measuring.

  Pick MANUAL selection in INTERVAL item to set up the measuring mode to MANUAL.

- UNIT

  Pick this item to set measurement unit. (Option: mmHg or kPa)

- CALIBRATE

  Calibrate the cuff pressure reading with a calibrated reference manometer. Pick the CALIBRATE item to start the calibration and the item will change into STOP CAL, which if picked, the system will stop calibration.

- DEFAULT

  Pick this item to access the NIBP DEFAULT CONFIG dialog box, in which the user may select whether the FACTORY DEFAULT CONFIG or the USER DEFAULT CONFIG is to be used. After selecting any of the items and exiting the dialog box, the system will pop up the dialog box asking for the user's confirmation.

⚠Warning⚠

The calibration of the NIBP measurement is necessary for every two years (or as frequently as dictated by your Hospital Procedures Policy). The performance should be checked according to the following details.

Procedure of the Pressure Transducer Calibration:

Replace the cuff of the device with a rigid metal vessel with a capacity of 500 ml ± 5%. Connect a calibrated reference manometer with an error less than 0.8 mmHg and a ball pump by means of a T-piece connector and hoses to the pneumatic system. Set the monitor in **CALIBRATE** mode. Inflate the pneumatic system to 0, 50 and 200 mmHg by ball pump separately. The difference between the indicated pressure of the reference manometer and the indicated pressure of the monitor will not exceed 3 mmHg. Otherwise, please contact our customer service.



Figure 14-3 Diagram of NIBP calibration

■ PNEUMATIC

This item is used for air leakage test. Turn the knob to pick the item to start the air leakage test. Then the item will change into STOP PENUM, which if picked, the system will stop air leakage test.

⚠Warning⚠

This pneumatic test other than being specified in the EN 1060-1 standard is to be used by the user to simply determine whether there are air leaks in the NIBP airway. If at the end of the test the system gives the prompt that the NIBP airway has air leaks, please contact the manufacturer for repair.

Procedure of the air leakage test:

1) Connect the cuff securely with the socket for NIBP air hole.
2) Wrap the cuff around the cylinder of an appropriate size.
3) Access the NIBP SETUP menu.
4) Turn the knob to the PNEUMATIC item and press the knob. Then the prompt “Pneum testing…” will appear on the bottom of the NIBP parameter area indicating that the system has started performing pneumatic test.
5) The system will automatically Inflate the pneumatic system to about 180mmHg.
6) After 20 seconds or so, the system will automatically open the deflating valve, which marks the completion of a pneumatic measurement.
7) If no prompt appears on the bottom of the NIBP parameter area, it indicates that the airway is in good situation and no air leaks exist. However if the prompt “PNEUMATIC LEAK” appears in the place, it indicates that the airway may have air leaks. In this case, the user should check for loose connection. After confirming secure connections, the user should re-perform the pneumatic test. If the failure prompt still appears, please contact the manufacturer for repair.



Figure 14-4 Diagram of NIBP air leakage test

14.4 NIBP Alarm Message

Among physiological alarms, those belonging to the type that the parameter has exceeded the limits may activate the recorder to automatically output the parameters and related measured waveforms when the alarms occur on the condition that the alarm record switch in the related menu is On.

Tables below describe the possible physiological alarms, technical alarms and prompt messages occurring during NIBP measurement.

Physiological alarms:

Message	Cause	Alarm Level
NS TOO HIGH	NIBP SYS measuring value is above upper alarm limit.	User-selectable
NS TOO LOW	NIBP SYS measuring value is below lower alarm limit.	User-selectable
ND TOO HIGH	NIBP DIA measuring value is above upper alarm limit.	User-selectable
ND TOO LOW	NIBP DIA measuring value is below lower alarm limit.	User-selectable
NM TOO HIGN	NIBP MAP measuring value is above upper alarm limit.	User-selectable
NM TOO LOW	NIBP MAP measuring value is below lower alarm limit.	User-selectable

Technical alarms 1: (display in information area)

Message	Cause	Alarm Level	Remedy
NS ALM LMT ERR	Functional safety failure	HIGH	Stop using alarming functions of NIBP module and notify biomedical engineer or Mindray service staff.
NM ALM LMT ERR	Functional safety failure	HIGH	Stop using alarming functions of NIBP module and notify biomedical engineer or Mindray service staff.
ND ALM LMT ERR	Functional safety failure	HIGH	Stop using alarming functions of NIBP module and notify biomedical engineer or Mindray service staff.

Technical alarms 2: (display in the area below the NIBP value)

Message	Cause	Alarm Level	Remedy

NIBP SELF TEST ERR	Sensor or other hardware of NIBP module is incorrect.	HIGH	Stop using measuring function of NIBP module, notify biomedical engineer or Mindray service staff.
NIBP COMM ERR	Communication with NIBP module is failed.	HIGH	If failure persists, stop using measuring function of NIBP module, notify biomedical engineer or Mindray service staff.
LOOSE CUFF	Cuff is no properly wrapped or no cuff exists.	LOW	Properly wrap the cuff
AIR LEAK	Cuff, hose or connector is damaged.	LOW	Check and replace the leaking parts, if required, notify biomedical engineer or Mindray service staff.
AIR PRESSURE ERROR	Stable pressure value is not available. e.g. hoses are tangled.	LOW	Check if the hoses are tangled, if failure persists, notify biomedical engineer or Mindray service staff.
WEAK SIGNAL	Cuff is too loose or patient pulse is too weak.	LOW	Use other method to measure blood pressure.
RANGE EXCEEDED	Measuring range exceeds the specified upper limit.	HIGH	Reset NIBP module, if failure persists, stop using measuring function of NIBP module, notify biomedical engineer or Mindray service staff.
EXCESSIVE MOTION	Affected by arm motion, signal noise is too large or pulse rate is not regular.	LOW	Make sure that the patient under monitoring is motionless.
OVER PRESSURE	Pressure has exceeded the specified upper safety limit.	HIGH	Measure again, if failure persists, stop using measuring function of NIBP module and notify biomedical engineer or Mindray service staff.
SIGNAL STURATED	Excessive motion	LOW	Stop the patient from moving.
PNEUMATIC LEAK	During pneumatic test, leak is detected.	LOW	Check and replace the leaking parts, if required, notify biomedical engineer or Mindray service staff.
NIBP SYSTEM FAILURE	Operation of blood pressure pump system is failed.	HIGH	Stop using measuring function of NIBP module, notify biomedical engineer or Mindray service staff.
CUFF TYPE ERR	Cuff type does not comply with the patient type.	LOW	Select appropriate cuff type
NIBP TIME OUT	Measuring time has exceeded 120 seconds (adult) or 90 seconds (neonatal).	HIGH	Measure again or use other measuring method.
NIBP ILLEGALLY RESET	Abnormal module reset	HIGH	Reset again
MEASURE FAIL	Problem happens when measuring the curve. The system cannot perform measurement, analysis or calculation.	HIGH	Check the cuff. Make sure that the patient under monitoring is motionless. Measure again.

Prompt message: (display in the prompt area below NIBP value)

Message	Cause	Alarm Level
Manual measure…	During manual measuring mode.	No alarm
Cont measuring…	During continuous measuring mode.	
Auto measuring…	During automatic measuring mode.	
Please start	After selecting interval between measurements in MENU	
Measurement over	Press START/STOP key during measuring to stop measurement	
Calibrating…	During calibrating	
Calibration over	Calibration over	
Pneum testing…	During pneumatic test	
Pneum test over	pneumatic test over	
Resetting...	NIBP module in resetting	
Reset failed	NIBP module reset failed	

14.5 Maintenance and Cleaning

⚠ **Warning** ⚠

- **Do not squeeze the rubber tube on the cuff.**
- **Do not allow liquid to enter the connector socket at the front of the monitor.**
- **Do not wipe the inner part of the connector socket when cleaning the monitor.**
- **When the reusable cuff is not connected with the monitor, or being cleaned, always place the cover on the rubber tube to avoid liquid permeation.**

Reusable Blood Pressure Cuff

The cuff can be sterilized by means of conventional autoclaving, gas, or radiation sterilization in hot air ovens or disinfected by immersion in decontamination solutions, but remember to remove the rubber bag if you use this method. The cuff should not be dry-cleaned.

The cuff can also be machine-washed or hand-washed, the latter method may prolong the service life of the cuff. Before washing, remove the latex rubber bag, and for machine-washing, close the Velcro fastening. Allow the cuff to dry thoroughly after washing, then reinsert the rubber bag.



Figure 14-5 Replace Rubber Bag in Cuff

To replace the rubber bag in the cuff, first place the bag on top of the cuff so that the rubber tubes line up with the large opening on the long side of the cuff. Now roll the bag lengthwise and insert it into the opening on the long side of the cuff. Hold the tubes and the cuff and shake the complete cuff until the bag is in position. Thread the rubber tubes from inside the cuff, and out through the small hole under the internal flap.

Disposable Blood Pressure Cuffs

Disposable cuffs are intended for one-patient use only. Do not use the same cuff on any other patient. Do not sterilize or use autoclave on disposable cuffs. Disposable cuffs can be cleaned using soap solution to prevent infection.

⚠ Note ⚠

For protecting environment, the disposable blood pressure cuffs must be recycled or disposed of properly.

Chapter 15 TEMP Monitoring

15.1 TEMP Monitoring

TEMP monitoring setup

- If you are using disposable TEMP probes you need to plug the TEMP cable into the monitor and then connect the probe to the cable. With a reusable TEMP probe you can plug the probe directly into the monitor
- Apply the TEMP probe(s) securely to the patient.
- Switch on the system.

⚠ **Warning** ⚠

Verify probe cables fault detection before beginning of monitoring phase. Unplug the temperature probe cable from the socket, the screen will display the error message "TEMP SENSOR OFF" and the audible alarm is activated.

⚠**Warning**⚠

If use two TEMP probes, them must be made by the same manufacture.

⚠ **Note** ⚠

Disposable TEMP probe can only be used once for one patient.

⚠ **Warning** ⚠

The calibration of the temperature measurement is necessary for every two years (or as frequently as dictated by your Hospital Procedures Policy). When you need calibrate the temperature measurement, contact the manufacture please.

⚠ **Note** ⚠

The self-test of the temperature measurement is performed automatically once per hour during the monitoring. The test procedure lasts about 2 seconds and does not affect the normal measurement of the temperature monitoring.

15.2 TEMP SETUP Menu

Pick the TEMP hot key on the screen to call up the TEMP SETUP menu shown as below:



Figure 15-1 TEMP SETUP Menu

- TEMP alarm setting
  - ALM: pick "ON" to enable prompt message and data record during the TEMP alarm; pick "OFF" to disable the alarm function, and prompt the symbol beside TEMP numeric.
  - ALM LEV: used to set up the alarm level, selectable from HIGH, MED or LOW.
  - ALM REC: used to start/stop recording TEMP alarms. Pick "ON" to enable report printing upon TEMP alarm.

  Alarm for TEMP occurs when the measured temperature exceeds set alarm high limit or falls below alarm low limit.

TEMP alarm limits:

	Max. TEMP HI	Min. TEMP LO	Step
TEMP	50	0	0.1

- UNIT

  To set temperature unit (ºC or ºF).
- DEFAULT

  Pick this item to access the TEMP DEFAULT CONFIG dialog box, in which the user may select whether the FACTORY DEFAULT CONFIG or the USER DEFAULT CONFIG is to be used. After selecting any of the items and exiting the dialog box, the system will pop up the dialog box asking for the user's confirmation.

15.3 TEMP Alarm message

Among physiological alarms, those belonging to the type that the parameter has exceeded the limits may activate the recorder to automatically output the parameters and related

measured waveforms when the alarms occur on the condition that the alarm record switch in the related menu is On.

Tables below describe the possible physiological alarms, technical alarms and prompt messages occurring during TEMP measurement.

Physiological alarms:

Message	Cause	Alarm Level
TEMP TOO HIGH	Measuring value of sensor is above upper alarm limit.	User-selectable
TEMP TOO LOW	Measuring value of sensor is below lower alarm limit.	User-selectable

Technical alarms:

Alarm Message	Cause	Alarm Level	Remedy
TEMP SENSOR OFF	Temperature cable may be disconnected from the monitor.	LOW	Make sure that the cable is properly connected.
TEMP ALM LMT ERR	Functional safety failure	HIGH	Stop using alarming function of TEMP module, notify biomedical engineer or Mindray service staff.

Prompt message:

Message	Cause	Alarm Level
TEMP EXCEED	Measuring value of sensor is beyond measuring range.	HIGH

15.4 Care and Cleaning

⚠ **Warning** ⚠

Before cleaning the monitor or the probe, make sure that the equipment is switched off and disconnected from the power line.

Reusable TEMP Probes

1 The TEMP probe should not be heated above 100℃ (212℉). It should only be subjected briefly to temperatures between 80℃ (176℉) and 100℃ (212℉).

2 The probe must not be sterilized in steam.

3 Only detergents containing no alcohol can be used for disaffection.

4 The rectal probes should be used, if possible, in conjunction with a protective rubber cover.

5 To clean the probe, hold the tip with one hand and with the other hand rubbing the probe down in the direction of the connector using a moist lint-free cloth.

⚠ Note ⚠

Disposable TEMP probe must not be re-sterilized or reused.

⚠ Note ⚠

For protecting environment, the disposable TEMP probe must be recycled or disposed of properly.

Chapter 16 IBP Monitoring

16.1 Introduction

This chapter introduces IBP measurement, maintenance and cleaning of relevant accessories.

The Monitor measures blood pressure (SYS, DIA and MAP) of selected blood vesse.

The available pressure labels are:

Label	Definition
ART	Arterial Blood Pressure
PA	Pulmonary Arterial Pressure
CVP	Center Venous Pressure
RAP	Right Atrial Pressure
LAP	Left Atrial Pressure
ICP	Intracranial Pressure (ICT/B Transducer information Refer to **16.7**)
P1,P2	Expand Pressure

16.2 Precautions during IBP Monitoring

⚠ **Warning** ⚠

Parts and accessories used must meet the safety requirements of the medical electrical equipment standards.

⚠ **Warning** ⚠

Do not contact the metal part connected to the electrical appliance when connecting or using the accessory.

⚠ **Warning** ⚠

When the monitor is used with HF surgical equipment, do not let the transducer and cable contact the HF surgical equipment to prevent the patient from burning caused by leakage current.

⚠ **Warning** ⚠

Disposable IBP transducer or domes should not be reused.

⚠ **Note** ⚠

Use only the pressure transducer specified in this operation manual.

The specified transducer （except for ICT/B transducer）has the function of protecting against the electric shock (especially the leakage current) and the influence of cardiac defibrillator. It can be used in surgical operation. When the patient is in the defibrillation, the pressure waveform may become temporarily distorted. However the monitor will work normally after defibrillation with the operation mode and user configuration being not affected.

⚠ **Warning** ⚠

Inspect the transducer cable is in normal condition before monitoring. Unplug the transducer of the channel 1, the monitor should display the error message "IBP: SENSOR OFF" and trigger audible alarm. The other channel should act the same.

⚠ **Note** ⚠

Periodically calibrate the transducer either new or used according to the Hospital Regulation.

⚠ **Warning** ⚠

If any kind of liquid, other than the solution to be infused in the pressure line or transducer, is splashed on the equipment or its accessories, especially enters the transducer or the monitor, contact the Service Center of the Hospital immediately.

16.3 Monitoring Procedure

Preparation before IBP measurement:

1. Plug the pressure cable into corresponding socket and check that the monitor is switched on.
2. Any entrapped air should be removed from the pressure system (pressure line and transducer) by filling with normal saline.
3. Connect the arterial catheter to the pressure line, ensure any entrapped air removed.

⚠ **Warning** ⚠

If any entrapped air in pressure system, re-fill system with normal saline.

4. Position the transducer at the same level of the patient's heart, approximately mid-axillary line.
5. Ensure the correct label name has been selected, Refer to the next section for details.
6. Zero the transducer. Refer to the next section for details.



Figure 16-1 IBP Monitoring

16.4 IBP Menu

Rotate the knob to move the cursor onto IBP hot key in the parameter area; press the knob to popup “IBP SELECT” menu shown as below: Pick the IBP SETUP to popup “IBPSETUP” menu shown as below:



Figure 16-2 IBP SELECT Menu



Figure 16-3 IBP SETUP Menu

Options that could be set up are:

- ALM: Select “ON” to enable alarm and data storage during IBP alarm. Select “OFF” to disable physiological alarm and display the symbol beside “IBP” numeric.
- ALM LEV: Set up the alarm level. Three levels are available: HIGH, MED, LOW.
- ALM REC: Select “ON” to enable recording once IBP alarm occurs. Select ”OFF” to disable recording function.
- AMP ADJUST: adjust waveform amplitude. Two selections are available: MANUAL, AUTO. Set it to AUTO, the pressure names of IBP become P, and the IBP scale is adjusted by system automatically. Set it to MANUAL, the pressure names of IBP can choose one of ART, PA, CVP, RAP, LAP, ICP, P, and the IBP scale is adjusted by the user via SCALE ADJUST item.
- SWEEP: Select the scanning speed of the IBP wave. Two selections are available: 12.5 mm/s or 25 mm/s.
- UNIT: Select the pressure unit (mmHg or kPa).
- FILTER: Select filtering mode of system. Three selections are available: NORMAL (filter the waveform at the frequency of 16Hz), SMOOTH (filter the waveform at the frequency of 8Hz) and NO FILTER (display the original waveform). The default is NO FILTER.
- ALM LIMIT SETUP: Access the sub-menu of IBP ALM LIMIT SETUP, in which user may set up the upper and lower alarm limit of systolic pressure, diastolic pressure and mean pressure.
- SCALE ADJUST: Access the sub-menu of IBP SCALE ADJUST, in which user may adjust the position of the high, reference and low scales for the two waveforms displayed on the screen.
- EXPAND PRESSURE: Access the sub-menu of IBP EXPAND PRESSURE, user could select the pressure type.
- DEFAULT: Access the IBP DEFAULT CONFIG dialog, in which user could select FACTORY DEFAULT CONFIG or USER DEFAULT CONFIG. After selecting an

option and exiting the dialog, the system will pop up a dialog asking for confirmation.

- EXIT: Exit the menu and return to the upper menu

⚠ Warning ⚠

Before setting the alarm limits, confirm to choose the correct label.



Figure 16-4 IBP ALM LIMIT SETUP

When the value exceeds the alarm limits, an alarm will occur.

IBP alarm limits:

Pressure Label	Max. Alarm High (mmHg)	Min. Alarm Low (mmHg)	Step (mmHg)
ART	300	0	1
PA	120	-6	1
CVP	40	-10	1
RAP	40	-10	1
LAP	40	-10	1
ICP	40	-10	1
P1/P2	300	-50	1

IBP Zeroing

Press the IBP PRESSURE ZERO button on the IBP SELECT menu to call up IBP PRESSURE ZERO menu as shown below:



Figure 16-5 IBP PRESSURE ZERO

⚠ **Note** ⚠

User should ensure that the transducer has been zeroed before measurement; otherwise the device does not have valid zero value, which may result in inaccurate measuring data.

Zero Transducer

Select “ZERO”, IBP is zeroed

Cautions:

- Close the transducer stopcock to the patient before zeroing.
- Open the venting stopcock to atmosphere.
- The transducer should be placed at the same level of the patient’s heart, approximately mid-axillary line.
- Zero procedure should be performed before starting the monitoring or at least once a day (or each time after connecting/disconnecting the cable).

Information related to zero:

- “SUCCESSFUL ZERO”
  Indicate the zero procedure has finished, open the transducer stopcock to the patient and close the venting stopcock to atmosphere.
- “SENSOR OFF, FAIL”
  Verify that the transducer does not fall off, then execute zeroing. If problem still exists, contact the serviceman.
- “IN DEMO FAIL”
  Ensure that the monitor is not in DEMO mode. Contact the serviceman if necessary.
- “PRESSURE OVER RANGE, FAIL”
  Ensure that the venting stopcock is opened to atmosphere, then execute zeroing. If the problem still exists, replace the transducer and contact the serviceman.
- “PULSATILE PRESSURE, FAIL”
  Ensure that the transducer is not opened to the patient and the stopcock is vented to atmosphere. Then execute zeroing. If the problem still exists, contact the serviceman.

IBP Calibration

Pick IBP PRESSURE CALIBRATE in the IBP SELECT menu to popup IBP PRESSURE CALIBRATE menu as shown below:



Figure 16-6 IBP Calibration Menu

Calibrate the transducer:

Turn the knob to select the item CAL VALUE, press and turn the knob to select the pressure value to be calibrated. Then turn the knob to select CALIBRATE to start calibrating.

- The pressure calibration of PM-8000



Figure 16-7 IBP Calibration

- Mercury calibration should be performed by the biomedical engineering department either whenever a new transducer is used or as periodically as requested by your Hospital regulation.
- The purpose of the calibration is to ensure that the system gives you accurate measurements.
- Before starting a mercury calibration, a zero procedure must be performed.
- If you need to perform this procedure yourself you will need the following pieces of equipment:

- Standard sphygmomanometer
- 3-way stopcock
- Tubing approximately 25 cm long

The Calibration Procedure: (SEE Figure 16-7)

⚠ **Warning** ⚠

You must never perform this procedure while patient is being monitored.

1. Disconnect transducer with patient.(when patient is monitored)
2. By using of tube, one end of T-type connector links to 3-way stopcock of transducer, another end links to inflation orb and the third end links to sphygmomanometer.
3. Vent the stopcock of transducer to atmosphere and run zeroing procedure. Open the stopcock to the sphygmomanometer side after successful zeroing.

4. Select the calibrated channel in “IBP calibration” menu and preset the calibration pressure of this channel.
5. Inflate sphygmomanometer and obtain the value of pressure to preset value in menu.
6. Repeatedly adjust the calibrating pressure value in the menu or the pressure value of sphygmomanometer until they are equal.
7. Push CALIBRATION button, the monitor starts calibrating process.
8. Wait for the result of calibration; determine the action according to the prompt message.
9. After calibration, disconnect the tube of sphygmomanometer and the T-type connector; then connect the transducer to the patient by following specified steps.

Related information to calibration

- “SUCCESSFUL CALIBRATION”
  Work properly, User could perform IBP monitoring.
- “SENSOR OFF , FALL”
  Check the connection of transducer, Ensure no “SENSOR OFF, FALL” message prompts, and Execute calibration. If problem still exists, contact serviceman.
- “IN DEMO, FAIL”
  Ensure that the monitor is not in DEMO mode, Execute calibration. If problem still exists, contact serviceman.
- “PRESSURE OVER RANGE, FAIL”
  Make sure that you have selected transducer value in IBP CAL, then proceed calibration.
- “PULSATILE PRESSURE, FALL”
  Ensure that the pressure value of the sphygmomanometer is constant. Execute calibration. If problem still exists, contact serviceman.

Changing the Label

- IBP SCALE ADJUST submenu:



Figure 16-8 IBP SCALE ADJUST Menu

The waveform and corresponding scale appears in the IBP Waveform Area with 3 dotted lines representing Higher Scale, Reference Scale, and Lower Scale from the top to the bottom.

Values of the three scales can be set according to the instruction given below.

- IBP label: Selectable from ART, PA, CVP, RAP, LAP, ICP, P;
- HI: IBP value of Higher scale , the range of which is the measurable range of current pressure.

⚠ **Note** ⚠

The HI value must be higher than the LO value.

- LO: IBP value of Lower scale , the range of which is the measurable range of current pressure.

⚠ **Note** ⚠

The LO value must be lower than the HI value.

- VAL: IBP value of Reference scale (between HI and LO).

.

⚠ **Note** ⚠

HI scale, LO scale, Reference scale and IBP waveform are displayed simultaneously on the screen, user could obviously view the change of the waveform after the scale has been adjusted.

16.5 Alarm Information and Prompts

Alarm Messages

Physiological alarm, caused by the parameter value exceeds the limits, will activate the recorder to automatically outputting the parameters and related measuring waveforms once the alarm occur while ALARM REC in related menu switch ON.

Tables below describe the possible physiological alarms, technical alarms and prompt messages occurring during IBP measurement.

Physiological alarms:

Message	Cause	Alarm Level
IS TOO HIGH	SYS measuring value is above upper alarm limit.	User-selectable
IS TOO LOW	SYS measuring value is below lower alarm limit.	User-selectable
ID TOO HIGH	DIA measuring value is above upper alarm limit.	User-selectable
ID TOO LOW	DIA measuring value is below lower alarm limit.	User-selectable
IM TOO HIGH	MAP measuring value is above upper alarm limit.	User-selectable
IM TOO LOW	MAP measuring value is below lower alarm limit.	User-selectable

Technical alarms:

Message	Cause	Alarm Level	Remedy
IBP SENSOR OFF	IBP cable of channel falls off from monitor.	LOW	Make sure that cable is properly connected.
IBP INIT ERR	IBP module failure	HIGH	Stop using measuring function of IBP module, notify biomedical engineer or Mindray service staff.
IBP INIT ERR1			
IBP INIT ERR2			
IBP INIT ERR3			
IBP INIT ERR4			
IBP INIT ERR5			
IBP INIT ERR6			
IBP INIT ERR7			
IBP INIT ERR8			
IBP COMM STOP	IBP module failure or communication failure	HIGH	Stop using ALARM function of IBP module, notify biomedical engineer or Our service staff.
IBP COMM ERR	IBP communication error	HIGH	Stop using ALARM function of IBP module, notify biomedical engineer or Our service staff.
IBP ALM LMT ERR	Functional safety failure	HIGH	Stop using ALARM function of IBP module, notify biomedical engineer or Our service staff.

Prompt message:

Message	Cause	Alarm Level
IBP SYS EXCEED	Systolic value is beyond measurement range.	HIGH
IBP DIA EXCEED	Diastolic measuring value is beyond measurement range.	HIGH
IBP MAP EXCEED	Mean measuring value is beyond measurement range.	HIGH
IBP NEED ZERO-CAL	IBP has not been zeroed.	LOW

16.6 Maintenance and Cleaning

16.6.1 Care and cleaning

⚠ **Warning** ⚠

Before cleaning the monitor or the transducer, turn off the power and disconnect from power line.

Cleaning of IBP Transducer (Reusable)

After the IBP monitoring operation is completed, remove the tubing and the dome from the transducer and wipe the transducer diaphragm with water. To clean the transducer and the cable, soak or wipe them by using soap or the detergents listed below:

Cetylcide
Wavicide-01
Wescodyne
Cidex
Lysol
Vesphene

Do not immerse the connector in any liquid. After cleaning, dry the transducer thoroughly before storing. Slight discoloration or temporary increase of surface stickiness of the cable should not be considered abnormal If adhesive tape residue must be removed from the transducer cable, double seal tape remover is effective and will cause a minimum of damage to the cable if used sparingly. Acetone, Alcohol, Ammonia and Chloroform, or other strong solvents are not recommended because over time the vinyl cabling will be damaged by these agents.

⚠ **Note** ⚠

The disposable transducers or domes must not be re-sterilized or re-used.

⚠ **Note** ⚠

For protecting environment, the disposable transducers or domes must be reclaimed or disposed properly

Sterilization

■ **Chemical solution Sterilization**

Remove obvious contamination by using the cleaning procedure described previously. Select a sterilant that has been found effective to your hospital or institution for chemical solution sterilization of operating room equipment. Buffered glutaraldehyde (e.g. Cidex or Hospisept) has been found to be effective. Do not use quaternary cationic detergents such as zephiran chloride. If the whole unit is to be sterilized, immerse the transducer but not the electrical connector into the sterilant for the recommended sterilizing period. Ensure that the dome has been removed. Then rinse all transducer parts except the electrical connector with sterilized water or saline. The transducer must be thoroughly dried before storing.

■ Gas Sterilization

For more complete asepsis, use gas sterilization.

Remove obvious contamination by using the cleaning procedure described previously. To inhibit the formation of ethylene glycol when ethylene oxide gas is used as the disinfectant, the transducer should be completely dry.

Follow the operating instructions provided by the manufacturer of the gas disinfectant.

⚠ **Warning** ⚠

The sterilize temperature must not exceed 70°C (158°F). Plastics in the pressure transducer may deform or melt above this temperature.

16.7 ICP Transducer ICT/B （Optional Accessory）

16.7.1 Introduction

The ICT/B is one of a of catheter tip transducers manufactured by Gaeltec. It is designed for measuring intracranial pressure by the epidural method. There are many advantages of catheter tip measurement including simplicity of use and excellent frequency response without artefacts.

The ICT/B has an atmospheric reference pressure channel that connects the back of the sensing area to the ambient air pressure via the luer fitting on the connector. All measurements are differential with respect to ambient air pressure.



Figure 16-9 ICT/B transducer

A significant feature of the ICT/B is the ability to check the zero drift of the ICT/B and pressure monitor in-vivo. Not only does this allows for accurate measurements, but also allows moving the patient with the ICT/B in the epidural space and reconnection to another monitor quickly and easily.

There is a flat silicone rubber membrane, or balloon, covering the pressure sensing diaphragm. Two internal tubes connect the two sides of the diaphragm to a female luer fitting on the connector shell. By introducing approximately 0.2 to 0.3ml of air from a 1ml syringe the pressure in these tubes will be greater than the ICP being measured. The exact amount of air is not critical, subject to the permitted maximum. When this air is injected, the pressure will cause the balloon to be lifted from the surface of the sensor and the same pressure will be applied to the back of the sensor. The strain gauge senses equal pressure above and below which is equivalent to having zero pressure applied. Thus by injecting a small volume of air, one undeflects the pressure sensor and checks the zero of the transducer and amplifier.

16.7.2 Cautions

⚠ **Note** ⚠

Gaeltec catheter tip pressure transducers are designed for use by trained physicians practicing a specialized branch of medicine. Use of the transducers should be restricted to those trained to perform the procedures.

⚠ **Note** ⚠

All pressure transducers must be used with patient monitors which meet the current safety standards for the country in which they are used and which are intended for use with strain gauge pressure transducers. The PATIENT MONITOR must provide electrical isolation between the transducer and any mains powered equipment to which the monitor is connected.

⚠ **Warning** ⚠

Disconnect the catheter from the monitor before defibrillation or electrosurgery.

⚠ **Warning** ⚠

Do not plug the female luer on the proximal end of the catheter during ethylene oxide sterilization or damage to the transducer may result.

⚠ **Warning** ⚠

Do not immerse or soak the electrical connector end in any kind of fluid or liquid.

⚠ **Warning** ⚠

The total volume of air injected to check the zero or baseline must not exceed 0.5ml or the membrane over the sensor may be ruptured.

⚠ **Warning** ⚠

Do not press with thumb and forefinger on the tip of the ICT/B. Enormous pressures will be generated this way and the device will be subject to possible damage. To see if the ICT/B is operating, gently touch the sensor tip.

⚠ **Note** ⚠

Carefully check for cuts on the silicone of the catheter and sensor tip before use.

- **PLACING THE ICT/B**

⚠ **Note** ⚠

Burr hole edges must be rounded where the catheter makes an "S" bend into the epidural space. Evacuate all bone chips.

⚠ **Note** ⚠

The catheter should be protected by suitable means where sutures are placed. This will prevent damage to the catheter when pulling sutures tight.

⚠ **Warning** ⚠

Do not use haemostats or forceps as these will damage the device.

⚠ **Note** ⚠

When removing the catheter, care should be taken not to nick the device while cutting sutures. Pull slowly on the catheter to remove the ICT/B.

16.7.3 Connection to the pressure monitor

⚠ **Note** ⚠

Although the catheter tip pressure transducer sensor is electrically isolated from the patient, it is recommended that pressure monitors with patient isolation be used for safety. Consult the manufacturer of the monitoring equipment for questions relating to monitor safety.

- **Calibration**

The ICT/B is supplied with minimal zero offset and the sensitivity is set at 5 uV/V/mmHg. In order to set up the amplifier and recorder accurately the controls should be zeroed at ambient pressure and then a known pressure applied, for instance using the calibration tube, syringe and manometer, or immersion to a known depth in a water column. The gain of the system is then set to the required level. The procedure should be repeated to check that the zero baseline has not changed due to the change in gain.


Manometer and syringe to apply a known pressure.
100
0
Patient Monitor
Use a 1ml syringe to inflate the balloon for zero check. Luer fitting must be open to ambient pressure during normal operation.
Calibration tube

Figure 16-10 ICT/B calibration

Tightening the collet on calibration tube over the sliding calibration sleeve will seal around the ICT/B catheter. Using the male luer fitting a connection can be made to a reference pressure, such as a syringe and manometer. The output of the transducer and amplifier system can be reliably and quickly confirmed.

- **Zeroing**

■ **Checking the Zero when the ICT/B is in the Epidural Space**

Using a 1ml syringe, inject approximately 0.3cc of air into the female luer connector on the proximal end of the ICT/B. Leave the syringe attached and note the value on the pressure monitor or scope. The ICP will decrease to zero or a value very close to it. If the monitor/transducer combination has drifted from zero, reset the zero control to zero value on the meter. Remove the syringe and the monitor will immediately begin to measure intracranial pressure.

⚠ **Warning** ⚠

The total volume of air injected from a 1ml syringe to check the zero must not exceed 0.5ml or the membrane over the sensor may be ruptured.

■ **Connecting to a new monitor when the ICT/B is in the Epidural Space**

- ✓ Set correct pressure range on monitor.
- ✓ Inject 0.3cc of air from a 1ml syringe.
- ✓ Adjust the monitor for zero reading.
- ✓ With the air still injected, set the calibration number on the monitor, if applicable.
- ✓ Remove the syringe and the ICP will be displayed immediately.

⚠ **Note** ⚠

Always leave the luer fitting open to ambient pressure during measurement.

⚠ **Warning** ⚠

Disconnect the catheter from the monitor before defibrillation or electrosurgery.

16.7.4 Practicing with the ICT/B

It is a good idea to obtain experience using the ICT/B and monitor combination before actually using the device with a patient.

Set up the monitor and the ICT/B as already described. Use either a water column or the calibration tube to apply a known pressure of from 10 to 25mmHg to the ICT/B. Recall that 13.6cm of water is about equal to 10mmHg.

With the known pressure applied to the ICT/B, inject approximately 0.3cc of air into the female luer using a 1ml syringe and note that the monitor does indeed immediately go to zero. Also note that if the ICT/B is moved rapidly up and down in a column of water pressure waves of high fidelity are seen. The ICT/B has a very high frequency response and you will observe excellent pressure waves in actual practice. It can also be confirmed that the exact amount of air injected to check the zero is not important.

16.7.5 Note to the neurosuregeon

The ICT/B is intended for the measurement of epidural pressures. Use of the transducer for the measurement of intraventricular pressures is not recommended. The ICT/B is designed for the measurement of positive pressures only.

⚠ **Note** ⚠

Catheter Tip Pressure Transducers must be used under the supervision of a suitably qualified Physician.

- **Method of Application of the ICT/B**

The application of the ICT/B may be accomplished through a variety of surgical techniques. Therefore, the surgeon is best advised to use the method which his own practice and discretion dictate to be best for the patient. The following are some general guidelines.

The ICT/B may be inserted during surgery or through a burr hole. When in place, the catheter tip transducer should have its pressure sensing surface facing against the dura, under the cranium. There are 2.5cm marks on the back of the catheter and these are visible when the sensor is facing in the proper direction.

The site of placement should be away from any craniotomy flap, preferably via a contralateral burr hole.

The dura mater should be carefully stripped at least 2cm under the skull and 180° in arc

before insertion. Failure to do this will result in wedging of the pressure sensor and inaccurate readings.

Protect the catheter :Use thick sutures and put tape around the catheter before suturing. Remove all bone chips. Use bone wax.


Use of a tunnel is recommended. A short length of a large bore catheter can be passed first to act as a guide for the sensor tip. Remove the guide from the burr hole end after the ICT/B has been passed to the burr hole from the tunnel entrance.
Note that both edges are smoothed. This will help to prevent damage to the catheter.
Strip the dura at least 2 cm under the inner table of the skull. Strip in an arc of at least 180°

Figure 16-11 ICT/B application

Reseat the transducer tip after a few days since the dura may rapidly tighten and change its physical characteristics.

If possible, round the bone at the point where the catheter makes its first bend into the burr hole and round the bone where the catheter makes the second bend under the cranium. This will help to prevent tearing the catheter or tip during insertion or removal. A tear will require the device to be returned for repair, an inconvenience that may be avoided by smoothing areas of bone in contact with the catheter transducer.

The catheter is led out through the wound in the manner of a drain. It may make sharp bends without disturbing the operation of the ICT/B. Care should be taken though, not to pinch the catheter by bending onto itself at acute angles for this will seal and possibly damage the internal lumens required for proper operation.

The catheter should be restrained from moving once the tip is in place. It may be fixed to the scalp by encircling sutures or with a silicone rubber suture collar available from manufacturers of such items as peritoneal shunt systems. The latter method is preferred as it will help prevent damage to the catheter by sutures or during removal of sutures.

Another method is by first approaching the burr hole through a tunnel under the skin (entering the tunnel from a point distal to the burr hole by making a small incision in the skin). The ICT/B can be guided in the tunnel by using a disposable tube removable from the burr hole side.

This latter method is to be preferred from both a mechanical stability point of view and from

the reported low incidence of infection. The catheter can then be led out in the manner of a drain and the burr hole incision sutured.

The physician is urged to examine the ICT/B for physical damage to the silicone rubber covering anywhere on the tip or catheter before use. If damage is suspected, do not use the catheter and return it to Manufactory for repair.

Proper function before insertion into the epidural space should be confirmed by gently touching the tip of the transducer and observing a deflection on the operating room pressure monitor.

Once the ICT/B has been inserted into the epidural space, the physician should check the proper function again, by injecting 0.3cc of air to check the zero of the ICT/B. The monitor should respond correctly as previously described.

- **Review of techniques to prevent damage to the catheter.**

1. In preparing the burr hole, it is imperative that the hole be rounded at the edges where the catheter makes an "S" bend into the epidural space.
2. Evacuate all bone chips.
3. A small pledgelet of woven bandage should be placed around the catheter where sutures will be placed. This will prevent damage to the catheter when pulling sutures tight. Otherwise you may cut the catheter.
4. Use some bone wax on the edges of bone where the catheter and tip make contact with bone.
5. When removing the catheter, care should be taken not to cut the device while cutting sutures. Remove the ICT/B by pulling slowly on the catheter.
6. The dura mater should be stripped sufficiently so that the tip of the sensor is not forced or wedged into place.
7. Do not use haemostats or forceps, they will damage the device. Do not squeeze the sensor between thumb and forefinger.

16.7.6 Cleaning and sterilization

⚠ **Warning** ⚠

Do not autoclave

⚠ **Warning** ⚠

Do not use radiation sterilization

⚠ **Warning** ⚠

Do not use ultrasonic cleaning

⚠ **Warning** ⚠

Do not use chlorinated hydrocarbons

⚠ **Warning** ⚠

Do not use toluene

⚠ **Warning** ⚠

Do not use sodium hypochlorite solution

⚠ **Note** ⚠

The ICT/B is supplied non-sterile. It must be cleaned and sterilised before each use.

Inspect for cuts or damage to silicone coating before immersing in any liquid. Be careful not to get liquid on the connector pins or inside the connector via the luer fitting

Wash the catheter with soap solution being careful not to poke the sensing area. Do not use synthetic detergents or oil based soaps as this may result in a foreign body reaction.

Transducers may be cleaned gently with alcohol wipes. Do not soak in alcohol.

Sterilisation is by means of cold aqueous solutions of detergicide (e.g. Cidex), formalin or by ethylene oxide gas.

⚠ **Warning** ⚠

Do not use the sterilizing cap during ETO gas sterilizing.

⚠ **Note** ⚠

Immediately after removal of the catheter from the patient, Checking for cuts in the silicone rubber.

Use a 1ml syringe to inject 0.5cc of air into the luer and immerse the catheter in water. If small bubbles are seen from any part of the catheter or tip, wipe dry and sterilize. Return to Manufactory for repair.

⚠ **Note** ⚠

It is recommended that each institution establish the efficacy of its sterilization procedure by a method which includes the sterilization of an intentionally contaminated product.

⚠ **Note** ⚠

There are only two chemical sterilization techniques recognized by the U.S. Department of Agriculture as effective and truly sporicidal, gas sterilization by ethylene oxide and liquid sterilization by a glutaraldehyde.

- **Ethylene oxide (ETO) Procedure**

Unplug the female luer on the proximal connector before the ETO sterilization cycle. Failure to do this will result in damage to the ICT/B and render it unusable. The luer must be open to allow free passage of ETO gas both internally and externally.

- Package the ICT/B in a coil in disposable ETO packaging. Include an approved sterilization indicator.
- Sterilize - "Normal Cycle" in an accepted commercially available hospital sterilizer. Follow the manufacturer's instructions for the sterilizer.

Use the following as a guide only. In an actual hospital sterilization facility, the following parameters were found to provide acceptable sterilization via ETO:

Sterilizer make and model	- AMSCO Eagle 2000
Prevacuum	- 15 minutes, 24 inches Hg
Relative humidity	- 40%
Temperature	- 140°F
ETO mix	- 12:88
Gas pressure	- 8 psi
Exposure time	- 1 hour 45 minutes
Post Vacuum	- 15 minutes, 24 inches Hg
Aeration Cycle	- 12 hours
Calculated ETO Concentration	- 600 mg/l

- **Liquid Sterilization Procedure**

Prevent liquids from entering the female luer on the electrical connector. A male plug may be used to do this. This plug must be removed during normal use and ETO sterilization.

1. Rinse and cold soak the catheter transducer in a solution of glutaraldehyde such as Cidex, following the chemical manufacturer's instructions. Note that disinfecting does not equal sterilisation and the strength of the glutaraldehyde must be confirmed by the chemical manufacturer's instructions.
2. After sterilisation of the catheter and just before use, rinse the device with pyrogen-free, sterile distilled water or saline solution as recommended by the manufacturer of the sterilising agent.

- **Care of the ICT/B**

The metal sensor is very robust and can withstand severe shocks and vibrations. It can be irreversibly damaged by contact with sharp objects or overpressure, for instance by squeezing the tip between finger and thumb.

The silicone coating on the sensing area allows a small amount of water absorption. During this process, which may take an hour or more, the baseline may drift a few mmHg. The device should be allowed to stabilize in water or saline before use for a few hours.

Liquids entering the back of the sensor will cause damage to the sensor. Cuts to the outer coating should be avoided and repaired immediately if any are found. Return to manufactory or apply a temporary repair using a suitable silicone sealant to the damage.

The most common reason for failure of the ICT/B pressure transducer is physical damage to the device's silicone catheter and/or tip. The cuts are usually caused by sharp bone segments and are not always visible to the naked eye. If such damage remains undetected fluids may enter the device and damage the sensing element. Check for damage as described in the Cleaning section of this manual.

16.7.7 SOME COMMONLY ASKED QUESTIONS

QUESTION	ANSWER
Is the ICT/B a single use device?	No, it is designed to be reused many times.
If it is damaged, what shall we do?	Sterilize first. Then, obtain a purchase order for repair and send it back to Manufactory for repair.
Does the air used for checking the zero get into the patient?	No. Air used for checking the zero stays in the fine lumens and tip of the ICT/B until the syringe is removed.
What happens if we autoclave the ICT/B?	It will have to be returned to Manufactory for repair.
We inject air to check the zero but the baseline on the scope always returns to ICP even if we leave the syringe attached. What is wrong?	There is a leak in the catheter or sensor tip. Remove, wipe clean with alcohol and then sterilize. Return the device to Manufactory for repair.
Readings were taken with the syringe left attached until we noticed it. Can we rely on these readings?	No. They are incorrect. All pressure readings must be made with the proximal female luer open to atmosphere.
Someone new on the staff began injecting water into the luer but we caught it just as a little went into the ICT/B. Is the device ruined?	Probably not. Return to Manufactory for repair.
Why will the ICT/B be damaged if we ETO sterilize it with the luer plugged?	When you plug the luer you are sealing the internal lumens at normal atmospheric pressure. Part of the ETO cycle is a partial vacuum. Thus, the trapped air at atmospheric pressure will expand and rupture the balloon.

16.7.8 TROUBLE-SHOOTING

TROUBLE	CAUSE	REMEDY
You inject air to check the zero and cal but the baseline reappears with the waveform showing.	The catheter or tip is cut and cannot hold zero long enough.	Readings cannot be trusted. Remove the ICT/B and use a spare. The waveforms will be accurate if that is all you need.
The monitor indicates 'damaged gauge' or 'over range' and if you inject air or not, you cannot see the waveform.	Either the tip is 'wedged' or the tip sensor was overpressured against the dura during insertion. Therefore the monitor is seeing a transducer that has a very high initial zero and finds this zero out of its range.	If the tip is wedged, pull back a few millimeters to free it. This will allow the monitor to be zeroed. If this does not help, the transducer has been strained and must come out and be returned for repair. Sometimes raising the scale on the monitor will allow it to manage a transducer with a high zero offset. Try raising the pressure scale to 90, 120 or 300mmHg and then setting

		zero. If this works, the only thing that you will sacrifice is the waveform resolution. Return the catheter for repair when the measurement is finished.
Everything was alright for several hours and then the 'damaged gauge' or 'over-range' light came on.	Although overpressured or wedged, the sensor zero must have been just within the range of the monitor. As conditions changed, the total pressure (=zero amount+ICP), pushed the monitor beyond its capabilities.	Try raising the pressure scale to 90, 120 or 300mmHg and then setting zero.
The transducer can be zeroed and we have good pressure waves but the ICP reads constantly near zero mmHg.	The sensor face must be flat (planar) against the dura. If its facing the inner table of the skull for example, then you will get pressure waves and be able to zero it but not obtain actual ICP readings. If indeed placed properly, the brain may have moved away from the skull substantially enough so that there is poor contact between the skull, transducer and the dura. This may happen soon after the transducer is placed but may correct itself in a short time.	It is important that the transducer face be placed against an intact section of dura. If required, use a contralateral burr hole.
We read negative ICP but get good waveforms on the monitor	Not proper zeroing.	The ICT/B cannot read negative pressure. Rezero the monitor /transducer combination. Make sure that you are not plugging the female luer during readings.
The waveform on the monitor makes large cyclical swings	If you are using a respirator or some device that applies pressure even indirectly it may affect ICP. The transducer is responding normally by showing this accurately.	/

Chapter 17 Accessories

This chapter lists the recommendation accessories used in this device.

⚠ **Warning** ⚠

Please use the specified accessories listed below with this patient monitor. The device will be possibly damaged or lead some harm if any other accessories are used.

17.1 ECG Accessories

Description	PN
Monitoring Electrode (10 electrodes per pack)	0010-10-12304
Monitoring Electrode (Pediatric, 2245, 25 electrodes per pack)	9000-10-07469
Monitoring Electrode (Neonatal, 2258-3, 3 electrodes per pack)	900E-10-04880
5 Lead Leadwires of snap (LL-22305)	6000-10-02006
6 Pin 5 Lead ECG Cable (LL-2514)	6000-10-02007
6 Pin 5 Lead ECG Cable (LL-2540)	9000-10-05163
3 Lead Leadwires of snap AHA (LL-22363)	9000-10-07445
5 Lead Leadwires of snap IEC	9000-30-07338
6 Pin 5 Lead ECG Cable IEC	9000-30-07339
6 Pin 3 Lead ECG Cable (LL-2325)	0509-10-00093
3 Lead Leadwires of snap IEC	9000-30-07470
6P 5 Lead ECG Cable with no resistance AHA	0010-30-12240
6P 5 Lead ECG Cable IEC with no resistance	0010-30-12241
6P 3 Lead ECG Cable with no resistance AHA	0010-30-12242
6P 3 Lead ECG Cable IEC with no resistance	0010-30-12243
6P 5 Lead ECG Cable with 1K resistance AHA	0010-30-12244
6P 5 Lead ECG Cable with 1K resistance IEC	0010-30-12245
6P 3 Lead ECG Cable with 1K resistance AHA	0010-30-12246
6P 3 Lead ECG Cable with 1K resistance IEC	0010-30-12247
6P ECG Trunk Cable with no resistance	0010-30-12256
6P ECG Trunk Cable with 1K resistance	0010-30-12257
5 Lead AHA Leadwires of clip	0010-30-12262
3 Lead AHA Leadwires of clip	0010-30-12263
5 Lead IEC Leadwires of clip	0010-30-12264

Description	PN
3 Lead IEC Leadwires of clip	0010-30-12265
5 Lead AHA Leadwires of snap	0010-30-12266
3 Lead AHA Leadwires of snap	0010-30-12267
5 Lead IEC Leadwires of snap	0010-30-12268
3 Lead IEC Leadwires of snap	0010-30-12269
6Pin 3-lead separable trunk cable with 1k resistance	0010-30-12377
6Pin 3-lead separable trunk cable with no resistance	0010-30-12378
Neonate 3-lead AHA leadwire of clip	0010-30-12381
Neonate 3-lead IEC leadwire of clip	0010-30-12382
Pediatric 3-lead AHA leadwire of clip	0010-30-12383
Pediatric 3-lead IEC leadwire of clip	0010-30-12384
Pediatric 3-lead AHA leadwire of snap	0010-30-12385
Pediatric 3-lead IEC leadwire of snap	0010-30-12386
Mixed-length 5-lead AHA leadwire of clip	0010-30-12387
Long 3-lead AHA leadwire of clip	0010-30-12388
Mixed-length 5-lead IEC leadwire of clip	0010-30-12389
Long 3-lead IEC leadwire of clip	0010-30-12390

17.2 SpO_2 Accessories

Description	PN
Mindray SpO_2	
DS-100A Adult SpO2 Sensor (Reusable)	9000-10-05161
OXI-P/I Pediatric/Infant Sensor and Sensor Wraps	9000-10-07308
OXI-A/N Adult/Neonatal Sensor and Sensor Wraps	9000-10-07336
Multisite SpO2 Sensor (Reusable, 518A)	518A-30-90226
Finger SpO2 Sensor (Reusable, 512B)	512B-30-90134
Finger SpO2 Sensor (Reusable, 512D)	512D-30-90200
Small SPO2 Ear Sensor (ES-3212-9)	0010-10-12392
6Pin SpO2 Cable	512D-30-16752
Masimo SpO_2	
LNOP-PDT Pediatric SpO2 Adhesive Sensor	0010-10-12107
LNOP-NEO Neonatal SpO2 Adhesive Sensor	0010-10-12108
LNOP-ADT Adult SpO2 Adhesive Sensor	0010-10-12109

Description	PN
LNOP-NEOPT Neonatal SpO2 Adhesive Sensor	0010-10-12110
NR7 Adult SpO2 Disposable Adhesive Sensor	0010-10-12100
LNOP-DCIP Pediatric SpO2 Reusable Sensor	0010-10-12102
LNOP-YI Multisite SpO2 Reusable Sensor	0010-10-12103
LNOP-EAR Ear SpO2 Reusable Sensor	0010-10-12104
LNOP-DC195 Adult Reusable Sensor (MASIMO P/N: 1560)	0010-10-12342
LNOP-DCI Adult SpO2 Reusable Sensor (MASIMO KIT)	0010-10-12274
6Pin SPO2 Cable (MASIMO KIT)	9200-30-10707

17.3 NIBP Accessories

Description	PN
NIBP Hose	509B-30-06259
Neonatal NIBP Hose	509B-30-06260
Infant 10 to 19 cm Arm Circumference (CM1201)	0010-30-12157
Child 18 to 26 cm Arm Circumference (CM1202)	0010-30-12158
Adult 25 to 35 cm Arm Circumference (CM1203)	0010-30-12159
Large Adult 33 to 47 cm Arm Circumference (CM1204)	0010-30-12160
Adult Thigh 46 to 66 cm Arm Circumference (CM1205)	0010-30-12161
M1872A Disposable Cuff (Size #4/7.1-13.1cm)	900E-10-04873
M1870A Disposable Cuff (Size #3/5.8-10.9cm)	900E-10-04874
M1868A Disposable Cuff (Size #2/4.3-8.0cm)	900E-10-04875
M1868A Disposable Cuff (Size #1/3.1-5.7cm)	900E-10-04876
W.A.BAUM Adult (Size 25-35cm Arm Circumference)	0010-30-12059
W.A.BAUM Child (Small Size 18-26cm Arm Circumference)	0010-30-12060
W.A.BAUM Infant (Size 10-19cm Arm Circumference)	0010-30-12061
Cuff without connector (Adult, CM1203, 25-35cm)	0010-10-12146
Cuff without connector (Infant, CM1201, 10-19cm)	0010-10-12147
Cuff without connector (Child, CM1202, 18-26cm)	0010-10-12148
Cuff without connector (Large Adult, CM1204, 33-47cm)	0010-10-12149
Cuff without connector (Adult Thigh, CM1205, 46-66cm)	0010-10-12150

17.4 TEMP Accessories

Description	PN
REF 427 Reusable Temperature Probe -Skin (Pediatric)	0010-10-12124
REF 401 Reusable Temperature Probe-Esophagesal/Rectal (Adult)	0509-10-00095
REF 402 Reusable Temperature Probe-Esophagesal/Rectal (Pediatric)	6000-10-01969
REF 409B Reusable Temperature Probe -Skin (Adult)	900E-10-04881
Adult reusable esophageal/rectal temperature probe	0011-30-90440
Pediatric/neonatal reusable esophageal/rectal temperature probe	0011-30-90441
Adult reusable skin-surface temperature probe	0011-30-90442
Pediatric/neonatal reusable skin-surface temperature probe	0011-30-90443
Reusable temperature probe extension cable	0011-30-90444
Disposable esophageal/rectal temperature probe	0011-30-90446
Disposable skin-surface temperature probe	0011-30-90447

17.5 IBP Accessories

Description	PN
Truware disposable pressure Transducer (PX260)	0010-10-12176
Transducer interface Cable (PX1800/896019021)	0010-10-12177
Edwards Transducer mount (DTH4)	0010-10-12192
Edwards Transducer mount (DTSC)	0010-10-12193
Truware disposable pressure transducer	0010-10-12208
Intracranial pressure transducer (ICT/B, Gaeltec)	0010-10-12151
6Pin ICP Cable	0010-21-12154
Transducer interface Cable (reusable)	6000-10-02106
Disposable pressure Transducer (DT-4812)	6000-10-02107
Transducer/manifold mount	0010-10-12156

Appendix A EC Declaration of Conformance

Manufacturer	Shenzhen Mindray Bio-Medical Electronics Co., Ltd.
Address	Mindray Building, Keji 12th Road South, Hi-tech Industrial Park, Nanshan, Shenzhen, 518057, P. R. China.
European Representative	Shanghai International Holding Corp. GmbH (Europe) Eiffestrasse 80 D-20537 Hamburg Germany
Product	Patient Monitor
Model Code	PM-8000 Standard Configuration including: ECG/RESP, NIBP, SpO_2, Battery Options: 1. Thermal Recorder 2. Temp 3. IBP

Classification (MDD, Annex IX): IIb

We herewith declare that the above mentioned products meet the provisions of the following EC Council Directives and Standards. All supporting documentations are retained under the premises of the manufacturer and the notified body.

DIRECTIVES

General applicable directives:

Medical Device Directive: COUNCIL DIRECTIVE 93/42/EEC of 14 June 1993 concerning medical devices(MDD 93/42/EEC).

Standards:

Harmonized Standards(published in the Official Journal of the European Communities) applicable to this product are:

ISO14971:2000+A1:2003, EN1041:1998, EN980:2003, IEC60878:2003, ISO1000:1992+A1:1998, ISO10993-1:2003, ISO3744:1994, EN540:1993, EN60601-1:1990+A1:1993+A2:1995+A13:1996, EN60601-1-1:2001, EN60601-1-2:2001, EN60601-1-4:2000, EN60601-2-27:1994, EN60601-2-30:2000, EN60601-2-34:2000, EN475:1995, EN865:1997, EN12470-4:2000, EN1060-1:1995, EN1060-3:1997, IEC60601-2-49:2001, ANSI/AAMI SP-10:1996.

Notified Body: TÜV Product Service GmbH, Ridlerstrasse 65 D- 80339 Münich, Germany.

Appendix B Product Specification

1 Classification

Anti-electroshock type	Class I equipment with internal power supply
Anti-electroshock degree	ECG (RESP), SpO_2, NIBP, IBP, TEMP: CF
Harmful liquid proof degree	Ordinary equipment (sealed equipment without liquid proof)
Disinfection/sterilizing method	Refer to Operation manual for details.
Working system	Continuous running equipment
EMC type	Class A

2 Specifications

B.1.1 Size and Weight

Size	Monitor	258×128×244 mm
Weight	Monitor	4.5 kg including Battery

B.1.2 Environment

Temperature		
	Working	0 ~ 40 °C
	Storage	-20 ~ 60 °C
Humidiity		
	Working	15%~95 % (noncomdensing)
	Storage	10%~ 95 % (noncomdensing)
Altitude		
	Working	-500 ~ 4,600m (-1,600 to 15,000ft)
	Storage	-500 ~ 13,100m (-1,600 to 43,000ft)
Power Supply		
		100 ~ 240 (V) AC, 50/60 (Hz)
		Pmax=100 VA
		FUSE T 1.6A

B.1.3 Display

Device	8.4 (in.) Color TFT, 800x600 Resolution, 3 LED
Messages	5 Waveforms Maximum

1 Alarm LED (Yellow/Red)
1 Power LED (Green)
1 Battery LED (Green)
3 Sound Mode corresponding Alarm Mode

B.1.4 Signal Interface

External Display	Standard VGA
Aux Output	BNC
ECG Output	
Sensitivity	1 V/mV ± 5% (Reference 10Hz)
Impedance	50 Ω
Signal Delay	≤25 ms
IBP Output	
Sensitivity	1 V/mmHg ± 5% (Reference 1Hz)
Impedance	50 Ω
Signal Delay	≤55 ms
NURSE CALL output	
Drive mode:	Relay Driven
Max. voltage:	36V DC, 25V AC
Max. load current:	2A
On resistance	<1Ω
Isolation voltage:	>1500VAC
Relay Drive:	ON/OFF

B.1.5 Battery

Single Rechargeable 2.3 Ah12V Lead-Acid battery
Operating time under the normal use and full charge greater than 60minutes
Operating time after the first alarm of low battery will be more than 5 minutes
Maximum charging time of single battery is 4 hours.

B.1.6 Recorder (Option)

Record Width	48 (mm)
Paper Speed	25/50(mm/s)
Trace	2
Recording types:	

Continuous real-time recording
8 second real-time recording
Auto 8 second recording
Parameter alarm recording
Waveform freeze recording
Trend graph/table recording
ARR events review recording

	Alarm event review recording
	NIBP review recording
	Hemodynamic Calculation result recording
	Monitor information recording
	OxyCRG recording

B.1.7 Recall

Trend Recall	
Short	1 hour, 1 second or 5 second Resolution
Long	72 (hrs), 1 min, 5min or 10 min Resolution
Alarm Event Recall	60 alarm events of all parameters and 8, 16 or 32 seconds of corresponding waveform.
NIBP Measurement Recall	400 NIBP measurement data.

B.1.8 ECG

Lead Mode	5 Leads (R, L, F, N, C or RA, LA, LL, RL, V)
Lead selection	I, II, III, avR, avL, avF, V, CAL
Waveform	2 ch
Lead mode	3 Leads (R, L, F or RA, LA, LL)
Lead selection	I, II, III, CAL
Waveform	1 ch
Gain	×0.25 mm/mV, ×0.5 mm/mV, ×1 mm/mV, ×2 mm/mV, auto
HR and Alarm	
Range	
Adult	15 ~ 300 (bpm)
Neo/Ped	15 ~ 350 (bpm)
Accuracy	± 1% or ± 1bpm, use the greater
Resolution	1 (bpm)
Sensitivity	≥200 (uV) P-P
Differential Input Impedance	≥5 MΩ
CMRR	
Monitor	≥ 105 dB
Surgery	≥ 105 dB
Diagnostic	≥ 90 dB
Electrode offset potential	±300mV
Patient Leakage Current	< 10 uA
Baseline Recovery After Defi	< 3 s
ECG Signal Range	± 8 mV p-p
Bandwidth	
Surgery	1 ~ 20 Hz

Monitor	0.5 ~ 40 Hz
Diagnostic	0.05 ~ 100 Hz
Calibration Signal	1 (mV) p-p, ±5% Accuracy
ST Segment Monitoring Range	
Measure and Alarm	-2.0 ~ +2.0 mV
Accuracy	-0.8 ~ +0.8 mV: ±0.02mV or ±10%, use the greater
ARR Detecting	
Type	ASYSTOLE, VFIB/VTAC, COUPLET, BIGEMINY, TRIGEMINY, R ON T, VT>2, PVC, TACHY, BRADY, MISSED BEATS, PNP, PNC
Alarm	Available
Review	Available

B.1.9 RESPARATION

Method	Impedance between RA-LL
Differential input Impedance:	>2.5 MOhm
Measuring Impedance Range:	0.3 ~ 5.0 Ω
Base line Impedance Range:	200Ω ~ 2500Ω
Bandwidth	0.2 ~ 2Hz (-3dB)
Resp.Rate	
Measuring and Alarm Range	
Adult	0 ~ 120 BrPM
Neo/Ped	0 ~ 150 BrPM
Resolution	1 BrPM
Accuracy	7 ~ 150 BrPM: ±2 BrPM or ±2%, whichever is greater. 0 ~ 6 BrPM: unspecified
Apean Alarm	10 ~ 40 (s)

B.1.10 NIBP

Method	Oscillometric
Mode	Manual, Auto, STAT
Measuring Interval in AUTO Mode	1, 2, 3, 4, 5, 10, 15, 30, 60, 90, 120, 180, 240, 480 min
Measuring Period in STAT Mode	5 (Min)
Alarm	
Type	SYS, DIA, MEAN
Measuring range	
Adult Mode	
SYS	40 ~ 270 mmHg
DIA	10 ~ 210 mmHg
MEAN	20 ~ 230 mmHg

Pediatric Mode		
	SYS	40 ~ 200 mmHg
	DIA	10 ~ 150 mmHg
	MEAN	20 ~ 165 mmHg
Neonatal Mode		
	SYS	40 ~ 135 mmHg
	DIA	10 ~ 100 mmHg
	MEAN	20 ~ 105 mmHg
Resolution		
Pressure		1mmHg
Cuff pressure accuracy		± 3mmHg
Accuracy		
Pressure		
Maximum Mean error		±5mmHg
Maximum Standard deviation		8mmHg
Overpressure Protection		
Adult Mode		297±3 mmHg
Pediatric Mode		240±3 mmHg
Neonatal Mode		147±3 mmHg

B.1.11 SpO_2

Measuring Range	0 ~ 100 %	
Alarm Range	0 ~ 100 %	
Resolution	1 %	
Accuracy	70% ~ 100%:	±2 %
	0% ~ 69%:	unspecified
Actualization interval	about 1(Sec.)	
Alarm Delay	10 (Sec.)	
Pulse Rate		
Measuring Range	20~254bpm	
Resolution	1bpm	
Accuracy	±3bpm	

MASIMO Specification:

Range		
Saturation (%SpO_2)	1% ~ 100%	
Pulse Rate (bmp)	25 ~ 240 bpm	
Accuracy		
Saturation(%SpO_2) — During No Motion Conditions		
Adults/pediatric	70% ~ 100%:	±2%
	0% ~ 69%:	unspecified

Neonates	70% ~ 100%:	±3%
	0% ~ 69%:	unspecified
Saturation(%SpO_2) — During Motion Conditions		
Adults/ pediatric/ Neonates	70% ~ 100%:	±3%
	0% ~ 69%:	unspecified
Pulse(bpm) — During No Motion Condition		
	25 ~ 240:	± 3bpm
Pulse(bpm) — During Motion Condition		
	25 ~ 240:	± 5 bpm

Resolution

Saturation (%SpO_2)	1%
Pulse Rate (bpm)	1

B.1.12 TEMPERATURE

Channel	2
Measuring and Alarm Range	0 ~ 50 °C
Resolution	0.1°C
Accuracy	±0.1°C (0 ~ 50 °C, exclusive the sensor)
Actualization interval	about 1s
Average Time Constant	< 10s

B.1.13 IBP

Channel	1
Label	ART, PA, CVP, RAP, LAP, ICP, P1, P2
Measuring and alarm range	
ART	0 ~ 300 (mmHg)
PA	-6 ~ 120 (mmHg)
CVP/RAP/LAP/ICP	-10 ~ 40 mmHg
P1/P2	-50 ~ 300 mmHg
Press Sensor	
Sensitivity	5uV/V/mmHg
Impedance	300 ~ 3000Ω
Resolution	1 mmHg
Accuracy	±2% or ±1mmHg, use the greater
Actualization interval	about 1s

Appendix C EMC

The monitor meets the requirements of EN 60601-1-2:2001

⚠ **Note** ⚠

The monitor needs special precautions regarding EMC and needs to be installed and put into service according to the EMC information provided below.

⚠ **Note** ⚠

Portable and mobile RF communications equipment can affect this monitor. See tables 1,2,3, and 4 below.

TABLE 1

Guidance and manufacturer's declaration — electromagnetic emmissions		
The **monitor** is intended for use in the electromagnetic environment specified below. The customer or the user of the monitor should assure that it is used in such an environment.		
Emissions test	**Compliance**	**Electromagnetic environment — guidance**
RF emissions CISPR 11	Group1	The **MONITOR** uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR 11	Class A	The **MONITOR** is suitable for use in all establishments other than domestic and those directly connected to the public low-voltage power supply network that supplies buildings used for domestic purposes
Harmonic Emissions IEC	Class A	
Voltage Fluctuations/ Flicker Emissions IEC 61000-3-3	Compliant	

TABLE 2

Guidance and manufacturer's declaration — electromagnetic immunity			
The **monitor** is intended for use in the electromagnetic environment specified below. The customer or the user of the monitor should assure that it is used in such an environment.			
Immunity test	**IEC 60601 Test level**	**Compliance level**	**Electromagnetic environment — guidance**
Electrostatic Discharge (ESD) IEC 61000-4-2	±6KV contact ±8KV air	±6KV contact ±8KV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%.
Electrical fast Transient/burst IEC 61000-4-4	±2 kV for power supply lines; ±1kV for I/O cable (>3m)	±2 kV for power supply lines; ±1kV for I/O cable (>3m)	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV differential mode ±2 kV common mode	±1 kV differential mode ±2 kV common mode	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, Short interruptions and voltage variation on power supply input lines IEC 61000-4-11	<5% U_T (>95% dip in U_T) for 0.5 cycle 40% U_T (60% dip in U_T) for 5 cycle 70% U_T (30% dip in U_T) for 25 cycle <5% U_T (>95% dip in U_T) for 5 sec	<5% U_T (>95% dip in U_T) for 0.5 cycle 40% U_T (60% dip in U_T) for 5 cycle 70% U_T (30% dip in U_T) for 25 cycle <5% U_T (>95% dip in U_T) for 5 sec	Mains power quality should be that of a typical commercial or hospital environment. If the user of the monitor requires continued operation during power mains interruptions, it is recommended that the monitor be powered from an uninterruptible power supply or a battery.
Power frequency (50/60 HZ) magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial of hospital environment.
NOTE — U_T is the a.c. mains voltage prior to application of the test level.			

TABLE 3

Guidance and manufacturer's declaration — electromagnetic immunity			
Portable and mobile RF communications equipment should be used no closer to any part of the monitor, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.			
NOTE 1 — At 80 MHz and 800 MHz, the higher frequency range applies. **NOTE 2** — These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			
The **monitor** is intended for use in the electromagnetic environment specified below. The customer or the user of the monitor should assure that it is used in such an environment.			
Immunity test	**IEC 60601 Test level**	**Compliance level**	**Electromagnetic environment — guidance**
Conducted RF IEC 61000-4-6	3 Vrms 150kHz to 80MHz	3 Vrms	Recommended separation distance: $d = 1.2 \times \sqrt{P}$
Radiated RF IEC 61000-4-3	3 V/m 80MHz to 2.5 GHz	3 V/m	Recommended separation distance: $d = 1.2 \times \sqrt{P}$ 80MHz to 800MHz $d = 2.3 \times \sqrt{P}$ 800MHz to 2.5GHz Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in meters (m). Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey, [a] should be less than the compliance level in each frequency range. [b] Interference may occur in the vicinity of equipment marked with the following symbol:
[a] Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast, and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the monitor is used exceeds the applicable RF compliance level above, the monitor should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the monitor. [b] Over the frequency range 150kHz to 80 MHz, field strengths should be less than 3V/m.			

TABLE 4

Recommended separation distances between portable and mobile RF communications equipment and the monitor			
The monitor is intended for user in an electromagnetic environment in which radiated RF disturbance are controlled. The customer or the user of the monitor can help prevent electromagnetic interference by maintain a minimum distance between portable and mobile RF communication equipment (transmitters) and the monitor as recommended below, according to the maximum output power of the communication equipment.			
Rated maximum output power of transmitter W	**Separation distance according to frequency of transmitter m**		
	150kHz～80MHz $d = \left[\frac{3.5}{3}\right]\sqrt{P}$	80MHz～800MHz $d = \left[\frac{3.5}{3}\right]\sqrt{P}$	800MHz～2.5GHz $d = \left[\frac{7}{3}\right]\sqrt{P}$
0.01	0.12	0.12	0.23
0.1	0.37	0.37	0.74
1	1.17	1.17	2.34
10	3.69	3.69	7.38
100	11.67	11.67	23.34
For transmitters rated at a maximum output power not listed above, the recommended separation distanced d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer. **NOTE 1** — At 80 MHz and 800 MHz, the separation distance for the higher frequency range applies. **NOTE 2** — These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.			

Appendix D System Alarm Prompt

PROMPT	CAUSE	MEASURE
"XX TOO HIGH"	XX value exceeds the higher alarm limit.	Check if the alarm limits are appropriate and the current situation of the patient.
"XX TOO LOW"	XX value is below the lower alarm limit.	
XX represents the value of parameter such as HR, ST1, ST2, RR, SpO2, IBP, NIBP, etc in the system.		
"ECG WEAK SIGNAL"	The ECG signal of the patient is too small so that the system can not perform ECG analysis.	Check if the electrodes and lead wires are connected correctly and the current situation of the patient.
“NO PULSE”	The pulse signal of the patient is too small so that the system can not perform pulse analysis.	Check the connection of the sensor and the current situation of the patient.
"RESP APNEA"	The respiration signal of the patient is too small so that the system cannot perform RESP analysis.	Check the connection of the linking wire and the current situation of the patient.
"ASYSTOLE"	Patient suffers from Arr. Of ASYSTOLE.	Check the current situation of the patient. Check the connection of the electrodes and lead wires.
"VFIB/VTAC"	Patient suffers from Arr. of VFIB/VTAC.	Check the current situation of the patient. Check the connection of the electrodes and lead wires.
"COUPLET"	Patient suffers from Arr. of COUPLET.	Check the current situation of the patient. Check the connection of the electrodes and lead wires.
"BIGEMINY"	Patient suffers from Arr. Of BIGEMINY.	Check the current situation of the patient. Check the connection of the electrodes and lead wires.
"TRIGEMINY"	Patient suffers from Arr. of TRIGEMINY.	Check the current situation of the patient. Check the connection of the electrodes and lead wires.
"R ON T"	Patient suffers from Arr. of R ON T.	Check the current situation of the patient. Check the connection of the electrodes and lead wires.
"PVC"	Patient suffers from Arr. of PVC.	Check the current situation of the patient. Check the connection of the electrodes and lead wires.
"TACHY"	Patient suffers from TACHY.	Check the current situation of the patient. Check the connection of the electrodes and lead wires.
" BRADY"	Patient suffers from BRADY.	Check the current situation of the patient. Check the connection of the electrodes and lead wires.
"VT>2"	Patient suffers from Arr. of VT>2.	Check the current situation of the patient. Check the connection of the electrodes and lead wires.
“MISSED BEATS”	Patient suffers from Arr. of	Check the current situation of the

	MISSED BEATS.	patient. Check the connection of the electrodes and lead wires.
"PNP"	The pacemaker is not paced.	Check the connection of the pacemaker. Check the connection of electrodes and lead wires. Check the current situation of the patient.
"PNC"	No pacemaker signal is captured.	Check the connection of the pacemaker. Check the connection of electrodes and lead wires. Check the current situation of the patient.
"ECG LEAD OFF"	ECG lead is not connected correctly.	Check the connection of ECG lead wire.
"ECG V LEAD OFF";	The V lead wire of ECG is not connected correctly.	Check the connection of V lead wire.
"ECG LL LEAD OFF";	The LL lead wire of ECG is not connected correctly.	Check the connection of LL lead wire.
"ECG LA LEAD OFF";	The LA lead wire of ECG is not connected correctly.	Check the connection of LA lead wire.
"ECG RA LEAD OFF";	The RA lead wire of ECG is not connected correctly.	Check the connection of RA lead wire.
"ECG C LEAD OFF";	The C lead wire of ECG is not connected correctly.	Check the connection of C lead wire.
"ECG F LEAD OFF";	The F lead wire of ECG is not connected correctly.	Check the connection of F lead wire.
"ECG L LEAD OFF";	The L lead wire of ECG is not connected correctly.	Check the connection of L lead wire.
"ECG R LEAD OFF";	The R lead wire of ECG is not connected correctly.	Check the connection of R lead wire.
SPO2 SENSOR OFF	SpO_2 sensor may be disconnected from the patient or the monitor.	Make sure that the monitor and the patient are in correct connection with the cables.
SPO2 INIT ERR	SpO_2 module failure	Stop using the measuring function of SpO_2 module, notify biomedical engineer or Mindray service staff.
SPO2 INIT ERR 1		
SPO2 INIT ERR 2		
SPO2 INIT ERR 3		
SPO2 INIT ERR 4		
SPO2 INIT ERR 5		
SPO2 INIT ERR 6		
SPO2 INIT ERR 7		
SPO2 INIT ERR 8		
SPO2 COMM STOP	SpO_2 module failure or communication error	Stop using the measuring function of SpO_2 module, notify biomedical engineer or Mindray service staff.
SPO2 COMM ERR	SpO_2 module failure or communication error	Stop using the measuring function of SpO_2 module, notify biomedical engineer or Mindray service staff.

SPO2 ALM LMT ERR	Functional safety failure	Stop using the measuring function of SpO_2 module, notify biomedical engineer or Mindray service staff.
PR ALM LMT ERR	Functional safety failure	Stop using the measuring function of SpO_2 module, notify biomedical engineer or Mindray service staff.
MASIMO Alarm information:		
SpO2 NO SENSOR	Sensor not fully inserted into the connector.	May be an incorrect sensor, or a defective sensor or cable. Insert sensor into the connector. Disconnect and reconnect sensor. Refer to the instructions for the sensor being used.
	Sensor inserted upside down.	Disconnect and reconnect he sensor with the logos matching.
SpO2 SENSOR OFF	SpO2 sensor may be disconnected from the patient or the monitor.	Disconnect and reconnect the sensor. Reattach sensor.
SpO2 SENSOR FAULT	This message appears when the sensor is faulty	Stop using the measuring function of SpO2 module, notify biomedical engineer or our service staff.
SpO2 UNRECOGNIZED SENSOR	Masimo board does not recognize the sensor.	Make sure that the monitor and the patient are in correct connection with the cables.
SpO2 INCOMPATIBLE SENSOR	This message is displayed when the masimo sensor is finding incompatible sensor.	Make sure that the monitor use incompatible sensor.
SpO2 INTERFERENCE	Outside signal or energy preventing reading.	Remove outside interference.
SpO2 PULSE SEARCH	Unit is searching for the patients pulse.	If values are not displayed within 30 seconds, disconnect and reconnect sensor. If pulse search continues, remove sensor and replace on a better perfused site.
SpO2 LOW PERFUSTION	Signal too small.	Move sensor to better perfused site.
SpO2 TOO MUCH LIGHT	Too much light on patient(sensor). Inadequate tissue covering sensor detector.	Remove or reduce lighting. Cover sensor from light. Reposition sensor.
SpO2 LOW SIGNAL IQ	Low signal quality.	Ensure proper sensor application. Mover sensor to a better perfused site.
SpO2 BOARD FAULT	This message appears when the Masimo Set board malfunctions.	Stop using the measuring function of SpO2 module, notify biomedical engineer or our service staff.

SpO2 COMMUNICATION ERROR	This message is displayed when the front end module is having problems communicating (ie: framing errors or bad checksums) with the Masimo board.	Stop using the measuring function of SpO2 module, notify biomedical engineer or our service staff.
SpO2 COMMUNICATION STOP	This message is displayed when the host can not receive the data from Masimo board for 5 seconds	Stop using the measuring function of SpO2 module, notify biomedical engineer or our service staff.
SpO2 INIT ERR	This message is displayed when the SpO2 module initialization error happened.	Stop using the measuring function of SpO2 module, notify biomedical engineer or our service staff.
"TEMP SENSOR OFF"	TEMP sensor is not connected correctly.	Check the connection of TEMP sensor.
"IBP LEAD OFF"	IBP sensor is not connected correctly.	Check the connection of IBP sensor.
"IBP NEED ZERO-CAL"	Zero calibrating must be done before measuring in IBP	Do zero calibrating for IBP
"TB SENSOR OFF"	TB sensor is not connected correctly.	Check the connection of TB sensor.
"ECG NOISE"	Rather large interference signals appear in the ECG signals.	Check the connection of ECG lead wire. Check the current situation of the patient. Check if the patient moves a lot.
"XX INIT ERR X"	XX has error X during initialization.	Re-start up the monitor or re-plug in/out the module. If the error still exists, contact the manufacturer.
"XX COMM STOP"	XX cannot communicate with the host.	
"XX COMM ERR"	XX cannot communicate normally with the host.	
XX represents all the parameter modules in the system such as ECG, NIBP, SpO2, IBP, CO module, etc.		
"XX ALM LMT ERR"	The alarm limit of XX parameter is modified by chance.	Contact the manufacturer for repair.
"XX RANGE EXCEEDED"	The measured value of XX parameter has exceeded the measuring range of the system.	Contact the manufacturer for repair.
"REAL CLOCK NEEDSET"	When the system displays 2000-1-1, the system gives this prompt reminding the user that the current system time is not right.	Re-set up the system time. It is better to set up the time just after the start-up and prior to monitoring the patient. After modifying the time, the user had better re-start up the monitor to avoid storing error time.
"REAL CLOCK NOT EXIST"	The system has no cell battery or the battery has run out of the capacity.	Install or replace the rechargeable battery.

"SYSTEM WD FAILURE"	The system has serious error.	Re-start up the system. If the failure still exists, contact the manufacturer.
"SYSTEM SOFTWARE ERR"		
"SYSTEM CMOS FULL"		
"SYSTEM CMOS ERR"		
"SYSTEM EPGA FAILURE"		
"SYSTEM FAILURE2"		
"SYSTEM FAILURE3"		
"SYSTEM FAILURE4"		
"SYSTEM FAILURE5"		
"SYSTEM FAILURE6"		
"SYSTEM FAILURE7"		
"SYSTEM FAILURE8"		
"SYSTEM FAILURE9"		
"SYSTEM FAILURE10"		
"SYSTEM FAILURE11"		
"SYSTEM FAILURE12"		
"KEYBOARD NOT AVAILABLE";	The keys on the keyboard cannot be used.	Check the keys to see whether it is pressed manually or by other object. If the key is not pressed abnormally, contact the manufacturer for repair.
"KEYBOARD COMM ERR";	The keyboard has failure, which cannot be used.	Contact the manufacturer for repair.
"KEBOARD ERROR";		
"KEYBOARD ERR1";		
"KEYBOARD ERR2";		
"NET INIT ERR(G.)"	The network part in the system has failure. The system cannot be linked to the net.	Contact the manufacturer for repair.
"NET INIT ERR(Ram)"		
"NET INIT ERR(Reg)"		
"NET INIT ERR(Mii)"		
"NET INIT ERR(Loop)"		
"NET ERR(Run1)"		
"NET ERR(Run2)"		
"NET ERR(Run3)"		
"5V TOO HIGH"	The power part of the system has failure.	If the prompt appears repeatedly, contact the manufacturer for repair.
"5V TOO LOW"		
"POWER ERR3"		
"POWER ERR4"		
"12V TOO HIGH"		
"12V TOO LOW"		
"POWER ERR7"		
"POWER ERR8"		
"3.3V TOO HIGH"		
"3.3V TOO LOW"		
"CELL BAT TOO HIGH"	Cell battery has problem.	Replace the battery. If the failure still exists, contact the manufacturer.
"CELL BAT TOO LOW"	The cell battery has low capacity or the cell battery is not installed or the connection is loose.	
"RECORDER SELFTEST ERR"	During the selftest, the system fails connecting with the recorder module.	Execute ‘Clear Record Task’ function in the recorder setup menu to re-connect the host and the recorder.

		If the failure still exists, contact the manufacturer for repair.
"RECORDER VLT HIGH"	The recorder module has voltage failure.	Contact the manufacturer for repair.
"RECORDER VLT LOW"		
"RECORDER HEAD HOT"	The continuous recording time may be too long.	After the recorder becomes cool, use the recorder for output again. If the failure still exists, contact the manufacturer for repair.
"REC HEAD IN WRONG POSITION"	The handle for pressing the paper is not pressed down.	Press down the recorder handle for pressing the paper.
"RECORDER OUT OF PAPER"	No paper is in the recorder.	Place the paper into the recorder.
"RECORDER PAPER JAM"	The paper in the recorder is jammed.	Place the recorder correctly and try again.
"RECORDER COMM ERR"	The communication of the recorder is abnormal.	In the recorder setup menu, execute the function of clearing record task. The function can make the host and the recorder connect again. If the failure still exists, contact the manufacturer for repair.
"RECORDER S. COMM ERR"		
"RECORDER PAPER W.P."	The paper roll of the recorder is not placed in the correction position.	Place the paper roll in the correct position.
"REC NOT AVAILABLE"	Cannot communicate with the recorder.	In the recorder setup menu, execute the function of clearing record task. The function can make the host and the recorder connect again. If the failure still exists, contact the manufacturer for repair.
"NIBP INIT ERR"	NIBP initialization error	Execute the reset program in the NIBP menu. If the failure still exists, contact the manufacturer for repair.
"NIBP SELFTEST ERR"		
"NIBP ILLEGALLY RESET"	During NIBP measurement, illegal reset occurs.	Check the airway of NIBP to see if there are clogs. Then measure again, if the failure still exists, contact the manufacturer for repair.
"NIBP COMM ERR"	The NIBP communication part has problem.	Execute the reset program in the NIBP menu. If the failure still exists, contact the manufacturer for repair.
"LOOSE CUFF"	The NIBP cuff is not connected correctly.	Re-connect the NIBP cuff.
"AIR LEAK"	The NIBP cuff is not connected correctly or there are leaks in the airway.	Check the connection of each part or replace with a new cuff. If the failure still exists, contact the manufacturer for repair.
"AIR PRESSURE ERROR"	Problem happens when measuring the curve. The system cannot perform measurement, analysis or calculation.	Check the connection of each part or replace with a new cuff. If the failure still exists, contact the manufacturer for repair.
"WEAK SIGNAL"	Problem happens when measuring the curve. The system cannot perform measurement, analysis or calculation.	Check if the setup of patient type is correct. Check the connection of each part or replace with a new cuff. If the failure still exists, contact the manufacturer for repair
"RANGE EXCEEDED"	Problem happens when measuring the curve. The	Check the connection of each part or replace with a new cuff. If the failure

	system cannot perform measurement, analysis or calculation.	still exists, contact the manufacturer for repair.
"EXCESSIVE MOTION"	The patient arm moves.	Check the connection of each part and the patient situation. Measure again, if the failure still exists, contact the manufacturer for repair.
"OVER PRESSURE"	Perhaps folds exist in the airway.	Check for the smoothness in the airway and patient situation. Measure again, if the failure still exists, contact the manufacturer for repair.
"SIGNAL SATURATED"	Problem happens when measuring the curve. The system cannot perform measurement, analysis or calculation.	Check the connection of each part and the patient situation. Measure again, if the failure still exists, contact the manufacturer for repair.
"NIBP TIME OUT"	Problem happens when measuring the curve. The system cannot perform measurement, analysis or calculation.	Check the connection of each part and the patient situation. Measure again, if the failure still exists, contact the manufacturer for repair.
"CUFF TYPE ERR"	Perhaps the used cuff does not fit the setup patient type.	Check if the patient type is set up correctly. Check the connection of each part or replace with a new cuff. If the failure still exists, contact the manufacturer for repair.
"PNEUMATIC LEAK"	NIBP airway has leaks.	Check the connection of each part or replace with a new cuff. If the failure still exists, contact the manufacturer for repair.
"MEASURE FAIL"	Problem happens when measuring the curve. The system cannot perform measurement, analysis or calculation.	Check the connection of each part and the patient situation. Measure again, if the failure still exists, contact the manufacturer for repair.
"NIBP SYSTEM FAILURE"	Problem happens when measuring the curve. The system cannot perform measurement, analysis or calculation.	Check the connection of each part and the patient situation. Measure again, if the failure still exists, contact the manufacturer for repair.

FOR YOUR NOTES

P/N:8000-20-10197